

Data Path : Z:\HPCHEM1\BNA F\Data\BF040618\
 Data File : BF104307.D
 Acq On : 6 Apr 2018 15:33
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 06 23:07:31 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 06 16:44:53 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.609	0.604	0.8	99	0.00
3	Pyridine	1.714	1.682	1.9	99	0.00
4	n-Nitrosodimethylamine	0.599	0.573	4.3	96	0.00
5 S	2-Fluorophenol	1.308	1.230	6.0	96	0.00
6	Aniline	2.233	2.159	3.3	97	0.00
7 S	Phenol-d6	1.607	1.546	3.8	99	0.00
8	2-Chlorophenol	1.381	1.338	3.1	98	0.00
9	Benzaldehyde	0.802	0.814	-1.5	99	0.00
10 C	Phenol	1.705	1.639	3.9	99	0.00
11	bis(2-Chloroethyl)ether	1.361	1.329	2.4	99	0.00
12	1,3-Dichlorobenzene	1.564	1.511	3.4	98	0.00
13 C	1,4-Dichlorobenzene	1.559	1.508	3.3	98	0.00
14	1,2-Dichlorobenzene	1.445	1.394	3.5	97	0.00
15	Benzyl Alcohol	1.221	1.164	4.7	97	0.00
16	2,2'-oxybis(1-Chloropropane	1.827	1.756	3.9	97	0.00
17	2-Methylphenol	1.200	1.167	2.7	98	0.00
18	Hexachloroethane	0.556	0.544	2.2	100	0.00
19 P	n-Nitroso-di-n-propylamine	1.050	0.979	6.8	99	0.00
20	3+4-Methylphenols	1.488	1.437	3.4	99	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
22	Acetophenone	0.492	0.475	3.5	99	0.00
23 S	Nitrobenzene-d5	0.306	0.319	-4.2	102	0.00
24	Nitrobenzene	0.338	0.348	-3.0	100	0.00
25	Isophorone	0.648	0.617	4.8	97	0.00
26 C	2-Nitrophenol	0.138	0.152	-10.1	106	0.00
27	2,4-Dimethylphenol	0.286	0.281	1.7	98	0.00
28	bis(2-Chloroethoxy)methane	0.396	0.383	3.3	98	0.00
29 C	2,4-Dichlorophenol	0.273	0.267	2.2	98	0.00
30	1,2,4-Trichlorobenzene	0.299	0.290	3.0	98	0.00
31	Naphthalene	0.996	0.963	3.3	98	0.00
32	Benzoic acid	0.228	0.249	-9.2	103	0.00
33	4-Chloroaniline	0.427	0.413	3.3	98	0.00
34 C	Hexachlorobutadiene	0.164	0.162	1.2	100	0.00
35	Caprolactam	0.095	0.093	2.1	97	0.00
36 C	4-Chloro-3-methylphenol	0.292	0.285	2.4	98	0.00
37	2-Methylnaphthalene	0.644	0.629	2.3	100	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
39	1,2,4,5-Tetrachlorobenzene	0.573	0.546	4.7	101	0.00
40 P	Hexachlorocyclopentadiene	0.321	0.318	0.9	101	0.00
41 S	2,4,6-Tribromophenol	0.207	0.196	5.3	98	0.00
42 C	2,4,6-Trichlorophenol	0.397	0.386	2.8	98	0.00
43	2,4,5-Trichlorophenol	0.445	0.428	3.8	100	0.00
44 S	2-Fluorobiphenyl	1.266	1.193	5.8	100	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF040618\
 Data File : BF104307.D
 Acq On : 6 Apr 2018 15:33
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 06 23:07:31 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 06 16:44:53 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.680	1.594	5.1	100	0.00
46	2-Chloronaphthalene	1.327	1.258	5.2	100	0.00
47	2-Nitroaniline	0.328	0.357	-8.8	106	0.00
48	Acenaphthylene	2.082	1.983	4.8	99	0.00
49	Dimethylphthalate	1.577	1.485	5.8	100	0.00
50	2,6-Dinitrotoluene	0.292	0.315	-7.9	104	0.00
51 C	Acenaphthene	1.270	1.199	5.6	99	0.00
52	3-Nitroaniline	0.342	0.366	-7.0	104	0.00
53 P	2,4-Dinitrophenol	0.089	0.094	-5.6	112	0.00
54	Dibenzofuran	1.770	1.675	5.4	99	0.00
55 P	4-Nitrophenol	0.268	0.282	-5.2	101	0.00
56	2,4-Dinitrotoluene	0.383	0.403	-5.2	104	0.00
57	Fluorene	1.289	1.252	2.9	101	0.00
58	2,3,4,6-Tetrachlorophenol	0.332	0.327	1.5	103	0.00
59	Diethylphthalate	1.571	1.471	6.4	99	0.00
60	4-Chlorophenyl-phenylether	0.574	0.558	2.8	100	0.00
61	4-Nitroaniline	0.334	0.359	-7.5	102	0.00
62	Azobenzene	1.501	1.417	5.6	99	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
64	4,6-Dinitro-2-methylphenol	0.087	0.097	-11.5	108	0.00
65 c	n-Nitrosodiphenylamine	0.693	0.663	4.3	99	0.00
66	4-Bromophenyl-phenylether	0.213	0.202	5.2	99	0.00
67	Hexachlorobenzene	0.239	0.228	4.6	99	0.00
68	Atrazine	0.209	0.197	5.7	98	0.00
69 C	Pentachlorophenol	0.148	0.150	-1.4	99	0.00
70	Phenanthrene	1.114	1.047	6.0	98	0.00
71	Anthracene	1.132	1.068	5.7	98	0.00
72	Carbazole	1.106	1.036	6.3	98	0.00
73	Di-n-butylphthalate	1.351	1.264	6.4	97	0.00
74 C	Fluoranthene	1.164	1.096	5.8	98	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
76	Benzidine	0.692	0.663	4.2	97	0.00
77	Pyrene	1.482	1.434	3.2	98	0.00
78 S	Terphenyl-d14	0.877	0.829	5.5	99	0.00
79	Butylbenzylphthalate	0.698	0.684	2.0	98	0.00
80	Benzo(a)anthracene	1.195	1.133	5.2	98	0.00
81	3,3'-Dichlorobenzidine	0.445	0.440	1.1	97	0.00
82	Chrysene	1.227	1.161	5.4	95	0.00
83	Bis(2-ethylhexyl)phthalate	0.898	0.869	3.2	99	0.00
84 c	Di-n-octyl phthalate	1.501	1.492	0.6	97	0.00
85	Indeno(1,2,3-cd)pyrene	1.043	0.969	7.1	97	0.00
86 I	Perylene-d12	1.000	1.000	0.0	95	0.00
87	Benzo(b)fluoranthene	1.263	1.219	3.5	91	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF040618\
Data File : BF104307.D
Acq On : 6 Apr 2018 15:33
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 06 23:07:31 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 06 16:44:53 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.193	1.194	-0.1	96	0.00
89 C	Benzo(a)pyrene	1.130	1.110	1.8	93	0.00
90	Dibenzo(a,h)anthracene	0.928	0.910	1.9	96	0.00
91	Benzo(a,h,i)perylene	0.899	0.872	3.0	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0