

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040622\
 Data File : BF127669.D
 Acq On : 07 Apr 2022 01:26
 Operator : CG\JU
 Sample : N2272-08
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-3

Quant Time: Apr 07 02:41:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 06 13:55:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.963	152	99670	20.000	ng	0.00
21) Naphthalene-d8	8.245	136	369066	20.000	ng	# 0.00
39) Acenaphthene-d10	10.010	164	214700	20.000	ng	0.00
64) Phenanthrene-d10	11.498	188	413199	20.000	ng	0.00
76) Chrysene-d12	14.151	240	371186	20.000	ng	0.00
86) Perylene-d12	15.674	264	331194	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.587	112	791382	129.018	ng	0.00
7) Phenol-d6	6.581	99	947675	118.098	ng	0.00
23) Nitrobenzene-d5	7.528	82	774548	93.561	ng	0.00
42) 2,4,6-Tribromophenol	10.798	330	331326	145.018	ng	0.00
45) 2-Fluorobiphenyl	9.328	172	1357487	89.445	ng	0.00
79) Terphenyl-d14	13.098	244	1844118	82.005	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040622\
Data File : BF127669.D
Acq On : 07 Apr 2022 01:26
Operator : CG\JU
Sample : N2272-08
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-3

Quant Time: Apr 07 02:41:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 06 13:55:27 2022
Response via : Initial Calibration

