

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040622\
 Data File : BF127647.D
 Acq On : 06 Apr 2022 14:42
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF040622

Quant Time: Apr 06 16:42:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 06 13:55:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	97	0.00
2	1,4-Dioxane	40.000	40.226	-0.6	97	0.00
3	Pyridine	40.000	40.187	-0.5	97	0.00
4	n-Nitrosodimethylamine	40.000	40.399	-1.0	97	0.00
5 S	2-Fluorophenol	80.000	78.497	1.9	98	0.00
6	Aniline	40.000	40.502	-1.3	96	0.00
7 S	Phenol-d6	80.000	79.336	0.8	98	0.00
8	2-Chlorophenol	40.000	40.054	-0.1	100	0.00
9	Benzaldehyde	40.000	41.823	-4.6	104	0.00
10 C	Phenol	40.000	39.973	0.1	98	0.00
11	bis(2-Chloroethyl)ether	40.000	39.277	1.8	97	0.00
12	1,3-Dichlorobenzene	40.000	38.849	2.9	96	0.00
13 C	1,4-Dichlorobenzene	40.000	39.291	1.8	98	0.00
14	1,2-Dichlorobenzene	40.000	37.891	5.3	97	0.00
15	Benzyl Alcohol	40.000	40.219	-0.5	96	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.847	0.4	96	0.00
17	2-Methylphenol	40.000	40.133	-0.3	98	0.00
18	Hexachloroethane	40.000	39.616	1.0	96	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.682	-1.7	99	0.00
20	3+4-Methylphenols	40.000	40.227	-0.6	98	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	39.154	2.1	99	0.00
23 S	Nitrobenzene-d5	80.000	81.341	-1.7	98	0.00
24	Nitrobenzene	40.000	40.679	-1.7	98	0.00
25	Isophorone	40.000	40.228	-0.6	98	0.00
26 C	2-Nitrophenol	40.000	42.733	-6.8	98	0.00
27	2,4-Dimethylphenol	40.000	39.201	2.0	99	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.989	0.0	98	0.00
29 C	2,4-Dichlorophenol	40.000	39.770	0.6	98	0.00
30	1,2,4-Trichlorobenzene	40.000	39.515	1.2	97	0.00
31	Naphthalene	40.000	39.275	1.8	98	0.00
32	Benzoic acid	40.000	43.964	-9.9	100	0.00
33	4-Chloroaniline	40.000	39.667	0.8	98	0.00
34 C	Hexachlorobutadiene	40.000	39.172	2.1	97	0.00
35	Caprolactam	40.000	41.105	-2.8	98	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.232	-0.6	97	0.00
37	2-Methylnaphthalene	40.000	39.614	1.0	98	0.00
38	1-Methylnaphthalene	40.000	39.405	1.5	98	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	99	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.071	2.3	97	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.997	0.0	95	0.00
42 S	2,4,6-Tribromophenol	80.000	82.176	-2.7	103	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.466	1.3	97	0.00
44	2,4,5-Trichlorophenol	40.000	40.707	-1.8	100	0.00
45 S	2-Fluorobiphenyl	80.000	76.550	4.3	98	0.00
46	1,1'-Biphenyl	40.000	39.477	1.3	99	0.00
47	2-Chloronaphthalene	40.000	39.244	1.9	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040622\
 Data File : BF127647.D
 Acq On : 06 Apr 2022 14:42
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF040622

Quant Time: Apr 06 16:42:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 06 13:55:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.019	-10.0	102	0.00
49	Acenaphthylene	40.000	39.986	0.0	100	0.00
50	Dimethylphthalate	40.000	39.461	1.3	98	0.00
51	2,6-Dinitrotoluene	40.000	42.603	-6.5	101	0.00
52 C	Acenaphthene	40.000	39.467	1.3	98	0.00
53	3-Nitroaniline	40.000	42.026	-5.1	101	0.00
54 P	2,4-Dinitrophenol	40.000	40.222	-0.6	99	0.00
55	Dibenzofuran	40.000	39.706	0.7	100	0.00
56 P	4-Nitrophenol	40.000	42.896	-7.2	102	0.00
57	2,4-Dinitrotoluene	40.000	43.776	-9.4	104	0.00
58	Fluorene	40.000	39.305	1.7	100	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.185	-0.5	100	0.00
60	Diethylphthalate	40.000	39.940	0.2	100	0.00
61	4-Chlorophenyl-phenylether	40.000	39.577	1.1	100	0.00
62	4-Nitroaniline	40.000	43.510	-8.8	105	0.00
63	Azobenzene	40.000	40.565	-1.4	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.249	-8.1	101	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.803	0.5	100	0.00
67	4-Bromophenyl-phenylether	40.000	40.792	-2.0	102	0.00
68	Hexachlorobenzene	40.000	39.946	0.1	100	0.00
69	Atrazine	40.000	38.714	3.2	101	0.00
70 C	Pentachlorophenol	40.000	41.095	-2.7	100	0.00
71	Phenanthrene	40.000	39.330	1.7	101	0.00
72	Anthracene	40.000	39.957	0.1	102	0.00
73	Carbazole	40.000	39.650	0.9	102	0.00
74	Di-n-butylphthalate	40.000	40.457	-1.1	101	0.00
75 C	Fluoranthene	40.000	39.650	0.9	103	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	106	0.00
77	Benzidine	40.000	40.998	-2.5	126	0.00
78	Pyrene	40.000	40.157	-0.4	103	0.00
79 S	Terphenyl-d14	80.000	79.096	1.1	103	0.00
80	Butylbenzylphthalate	40.000	41.928	-4.8	105	0.00
81	Benzo(a)anthracene	40.000	40.671	-1.7	109	0.00
82	3,3'-Dichlorobenzidine	40.000	39.969	0.1	110	0.00
83	Chrysene	40.000	39.560	1.1	106	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.224	-3.1	106	0.00
85 c	Di-n-octyl phthalate	40.000	42.704	-6.8	109	0.00
86 I	Perylene-d12	20.000	20.000	0.0	107	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.284	-5.7	104	0.00
88	Benzo(b)fluoranthene	40.000	40.465	-1.2	109	0.00
89	Benzo(k)fluoranthene	40.000	40.295	-0.7	107	0.00
90 C	Benzo(a)pyrene	40.000	40.663	-1.7	105	0.00
91	Dibenzo(a,h)anthracene	40.000	42.685	-6.7	104	0.00
92	Benzo(g,h,i)perylene	40.000	42.062	-5.2	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040622\
Data File : BF127647.D
Acq On : 06 Apr 2022 14:42
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF040622

Quant Time: Apr 06 16:42:02 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 06 13:55:27 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0