

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040622\
 Data File : BF127650.D
 Acq On : 06 Apr 2022 16:10
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 06 17:05:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 06 13:55:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.584	0.585	-0.2	104	0.00
3	Pyridine	1.538	1.551	-0.8	105	0.00
4	n-Nitrosodimethylamine	0.827	0.828	-0.1	103	0.00
5 S	2-Fluorophenol	1.231	1.191	3.2	103	0.00
6	Aniline	1.920	1.954	-1.8	104	0.00
7 S	Phenol-d6	1.610	1.586	1.5	104	0.00
8	2-Chlorophenol	1.249	1.229	1.6	105	0.00
9	Benzaldehyde	0.910	0.837	8.0	111	0.00
10 C	Phenol	1.610	1.603	0.4	105	0.00
11	bis(2-Chloroethyl)ether	1.330	1.320	0.8	105	0.00
12	1,3-Dichlorobenzene	1.427	1.396	2.2	104	0.00
13 C	1,4-Dichlorobenzene	1.442	1.391	3.5	104	0.00
14	1,2-Dichlorobenzene	1.379	1.286	6.7	103	0.00
15	Benzyl Alcohol	1.408	1.400	0.6	102	0.00
16	2,2'-oxybis(1-Chloropropane	2.099	2.101	-0.1	104	0.00
17	2-Methylphenol	1.086	1.087	-0.1	105	0.00
18	Hexachloroethane	0.550	0.540	1.8	102	0.00
19 P	n-Nitroso-di-n-propylamine	1.041	1.043	-0.2	104	0.00
20	3+4-Methylphenols	1.384	1.368	1.2	104	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
22	Acetophenone	0.537	0.523	2.6	105	0.00
23 S	Nitrobenzene-d5	0.449	0.458	-2.0	105	0.00
24	Nitrobenzene	0.435	0.440	-1.1	105	0.00
25	Isophorone	0.783	0.786	-0.4	104	0.00
26 C	2-Nitrophenol	0.156	0.170	-9.0	107	0.00
27	2,4-Dimethylphenol	0.306	0.301	1.6	106	0.00
28	bis(2-Chloroethoxy)methane	0.483	0.479	0.8	104	0.00
29 C	2,4-Dichlorophenol	0.330	0.327	0.9	105	0.00
30	1,2,4-Trichlorobenzene	0.379	0.372	1.8	103	0.00
31	Naphthalene	1.048	1.019	2.8	104	0.00
32	Benzoic acid	0.222	0.239	-7.7	104	0.00
33	4-Chloroaniline	0.413	0.406	1.7	104	0.00
34 C	Hexachlorobutadiene	0.261	0.256	1.9	104	0.00
35	Caprolactam	0.097	0.100	-3.1	105	0.00
36 C	4-Chloro-3-methylphenol	0.346	0.347	-0.3	104	0.00
37	2-Methylnaphthalene	0.701	0.689	1.7	104	0.00
38	1-Methylnaphthalene	0.677	0.670	1.0	105	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
40	1,2,4,5-Tetrachlorobenzene	0.656	0.638	2.7	104	0.00
41 P	Hexachlorocyclopentadiene	0.395	0.395	0.0	102	0.00
42 S	2,4,6-Tribromophenol	0.213	0.217	-1.9	110	0.00
43 C	2,4,6-Trichlorophenol	0.445	0.432	2.9	103	0.00
44	2,4,5-Trichlorophenol	0.449	0.454	-1.1	106	0.00
45 S	2-Fluorobiphenyl	1.414	1.355	4.2	106	0.00
46	1,1'-Biphenyl	1.530	1.488	2.7	105	0.00
47	2-Chloronaphthalene	1.216	1.182	2.8	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040622\
 Data File : BF127650.D
 Acq On : 06 Apr 2022 16:10
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 06 17:05:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 06 13:55:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.378	0.404	-6.9	106	0.00
49	Acenaphthylene	1.849	1.808	2.2	105	0.00
50	Dimethylphthalate	1.450	1.429	1.4	105	0.00
51	2,6-Dinitrotoluene	0.273	0.286	-4.8	107	0.00
52 C	Acenaphthene	1.165	1.133	2.7	104	0.00
53	3-Nitroaniline	0.289	0.303	-4.8	108	0.00
54 P	2,4-Dinitrophenol	0.119	0.123	-3.4	103	0.00
55	Dibenzofuran	1.700	1.662	2.2	105	0.00
56 P	4-Nitrophenol	0.227	0.239	-5.3	108	0.00
57	2,4-Dinitrotoluene	0.339	0.368	-8.6	110	0.00
58	Fluorene	1.256	1.215	3.3	106	0.00
59	2,3,4,6-Tetrachlorophenol	0.392	0.389	0.8	106	0.00
60	Diethylphthalate	1.387	1.376	0.8	106	0.00
61	4-Chlorophenyl-phenylether	0.690	0.677	1.9	106	0.00
62	4-Nitroaniline	0.272	0.287	-5.5	109	0.00
63	Azobenzene	1.550	1.532	1.2	105	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	0.097	0.104	-7.2	106	0.00
66 c	n-Nitrosodiphenylamine	0.630	0.638	-1.3	107	0.00
67	4-Bromophenyl-phenylether	0.243	0.253	-4.1	110	0.00
68	Hexachlorobenzene	0.265	0.267	-0.8	105	0.00
69	Atrazine	0.209	0.201	3.8	105	0.00
70 C	Pentachlorophenol	0.158	0.166	-5.1	107	0.00
71	Phenanthrene	1.060	1.041	1.8	105	0.00
72	Anthracene	1.048	1.035	1.2	105	0.00
73	Carbazole	0.982	0.966	1.6	106	0.00
74	Di-n-butylphthalate	1.174	1.180	-0.5	106	0.00
75 C	Fluoranthene	1.156	1.121	3.0	106	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
77	Benzydine	0.522	0.565	-8.2	130	0.00
78	Pyrene	1.689	1.755	-3.9	104	0.00
79 S	Terphenyl-d14	1.212	1.254	-3.5	106	0.00
80	Butylbenzylphthalate	0.629	0.684	-8.7	107	0.00
81	Benzo(a)anthracene	1.336	1.367	-2.3	108	0.00
82	3,3'-Dichlorobenzidine	0.416	0.431	-3.6	112	0.00
83	Chrysene	1.282	1.294	-0.9	106	0.00
84	Bis(2-ethylhexyl)phthalate	0.886	0.936	-5.6	106	0.00
85 c	Di-n-octyl phthalate	1.296	1.400	-8.0	108	0.00
86 I	Perylene-d12	1.000	1.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	1.250	1.371	-9.7	105	0.00
88	Benzo(b)fluoranthene	1.287	1.303	-1.2	106	0.00
89	Benzo(k)fluoranthene	1.276	1.285	-0.7	104	0.00
90 C	Benzo(a)pyrene	1.046	1.082	-3.4	104	0.00
91	Dibenzo(a,h)anthracene	1.012	1.124	-11.1	106	0.00
92	Benzo(g,h,i)perylene	1.038	1.138	-9.6	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040622\
Data File : BF127650.D
Acq On : 06 Apr 2022 16:10
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 06 17:05:57 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 06 13:55:27 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0