

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040622\
 Data File : BF127650.D
 Acq On : 06 Apr 2022 16:10
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 06 17:05:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 06 13:55:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	40.107	-0.3	104	0.00
3	Pyridine	40.000	40.335	-0.8	105	0.00
4	n-Nitrosodimethylamine	40.000	40.047	-0.1	103	0.00
5 S	2-Fluorophenol	80.000	77.432	3.2	103	0.00
6	Aniline	40.000	40.692	-1.7	104	0.00
7 S	Phenol-d6	80.000	78.793	1.5	104	0.00
8	2-Chlorophenol	40.000	39.370	1.6	105	0.00
9	Benzaldehyde	40.000	41.673	-4.2	111	0.00
10 C	Phenol	40.000	39.836	0.4	105	0.00
11	bis(2-Chloroethyl)ether	40.000	39.713	0.7	105	0.00
12	1,3-Dichlorobenzene	40.000	39.139	2.2	104	0.00
13 C	1,4-Dichlorobenzene	40.000	38.588	3.5	104	0.00
14	1,2-Dichlorobenzene	40.000	37.311	6.7	103	0.00
15	Benzyl Alcohol	40.000	39.778	0.6	102	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.040	-0.1	104	0.00
17	2-Methylphenol	40.000	40.017	-0.0	105	0.00
18	Hexachloroethane	40.000	39.276	1.8	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.073	-0.2	104	0.00
20	3+4-Methylphenols	40.000	39.550	1.1	104	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	105	0.00
22	Acetophenone	40.000	38.900	2.8	105	0.00
23 S	Nitrobenzene-d5	80.000	81.610	-2.0	105	0.00
24	Nitrobenzene	40.000	40.492	-1.2	105	0.00
25	Isophorone	40.000	40.158	-0.4	104	0.00
26 C	2-Nitrophenol	40.000	43.609	-9.0	107	0.00
27	2,4-Dimethylphenol	40.000	39.290	1.8	106	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.661	0.8	104	0.00
29 C	2,4-Dichlorophenol	40.000	39.692	0.8	105	0.00
30	1,2,4-Trichlorobenzene	40.000	39.236	1.9	103	0.00
31	Naphthalene	40.000	38.876	2.8	104	0.00
32	Benzoic acid	40.000	42.991	-7.5	104	0.00
33	4-Chloroaniline	40.000	39.362	1.6	104	0.00
34 C	Hexachlorobutadiene	40.000	39.243	1.9	104	0.00
35	Caprolactam	40.000	41.103	-2.8	105	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.200	-0.5	104	0.00
37	2-Methylnaphthalene	40.000	39.342	1.6	104	0.00
38	1-Methylnaphthalene	40.000	39.622	0.9	105	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	107	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.922	2.7	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.028	-0.1	102	0.00
42 S	2,4,6-Tribromophenol	80.000	81.564	-2.0	110	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.815	3.0	103	0.00
44	2,4,5-Trichlorophenol	40.000	40.431	-1.1	106	0.00
45 S	2-Fluorobiphenyl	80.000	76.675	4.2	106	0.00
46	1,1'-Biphenyl	40.000	38.919	2.7	105	0.00
47	2-Chloronaphthalene	40.000	38.906	2.7	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040622\
 Data File : BF127650.D
 Acq On : 06 Apr 2022 16:10
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 06 17:05:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 06 13:55:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.808	-7.0	106	0.00
49	Acenaphthylene	40.000	39.126	2.2	105	0.00
50	Dimethylphthalate	40.000	39.416	1.5	105	0.00
51	2,6-Dinitrotoluene	40.000	41.859	-4.6	107	0.00
52 C	Acenaphthene	40.000	38.896	2.8	104	0.00
53	3-Nitroaniline	40.000	41.981	-5.0	108	0.00
54 P	2,4-Dinitrophenol	40.000	39.170	2.1	103	0.00
55	Dibenzofuran	40.000	39.112	2.2	105	0.00
56 P	4-Nitrophenol	40.000	42.210	-5.5	108	0.00
57	2,4-Dinitrotoluene	40.000	43.407	-8.5	110	0.00
58	Fluorene	40.000	38.695	3.3	106	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.707	0.7	106	0.00
60	Diethylphthalate	40.000	39.676	0.8	106	0.00
61	4-Chlorophenyl-phenylether	40.000	39.271	1.8	106	0.00
62	4-Nitroaniline	40.000	42.170	-5.4	109	0.00
63	Azobenzene	40.000	39.526	1.2	105	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.192	-8.0	106	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.507	-1.3	107	0.00
67	4-Bromophenyl-phenylether	40.000	41.735	-4.3	110	0.00
68	Hexachlorobenzene	40.000	40.186	-0.5	105	0.00
69	Atrazine	40.000	38.353	4.1	105	0.00
70 C	Pentachlorophenol	40.000	41.842	-4.6	107	0.00
71	Phenanthrene	40.000	39.290	1.8	105	0.00
72	Anthracene	40.000	39.512	1.2	105	0.00
73	Carbazole	40.000	39.329	1.7	106	0.00
74	Di-n-butylphthalate	40.000	40.199	-0.5	106	0.00
75 C	Fluoranthene	40.000	38.783	3.0	106	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzydine	40.000	43.327	-8.3	130	0.00
78	Pyrene	40.000	41.562	-3.9	104	0.00
79 S	Terphenyl-d14	80.000	82.792	-3.5	106	0.00
80	Butylbenzylphthalate	40.000	43.533	-8.8	107	0.00
81	Benzo(a)anthracene	40.000	40.950	-2.4	108	0.00
82	3,3'-Dichlorobenzidine	40.000	41.434	-3.6	112	0.00
83	Chrysene	40.000	40.363	-0.9	106	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.231	-5.6	106	0.00
85 c	Di-n-octyl phthalate	40.000	43.212	-8.0	108	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.872	-9.7	105	0.00
88	Benzo(b)fluoranthene	40.000	40.514	-1.3	106	0.00
89	Benzo(k)fluoranthene	40.000	40.269	-0.7	104	0.00
90 C	Benzo(a)pyrene	40.000	41.403	-3.5	104	0.00
91	Dibenzo(a,h)anthracene	40.000	44.399	-11.0	106	0.00
92	Benzo(g,h,i)perylene	40.000	43.861	-9.7	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040622\
Data File : BF127650.D
Acq On : 06 Apr 2022 16:10
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 06 17:05:57 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 06 13:55:27 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0