

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040715\
 Data File : BF078329.D
 Acq On : 8 Apr 2015 8:27
 Operator : TP/IZ
 Sample : G1745-05DL 10X
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-2DDL

Quant Time: Apr 08 08:07:35 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 08 03:14:41 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.93	152	37175	20.00	ng	0.00
21) Naphthalene-d8	8.51	136	156943	20.00	ng	0.00
38) Acenaphthene-d10	10.67	164	77536	20.00	ng	0.00
63) Phenanthrene-d10	12.50	188	153315	20.00	ng	0.00
75) Chrysene-d12	15.77	240	163616	20.00	ng	0.00
86) Perylene-d12	17.47	264	160599	20.00	ng	0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.21	112	18917	8.39	ng	0.01
7) Phenol-d6	6.53	99	23437	8.29	ng	0.01
23) Nitrobenzene-d5	7.63	82	14133	5.46	ng	0.00
41) 2,4,6-Tribromophenol	11.65	330	4814	7.49	ng	0.00
44) 2-Fluorobiphenyl	9.86	172	28866	5.32	ng	0.00
78) Terphenyl-d14	14.49	244	32764	4.52	ng	-0.01
Target Compounds						Qvalue
31) Naphthalene	8.53	128	387135	47.39	ng	99
37) 2-Methylnaphthalene	9.39	142	61666	11.12	ng	97
45) 1,1'-Biphenyl	9.97	154	37224	5.87	ng	97
48) Acenaphthylene	10.50	152	20607	2.56	ng	# 96
51) Acenaphthene	10.72	154	60176	13.26	ng	99
57) Fluorene	11.35	166	42739	7.61	ng	100
70) Phenanthrene	12.53	178	245070	27.63	ng	100
71) Anthracene	12.59	178	63882	7.31	ng	100
74) Fluoranthene	14.00	202	111912	11.97	ng	92
77) Pyrene	14.27	202	171653	16.71	ng	99
80) Benzo(a)anthracene	15.76	228	43896m	4.51	ng	
82) Chrysene	15.79	228	28753	3.14	ng	98
87) Benzo(b)fluoranthene	17.04	252	29262m	2.95	ng	
89) Benzo(a)pyrene	17.40	252	37503	4.33	ng	# 98
91) Benzo(a,h,i)perylene	19.09	276	22962	2.64	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040715\
Data File : BF078329.D
Acq On : 8 Apr 2015 8:27
Operator : TP/IZ
Sample : G1745-05DL 10X
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-2DDL

Quant Time: Apr 08 08:07:35 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 08 03:14:41 2015
Response via : Initial Calibration

