

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040716\
 Data File : BF086149.D
 Acq On : 8 Apr 2016 00:23
 Operator : UM/SJ
 Sample : H2282-08
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 AMBOY-SB-16(0.5-20.0)

Manual Integrations
 APPROVED

Sohil
 4/8/2016 6:09:26 PM

Quant Time: Apr 08 03:47:33 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 01 23:17:06 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.74	152	112326	20.00	ng	-0.03
21) Naphthalene-d8	8.02	136	467270	20.00	ng	-0.05
38) Acenaphthene-d10	9.77	164	203035	20.00	ng	-0.05
63) Phenanthrene-d10	11.25	188	397302	20.00	ng	-0.05
75) Chrysene-d12	13.89	240	262010	20.00	ng	-0.03
86) Perylene-d12	15.29	264	210839	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	785440	119.52	ng	-0.03
7) Phenol-d6	6.38	99	1012799	121.34	ng	-0.03
23) Nitrobenzene-d5	7.31	82	673199	84.96	ng	-0.03
41) 2,4,6-Tribromophenol	10.57	330	273963	143.76	ng	-0.03
44) 2-Fluorobiphenyl	9.09	172	916966	75.09	ng	-0.05
78) Terphenyl-d14	12.84	244	941217	84.44	ng	-0.03
Target Compounds						Qvalue
49) Dimethylphthalate	9.49	163	157538	10.47	ng	99
74) Fluoranthene	12.46	202	77659	3.44	ng	95
77) Pyrene	12.69	202	68808	3.73	ng	98
80) Benzo(a)anthracene	13.88	228	42105	2.73	ng	97
87) Benzo(b)fluoranthene	14.90	252	35799m	2.70	ng	
89) Benzo(a)pyrene	15.23	252	28948	2.46	ng	# 93
91) Benzo(g,h,i)perylene	17.04	276	19694	2.15	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040716\
Data File : BF086149.D
Acq On : 8 Apr 2016 00:23
Operator : UM/SJ
Sample : H2282-08
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
AMBOY-SB-16(0.5-20.0)

Manual Integrations
APPROVED

Sohil
4/8/2016 6:09:26 PM

Quant Time: Apr 08 03:47:33 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 01 23:17:06 2016
Response via : Initial Calibration

