

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040717\
 Data File : BF094275.D
 Acq On : 7 Apr 2017 16:14
 Operator : SJ/MA
 Sample : I2572-04
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RT-2891

Manual Integrations
 APPROVED

mohammad
 4/10/2017 2:07:35 PM

Quant Time: Apr 07 16:53:12 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 07 14:38:56 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.85	152	176679	20.00	ng	0.00
21) Naphthalene-d8	7.35	136	705965	20.00	ng	0.00
38) Acenaphthene-d10	9.49	164	319191	20.00	ng	0.00
63) Phenanthrene-d10	11.31	188	535060	20.00	ng	0.00
75) Chrysene-d12	14.56	240	346197	20.00	ng	-0.01
86) Perylene-d12	16.19	264	283766	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	4.39	112	1434450	137.13	ng	0.02
7) Phenol-d6	5.46	99	1825523	141.07	ng	0.00
23) Nitrobenzene-d5	6.50	82	1146042	92.64	ng	0.00
41) 2,4,6-Tribromophenol	10.46	330	369285	146.41	ng	0.00
44) 2-Fluorobiphenyl	8.67	172	1683633	80.45	ng	0.00
78) Terphenyl-d14	13.29	244	1217842	78.85	ng	0.00
Target Compounds						
10) Phenol	5.47	94	133648	9.08	ng	84
49) Dimethylphthalate	9.17	163	244477	10.77	ng	98
70) Phenanthrene	11.33	178	90796	3.05	ng	97
74) Fluoranthene	12.80	202	113398	3.72	ng	98
77) Pyrene	13.07	202	112695	4.49	ng	98
82) Chrysene	14.59	228	39059	2.08	ng	95
87) Benzo(b)fluoranthene	15.78	252	39970m	2.30	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040717\
Data File : BF094275.D
Acq On : 7 Apr 2017 16:14
Operator : SJ/MA
Sample : I2572-04
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RT-2891

Manual Integrations
APPROVED

mohammad
4/10/2017 2:07:35 PM

Quant Time: Apr 07 16:53:12 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 07 14:38:56 2017
Response via : Initial Calibration

