

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040717\
 Data File : BF094288.D
 Acq On : 7 Apr 2017 22:20
 Operator : SJ/MA
 Sample : I2534-11
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 GP-B-02(0-5)

Quant Time: Apr 08 00:05:26 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 06 13:20:29 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.86	152	147928	20.00	ng	0.00
21) Naphthalene-d8	7.36	136	559411	20.00	ng	0.00
38) Acenaphthene-d10	9.50	164	224729	20.00	ng	0.00
63) Phenanthrene-d10	11.32	188	327014	20.00	ng	0.00
75) Chrysene-d12	14.57	240	290528	20.00	ng	0.00
86) Perylene-d12	16.21	264	231209	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	4.43	112	1079727	123.28	ng	0.04
7) Phenol-d6	5.51	99	1418793	130.95	ng	0.05
23) Nitrobenzene-d5	6.51	82	882430	90.02	ng	0.00
41) 2,4,6-Tribromophenol	10.48	330	226264	127.41	ng	0.00
44) 2-Fluorobiphenyl	8.68	172	1435476	97.43	ng	0.00
78) Terphenyl-d14	13.30	244	848804	65.49	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.17	163	134585	8.42	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040717\
Data File : BF094288.D
Acq On : 7 Apr 2017 22:20
Operator : SJ/MA
Sample : I2534-11
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
GP-B-02(0-5)

Quant Time: Apr 08 00:05:26 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 06 13:20:29 2017
Response via : Initial Calibration

