

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040717\
 Data File : BF094297.D
 Acq On : 8 Apr 2017 3:55
 Operator : SJ/MA
 Sample : I2593-01
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TW-10

Quant Time: Apr 08 05:19:13 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 06 13:20:29 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.85	152	130645	20.00	ng	0.00
21) Naphthalene-d8	7.36	136	452810	20.00	ng	0.00
38) Acenaphthene-d10	9.50	164	177724	20.00	ng	0.00
63) Phenanthrene-d10	11.33	188	264880	20.00	ng	0.00
75) Chrysene-d12	14.57	240	261870	20.00	ng	0.00
86) Perylene-d12	16.20	264	210740	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	4.42	112	512636	66.27	ng	0.03
7) Phenol-d6	5.50	99	385464	40.28	ng	0.04
23) Nitrobenzene-d5	6.51	82	746123	94.04	ng	0.00
41) 2,4,6-Tribromophenol	10.49	330	198045	141.02	ng	0.01
44) 2-Fluorobiphenyl	8.69	172	1288004	110.54	ng	0.00
78) Terphenyl-d14	13.30	244	777339	66.54	ng	0.00
Target Compounds						
10) Phenol	5.51	94	24973	2.29	ng	Qvalue # 80
31) Naphthalene	7.39	128	2018978	92.73	ng	99
37) 2-Methylnaphthalene	8.24	142	1754539	120.89	ng	99
45) 1,1'-Biphenyl	8.80	154	134655	9.39	ng	97
51) Acenaphthene	9.54	154	25204	2.32	ng	89
54) Dibenzofuran	9.75	168	63041	4.26	ng	# 89
57) Fluorene	10.17	166	61678	5.02	ng	# 94
70) Phenanthrene	11.35	178	87935	5.97	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040717\
Data File : BF094297.D
Acq On : 8 Apr 2017 3:55
Operator : SJ/MA
Sample : I2593-01
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TW-10

Quant Time: Apr 08 05:19:13 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 06 13:20:29 2017
Response via : Initial Calibration

