

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040815\
 Data File : BF078355.D
 Acq On : 9 Apr 2015 1:19
 Operator : TP/IZ
 Sample : G1783-02
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FD-1

Quant Time: Apr 09 02:22:24 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 09 00:49:15 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.92	152	35840	20.00	ng	0.01
21) Naphthalene-d8	8.50	136	126556	20.00	ng	0.01
38) Acenaphthene-d10	10.67	164	40308	20.00	ng	0.02
63) Phenanthrene-d10	12.51	188	112623	20.00	ng	0.03
75) Chrysene-d12	15.77	240	144872	20.00	ng	0.00
86) Perylene-d12	17.51	264	140522	20.00	ng	-0.07
System Monitoring Compounds						
5) 2-Fluorophenol	5.20	112	167042	76.85	ng	0.02
7) Phenol-d6	6.52	99	226109	83.00	ng	0.01
23) Nitrobenzene-d5	7.62	82	209719	100.44	ng	0.01
41) 2,4,6-Tribromophenol	11.65	330	34725	103.87	ng	0.02
44) 2-Fluorobiphenyl	9.86	172	264282	93.72	ng	0.02
78) Terphenyl-d14	14.50	244	315901	49.27	ng	0.02
Target Compounds						Qvalue
31) Naphthalene	8.52	128	490916	74.53	ng	98
37) 2-Methylnaphthalene	9.40	142	958211	214.26	ng	96
45) 1,1'-Biphenyl	9.97	154	167820	50.90	ng	96
51) Acenaphthene	10.72	154	75812	32.14	ng	# 93
57) Fluorene	11.36	166	154349	52.86	ng	# 85
70) Phenanthrene	12.55	178	412967	63.39	ng	96
77) Pyrene	14.28	202	127624	14.03	ng	# 92
83) Bis(2-ethylhexyl)phthalate	15.84	149	125240	23.04	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040815\
Data File : BF078355.D
Acq On : 9 Apr 2015 1:19
Operator : TP/IZ
Sample : G1783-02
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FD-1

Quant Time: Apr 09 02:22:24 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 09 00:49:15 2015
Response via : Initial Calibration

