

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040822\
 Data File : BF127706.D
 Acq On : 08 Apr 2022 10:33
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 11 02:20:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 06 13:55:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	107	0.00
2	1,4-Dioxane	40.000	39.079	2.3	104	-0.02
3	Pyridine	40.000	38.904	2.7	104	-0.01
4	n-Nitrosodimethylamine	40.000	36.255	9.4	96	-0.01
5 S	2-Fluorophenol	80.000	76.943	3.8	106	0.00
6	Aniline	40.000	39.636	0.9	104	0.00
7 S	Phenol-d6	80.000	77.756	2.8	106	0.00
8	2-Chlorophenol	40.000	39.348	1.6	108	0.00
9	Benzaldehyde	40.000	40.340	-0.9	112	0.00
10 C	Phenol	40.000	38.657	3.4	105	0.00
11	bis(2-Chloroethyl)ether	40.000	38.354	4.1	104	0.00
12	1,3-Dichlorobenzene	40.000	38.993	2.5	107	0.00
13 C	1,4-Dichlorobenzene	40.000	38.987	2.5	108	0.00
14	1,2-Dichlorobenzene	40.000	37.735	5.7	107	0.00
15	Benzyl Alcohol	40.000	37.046	7.4	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.780	8.0	99	0.00
17	2-Methylphenol	40.000	39.245	1.9	106	0.00
18	Hexachloroethane	40.000	37.485	6.3	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.907	7.7	99	0.00
20	3+4-Methylphenols	40.000	38.776	3.1	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
22	Acetophenone	40.000	37.637	5.9	102	0.00
23 S	Nitrobenzene-d5	80.000	75.806	5.2	98	0.00
24	Nitrobenzene	40.000	38.117	4.7	99	0.00
25	Isophorone	40.000	38.225	4.4	100	0.00
26 C	2-Nitrophenol	40.000	40.807	-2.0	101	0.00
27	2,4-Dimethylphenol	40.000	38.992	2.5	106	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.372	4.1	101	0.00
29 C	2,4-Dichlorophenol	40.000	40.414	-1.0	107	0.00
30	1,2,4-Trichlorobenzene	40.000	40.060	-0.2	106	0.00
31	Naphthalene	40.000	38.940	2.7	105	0.00
32	Benzoic acid	40.000	37.647	5.9	92	0.00
33	4-Chloroaniline	40.000	40.206	-0.5	107	0.00
34 C	Hexachlorobutadiene	40.000	40.067	-0.2	107	0.00
35	Caprolactam	40.000	39.860	0.4	103	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.415	1.5	103	0.00
37	2-Methylnaphthalene	40.000	39.114	2.2	104	0.00
38	1-Methylnaphthalene	40.000	39.313	1.7	105	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	107	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.169	-0.4	108	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.989	2.5	100	0.00
42 S	2,4,6-Tribromophenol	80.000	87.333	-9.2	119	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.223	1.9	104	0.00
44	2,4,5-Trichlorophenol	40.000	41.156	-2.9	109	0.00
45 S	2-Fluorobiphenyl	80.000	75.808	5.2	105	0.00
46	1,1'-Biphenyl	40.000	39.269	1.8	107	0.00
47	2-Chloronaphthalene	40.000	39.305	1.7	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040822\
 Data File : BF127706.D
 Acq On : 08 Apr 2022 10:33
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 11 02:20:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 06 13:55:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.116	2.2	98	0.00
49	Acenaphthylene	40.000	39.225	1.9	105	0.00
50	Dimethylphthalate	40.000	38.793	3.0	104	0.00
51	2,6-Dinitrotoluene	40.000	39.111	2.2	100	0.00
52 C	Acenaphthene	40.000	38.657	3.4	103	0.00
53	3-Nitroaniline	40.000	40.591	-1.5	105	0.00
54 P	2,4-Dinitrophenol	40.000	39.474	1.3	105	0.00
55	Dibenzofuran	40.000	39.385	1.5	107	0.00
56 P	4-Nitrophenol	40.000	38.605	3.5	99	0.00
57	2,4-Dinitrotoluene	40.000	40.635	-1.6	104	0.00
58	Fluorene	40.000	39.186	2.0	108	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.583	-1.5	109	0.00
60	Diethylphthalate	40.000	38.380	4.0	103	0.00
61	4-Chlorophenyl-phenylether	40.000	39.138	2.2	106	0.00
62	4-Nitroaniline	40.000	43.685	-9.2	113	0.00
63	Azobenzene	40.000	37.449	6.4	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	111	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.283	-3.2	106	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.779	3.1	107	0.00
67	4-Bromophenyl-phenylether	40.000	40.940	-2.3	113	0.00
68	Hexachlorobenzene	40.000	41.462	-3.7	113	0.00
69	Atrazine	40.000	37.747	5.6	109	0.00
70 C	Pentachlorophenol	40.000	40.520	-1.3	109	0.00
71	Phenanthrene	40.000	38.306	4.2	108	0.00
72	Anthracene	40.000	39.276	1.8	110	0.00
73	Carbazole	40.000	39.292	1.8	111	0.00
74	Di-n-butylphthalate	40.000	37.962	5.1	104	0.00
75 C	Fluoranthene	40.000	38.939	2.7	111	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	119	0.00
77	Benzydine	40.000	38.233	4.4	131	0.00
78	Pyrene	40.000	38.597	3.5	110	0.00
79 S	Terphenyl-d14	80.000	77.030	3.7	112	0.00
80	Butylbenzylphthalate	40.000	38.541	3.6	108	0.00
81	Benzo(a)anthracene	40.000	39.938	0.2	120	0.00
82	3,3'-Dichlorobenzidine	40.000	43.155	-7.9	133	0.00
83	Chrysene	40.000	39.953	0.1	120	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	36.371	9.1	104	0.00
85 c	Di-n-octyl phthalate	40.000	37.798	5.5	108	0.00
86 I	Perylene-d12	20.000	20.000	0.0	127	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	44.308	-10.8	129	0.00
88	Benzo(b)fluoranthene	40.000	38.974	2.6	125	0.00
89	Benzo(k)fluoranthene	40.000	39.950	0.1	126	0.00
90 C	Benzo(a)pyrene	40.000	40.295	-0.7	124	0.00
91	Dibenzo(a,h)anthracene	40.000	45.092	-12.7	131	0.00
92	Benzo(g,h,i)perylene	40.000	44.427	-11.1	130	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040822\
Data File : BF127706.D
Acq On : 08 Apr 2022 10:33
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 11 02:20:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 06 13:55:27 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0