

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040915\
 Data File : BF078389.D
 Acq On : 10 Apr 2015 7:48
 Operator : TP/IZ
 Sample : G1782-22 5X
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FD03

Manual Integrations
 APPROVED

apatel
 4/10/2015 5:13:31 PM

Quant Time: Apr 10 07:28:54 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 10 01:53:29 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	34283	20.00	ng	0.00
21) Naphthalene-d8	8.46	136	142383	20.00	ng	0.00
38) Acenaphthene-d10	10.63	164	71480	20.00	ng	0.00
63) Phenanthrene-d10	12.45	188	136848	20.00	ng	0.01
75) Chrysene-d12	15.72	240	140104	20.00	ng	0.01
86) Perylene-d12	17.46	264	119330	20.00	ng	0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.15	112	39481	18.99	ng	0.01
7) Phenol-d6	6.49	99	51027	19.58	ng	0.01
23) Nitrobenzene-d5	7.58	82	28507	12.14	ng	0.00
41) 2,4,6-Tribromophenol	11.60	330	11712	19.76	ng	0.00
44) 2-Fluorobiphenyl	9.81	172	61986	12.40	ng	0.00
78) Terphenyl-d14	14.44	244	71018	11.45	ng	0.00
Target Compounds						
20) 3+4-Methylphenols	7.45	107	45733	17.96	ng	95
32) Benzoic acid	8.22	122	29411m	31.06	ng	
83) Bis(2-ethylhexyl)phthalate	15.80	149	35955	6.84	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040915\
Data File : BF078389.D
Acq On : 10 Apr 2015 7:48
Operator : TP/IZ
Sample : G1782-22 5X
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FD03

Manual Integrations
APPROVED

apatel
4/10/2015 5:13:31 PM

Quant Time: Apr 10 07:28:54 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 10 01:53:29 2015
Response via : Initial Calibration

