

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040915\
 Data File : BF078394.D
 Acq On : 10 Apr 2015 10:11
 Operator : TP/IZ
 Sample : G1793-03DL 50X
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-3DADL

Manual Integrations
 APPROVED

apatel
 4/10/2015 5:13:54 PM

Quant Time: Apr 10 13:08:15 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 09 11:29:31 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	35339	20.00	ng	0.00
21) Naphthalene-d8	8.46	136	151123	20.00	ng	0.00
38) Acenaphthene-d10	10.62	164	74698	20.00	ng	-0.01
63) Phenanthrene-d10	12.45	188	141195	20.00	ng	-0.01
75) Chrysene-d12	15.71	240	148567	20.00	ng	-0.03
86) Perylene-d12	17.44	264	144907	20.00	ng	-0.02

System Monitoring Compounds						
5) 2-Fluorophenol	5.16	112	3136	1.46	ng	-0.01
7) Phenol-d6	6.50	99	4412	1.64	ng	0.02
23) Nitrobenzene-d5	7.58	82	2353	0.94	ng	0.00
41) 2,4,6-Tribromophenol	11.61	330	632	1.02	ng	-0.01
44) 2-Fluorobiphenyl	9.81	172	6707	1.28	ng	0.00
78) Terphenyl-d14	14.44	244	6981	1.06	ng	-0.02

Target Compounds				Qvalue		
31) Naphthalene	8.49	128	875008	111.24	ng	99
37) 2-Methylnaphthalene	9.34	142	148736	27.85	ng	98
45) 1,1'-Biphenyl	9.93	154	67997	11.13	ng	99
48) Acenaphthylene	10.45	152	49179	6.34	ng	98
51) Acenaphthene	10.66	154	149493	34.20	ng	99
57) Fluorene	11.30	166	108398	20.03	ng	99
70) Phenanthrene	12.48	178	530856	65.00	ng	99
71) Anthracene	12.55	178	114099	14.18	ng	99
74) Fluoranthene	13.94	202	234166	27.19	ng	97
77) Pyrene	14.21	202	328705	35.24	ng	97
80) Benzo(a)anthracene	15.70	228	103046m	11.66	ng	
82) Chrysene	15.75	228	67532	8.13	ng	98
85) Indeno(1,2,3-cd)pyrene	18.69	276	32578	3.46	ng	# 86
87) Benzo(b)fluoranthene	17.00	252	68247m	7.62	ng	
88) Benzo(k)fluoranthene	17.03	252	21441m	2.69	ng	
89) Benzo(a)pyrene	17.37	252	82173	10.51	ng	# 95
91) Benzo(a,h,i)perylene	19.05	276	41068m	5.23	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040915\
Data File : BF078394.D
Acq On : 10 Apr 2015 10:11
Operator : TP/IZ
Sample : G1793-03DL 50X
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-3DADL

Manual Integrations
APPROVED

apatel
4/10/2015 5:13:54 PM

Quant Time: Apr 10 13:08:15 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 09 11:29:31 2015
Response via : Initial Calibration

