

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041117\
 Data File : BF094383.D
 Acq On : 12 Apr 2017 8:52
 Operator : SJ/MA
 Sample : I2463-11RE
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-BB-8-9-ESWRE

Quant Time: Apr 12 11:03:32 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 06 13:20:29 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.83	152	135308	20.00	ng	-0.03
21) Naphthalene-d8	7.33	136	488719	20.00	ng	-0.03
38) Acenaphthene-d10	9.47	164	192338	20.00	ng	-0.03
63) Phenanthrene-d10	11.30	188	312289	20.00	ng	-0.02
75) Chrysene-d12	14.60	240	676713	20.00	ng	0.02
86) Perylene-d12	16.23	264	161196	20.00	ng	0.04
System Monitoring Compounds						
5) 2-Fluorophenol	4.38	112	876249	109.38	ng	0.00
7) Phenol-d6	5.47	99	1064176	107.38	ng	0.00
23) Nitrobenzene-d5	6.48	82	617779	72.14	ng	-0.03
41) 2,4,6-Tribromophenol	10.45	330	168619	110.94	ng	-0.02
44) 2-Fluorobiphenyl	8.66	172	1062743	84.28	ng	-0.03
78) Terphenyl-d14	13.31	244	544726	18.04	ng	0.01
Target Compounds						Qvalue
9) Benzaldehyde	5.32	77	14277	2.13	ng	95
10) Phenol	5.48	94	49382	4.38	ng	95
17) 2-Methylphenol	6.16	107	202627	26.87	ng	95
20) 3+4-Methylphenols	6.35	107	279987	30.66	ng	# 60
27) 2,4-Dimethylphenol	6.98	122	434485	62.60	ng	94
29) 2,4-Dichlorophenol	7.19	162	99952	15.40	ng	96
31) Naphthalene	7.36	128	666771	28.37	ng	# 87
32) Benzoic acid	7.13	122	513442	111.13	ng	# 20
37) 2-Methylnaphthalene	8.20	142	57951	3.70	ng	97
49) Dimethylphthalate	9.15	163	69500	5.08	ng	99
70) Phenanthrene	11.32	178	92387	5.32	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041117\
Data File : BF094383.D
Acq On : 12 Apr 2017 8:52
Operator : SJ/MA
Sample : I2463-11RE
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-BB-8-9-ESWRE

Quant Time: Apr 12 11:03:32 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 06 13:20:29 2017
Response via : Initial Calibration

