

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041122\
 Data File : BF127749.D
 Acq On : 11 Apr 2022 20:18
 Operator : CG\JU
 Sample : N2349-17 5X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DEL-8D

Quant Time: Apr 12 04:56:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 06 13:55:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.963	152	121076	20.000	ng	0.00
21) Naphthalene-d8	8.251	136	398982	20.000	ng	# 0.00
39) Acenaphthene-d10	10.010	164	194232	20.000	ng	0.00
64) Phenanthrene-d10	11.504	188	303132	20.000	ng	0.00
76) Chrysene-d12	14.180	240	191638	20.000	ng	# 0.02
86) Perylene-d12	15.745	264	146579	20.000	ng	0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.575	112	48614	6.524	ng	0.00
7) Phenol-d6	6.575	99	95784	9.826	ng	0.00
23) Nitrobenzene-d5	7.522	82	78512	8.773	ng	-0.01
42) 2,4,6-Tribromophenol	10.798	330	11286	5.460	ng	0.00
45) 2-Fluorobiphenyl	9.322	172	152133	11.080	ng	0.00
79) Terphenyl-d14	13.098	244	111443	9.599	ng	0.00
Target Compounds						
84) Bis(2-ethylhexyl)phtha...	14.151	149	33589	3.955	ng	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041122\
Data File : BF127749.D
Acq On : 11 Apr 2022 20:18
Operator : CG\JU
Sample : N2349-17 5X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DEL-8D

Quant Time: Apr 12 04:56:39 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 06 13:55:27 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.963 min Scan# 795
Delta R.T. 0.000 min
Lab File: BF127749.D
Acq: 11 Apr 2022 20:18




Instrument :
BNA_F
ClientSampleId :
DEL-8D

Tgt Ion:152 Resp: 121076
Ion Ratio Lower Upper
152 100
150 149.5 117.2 175.8
115 66.6 49.8 74.6



#5
2-Fluorophenol
Concen: 6.524 ng
RT: 5.575 min Scan# 559
Delta R.T. -0.006 min
Lab File: BF127749.D
Acq: 11 Apr 2022 20:18

Tgt Ion:112 Resp: 48614
Ion Ratio Lower Upper
112 100
64 64.0 58.4 87.6
63 38.2 30.1 45.1









#7
Phenol-d6
Concen: 9.826 ng
RT: 6.575 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF127749.D
Acq: 11 Apr 2022 20:18




Instrument :
BNA_F
ClientSampleId :
DEL-8D


Tgt Ion: 99 Resp: 95784
Ion Ratio Lower Upper
99 100
42 19.2 17.8 26.6
71 38.5 29.6 44.4



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.251 min Scan# 1014
Delta R.T. 0.000 min
Lab File: BF127749.D
Acq: 11 Apr 2022 20:18

Tgt Ion: 136 Resp: 398982
Ion Ratio Lower Upper
136 100
137 10.7 8.6 13.0
54 8.3 9.2 13.8#
68 6.0 5.4 8.0









#23
Nitrobenzene-d5
Concen: 8.773 ng
RT: 7.522 min Scan# 89
Delta R.T. -0.012 min
Lab File: BF127749.D
Acq: 11 Apr 2022 20:18




Instrument :
BNA_F
ClientSampleId :
DEL-8D


Tgt Ion: 82 Resp: 78512
Ion Ratio Lower Upper
82 100
128 36.1 30.6 45.8
54 50.3 48.2 72.2



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.010 min Scan# 1313
Delta R.T. -0.006 min
Lab File: BF127749.D
Acq: 11 Apr 2022 20:18

Tgt Ion:164 Resp: 194232
Ion Ratio Lower Upper
164 100
162 100.0 81.2 121.8
160 44.5 34.6 52.0





#42

2,4,6-Tribromophenol

Concen: 5.460 ng

RT: 10.798 min Scan# 1447

Delta R.T. -0.006 min

Lab File: BF127749.D

Acq: 11 Apr 2022 20:18

Instrument :

BNA_F

ClientSampleId :

DEL-8D





Tgt Ion:330 Resp: 11286

Ion Ratio Lower Upper

330 100

332 99.8 79.1 118.7

141 35.8 26.2 39.4



#45

2-Fluorobiphenyl

Concen: 11.080 ng

RT: 9.322 min Scan# 1196

Delta R.T. -0.006 min

Lab File: BF127749.D

Acq: 11 Apr 2022 20:18




Tgt Ion:172 Resp: 152133


Ion Ratio Lower Upper

172 100

171 34.1 28.4 42.6

170 22.5 18.6 27.8









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.504 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF127749.D
Acq: 11 Apr 2022 20:18




Instrument :
BNA_F
ClientSampleId :
DEL-8D


Tgt Ion:188 Resp: 303132
Ion Ratio Lower Upper
188 100
94 8.3 7.4 11.0
80 9.9 8.0 12.0



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.180 min Scan# 2022
Delta R.T. 0.024 min
Lab File: BF127749.D
Acq: 11 Apr 2022 20:18

Tgt Ion:240 Resp: 191638
Ion Ratio Lower Upper
240 100
120 14.3 7.1 10.7#
236 29.3 20.7 31.1









#79
 Terphenyl-d14
 Concen: 9.599 ng
 RT: 13.098 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: BF127749.D
 Acq: 11 Apr 2022 20:18




Instrument :
 BNA_F
 ClientSampleId :
 DEL-8D


Tgt Ion:244 Resp: 111443
 Ion Ratio Lower Upper
 244 100
 212 8.5 5.7 8.5
 122 14.3 6.2 9.2#



#84
 Bis(2-ethylhexyl)phthalate
 Concen: 3.955 ng
 RT: 14.151 min Scan# 2017
 Delta R.T. 0.024 min
 Lab File: BF127749.D
 Acq: 11 Apr 2022 20:18

Tgt Ion:149 Resp: 33589
 Ion Ratio Lower Upper
 149 100
 167 42.5 22.9 34.3#
 279 8.6 3.2 4.8#





#86
Perylene-d12
Concen: 20.000 ng
RT: 15.745 min Scan# 21
Delta R.T. 0.065 min
Lab File: BF127749.D
Acq: 11 Apr 2022 20:18

Instrument :
BNA_F
ClientSampleId :
DEL-8D

Tgt Ion:264 Resp: 146579

Ion	Ratio	Lower	Upper
264	100		
260	25.7	19.1	28.7
265	25.2	17.4	26.2

