

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041218\
 Data File : BF104457.D
 Acq On : 12 Apr 2018 18:02
 Operator : JU/SJ
 Sample : J2163-05 2X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PL-01-041118-C

Manual Integrations
 APPROVED

Sohil
 4/13/2018 4:44:13 PM

Quant Time: Apr 13 00:40:51 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 12 17:07:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	129874	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	500944	20.00	ng	0.00
38) Acenaphthene-d10	9.85	164	197629	20.00	ng	0.00
63) Phenanthrene-d10	11.34	188	307380	20.00	ng	0.00
75) Chrysene-d12	13.99	240	268835	20.00	ng	0.00
86) Perylene-d12	15.46	264	201722	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	433221	51.00	ng	0.00
7) Phenol-d6	6.43	99	503780	48.27	ng	-0.01
23) Nitrobenzene-d5	7.37	82	292818	38.17	ng	-0.01
41) 2,4,6-Tribromophenol	10.64	330	88925	43.38	ng	0.00
44) 2-Fluorobiphenyl	9.17	172	488759	39.07	ng	0.00
78) Terphenyl-d14	12.93	244	373614	31.68	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.56	163	32498	2.085	ng	98
70) Phenanthrene	11.36	178	64263	3.754	ng	97
74) Fluoranthene	12.55	202	80618	4.508	ng	98
77) Pyrene	12.78	202	91457	4.590	ng	98
80) Benzo(a)anthracene	13.97	228	41782	2.602	ng	95
82) Chrysene	14.01	228	42193	2.558	ng	96
87) Benzo(b)fluoranthene	15.03	252	31557m	2.477	ng	
89) Benzo(a)pyrene	15.39	252	22956	2.014	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041218\
Data File : BF104457.D
Acq On : 12 Apr 2018 18:02
Operator : JU/SJ
Sample : J2163-05 2X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleID :
PL-01-041118-C

Manual Integrations
APPROVED

Sohil
4/13/2018 4:44:13 PM

Quant Time: Apr 13 00:40:51 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 12 17:07:03 2018
Response via : Initial Calibration

