

Data Path : Z:\HPCHEM1\BNA F\Data\BF041218\
 Data File : BF104446.D
 Acq On : 12 Apr 2018 13:02
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 12 17:07:10 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 12 17:07:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.609	0.561	7.9	83	0.00
3	Pyridine	1.714	1.570	8.4	83	0.00
4	n-Nitrosodimethylamine	0.599	0.522	12.9	79	0.00
5 S	2-Fluorophenol	1.308	1.197	8.5	85	0.00
6	Aniline	2.233	2.032	9.0	83	0.00
7 S	Phenol-d6	1.607	1.463	9.0	85	0.00
8	2-Chlorophenol	1.381	1.305	5.5	87	0.00
9	Benzaldehyde	0.802	0.741	7.6	81	0.00
10 C	Phenol	1.705	1.684	1.2	92	0.00
11	bis(2-Chloroethyl)ether	1.361	1.238	9.0	83	0.00
12	1,3-Dichlorobenzene	1.564	1.485	5.1	88	0.00
13 C	1,4-Dichlorobenzene	1.559	1.495	4.1	88	0.00
14	1,2-Dichlorobenzene	1.445	1.376	4.8	87	0.00
15	Benzyl Alcohol	1.221	1.121	8.2	85	0.00
16	2,2'-oxybis(1-Chloropropane	1.827	1.627	10.9	82	0.00
17	2-Methylphenol	1.200	1.133	5.6	87	0.00
18	Hexachloroethane	0.556	0.538	3.2	89	0.00
19 P	n-Nitroso-di-n-propylamine	1.050	0.894	14.9	82	0.00
20	3+4-Methylphenols	1.488	1.353	9.1	84	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
22	Acetophenone	0.492	0.446	9.3	84	0.00
23 S	Nitrobenzene-d5	0.306	0.322	-5.2	94	0.00
24	Nitrobenzene	0.338	0.344	-1.8	89	0.00
25	Isophorone	0.648	0.590	9.0	84	0.00
26 C	2-Nitrophenol	0.138	0.173	-25.4#	109	0.00
27	2,4-Dimethylphenol	0.286	0.251	12.2	80	0.00
28	bis(2-Chloroethoxy)methane	0.396	0.366	7.6	86	0.00
29 C	2,4-Dichlorophenol	0.273	0.261	4.4	87	0.00
30	1,2,4-Trichlorobenzene	0.299	0.286	4.3	88	0.00
31	Naphthalene	0.996	0.931	6.5	86	0.00
32	Benzoic acid	0.228	0.213	6.6	80	0.00
33	4-Chloroaniline	0.427	0.404	5.4	87	0.00
34 C	Hexachlorobutadiene	0.164	0.160	2.4	90	0.00
35	Caprolactam	0.095	0.090	5.3	85	0.00
36 C	4-Chloro-3-methylphenol	0.292	0.282	3.4	88	0.00
37	2-Methylnaphthalene	0.644	0.605	6.1	87	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
39	1,2,4,5-Tetrachlorobenzene	0.573	0.558	2.6	92	0.00
40 P	Hexachlorocyclopentadiene	0.321	0.326	-1.6	93	0.00
41 S	2,4,6-Tribromophenol	0.207	0.212	-2.4	94	0.00
42 C	2,4,6-Trichlorophenol	0.397	0.387	2.5	88	0.00
43	2,4,5-Trichlorophenol	0.445	0.436	2.0	91	0.00
44 S	2-Fluorobiphenyl	1.266	1.192	5.8	89	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF041218\
 Data File : BF104446.D
 Acq On : 12 Apr 2018 13:02
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 12 17:07:10 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 12 17:07:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.680	1.564	6.9	88	0.00
46	2-Chloronaphthalene	1.327	1.255	5.4	89	0.00
47	2-Nitroaniline	0.328	0.389	-18.6	103	0.00
48	Acenaphthylene	2.082	1.952	6.2	87	0.00
49	Dimethylphthalate	1.577	1.533	2.8	92	0.00
50	2,6-Dinitrotoluene	0.292	0.331	-13.4	98	0.00
51 C	Acenaphthene	1.270	1.145	9.8	85	0.00
52	3-Nitroaniline	0.342	0.388	-13.5	98	0.00
53 P	2,4-Dinitrophenol	0.089	0.154	-73.0#	164#	0.00
54	Dibenzofuran	1.770	1.659	6.3	87	0.00
55 P	4-Nitrophenol	0.268	0.293	-9.3	94	0.00
56	2,4-Dinitrotoluene	0.383	0.440	-14.9	102	0.00
57	Fluorene	1.289	1.275	1.1	92	0.00
58	2,3,4,6-Tetrachlorophenol	0.332	0.329	0.9	92	0.00
59	Diethylphthalate	1.571	1.488	5.3	90	0.00
60	4-Chlorophenyl-phenylether	0.574	0.584	-1.7	94	0.00
61	4-Nitroaniline	0.334	0.392	-17.4	100	0.00
62	Azobenzene	1.501	1.372	8.6	86	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
64	4,6-Dinitro-2-methylphenol	0.087	0.128	-47.1#	134	0.00
65 c	n-Nitrosodiphenylamine	0.693	0.632	8.8	89	0.00
66	4-Bromophenyl-phenylether	0.213	0.201	5.6	92	0.00
67	Hexachlorobenzene	0.239	0.228	4.6	93	0.00
68	Atrazine	0.209	0.200	4.3	94	0.00
69 C	Pentachlorophenol	0.148	0.142	4.1	88	0.00
70	Phenanthrene	1.114	1.019	8.5	89	0.00
71	Anthracene	1.132	1.024	9.5	89	0.00
72	Carbazole	1.106	1.006	9.0	90	0.00
73	Di-n-butylphthalate	1.351	1.260	6.7	91	0.00
74 C	Fluoranthene	1.164	1.083	7.0	91	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
76	Benzidine	0.692	0.619	10.5	88	0.00
77	Pyrene	1.482	1.363	8.0	91	0.00
78 S	Terphenyl-d14	0.877	0.808	7.9	93	0.00
79	Butylbenzylphthalate	0.698	0.672	3.7	94	0.00
80	Benzo(a)anthracene	1.195	1.144	4.3	96	0.00
81	3,3'-Dichlorobenzidine	0.445	0.425	4.5	90	0.00
82	Chrysene	1.227	1.124	8.4	90	0.00
83	Bis(2-ethylhexyl)phthalate	0.898	0.903	-0.6	99	0.00
84 c	Di-n-octyl phthalate	1.501	1.389	7.5	88	0.00
85	Indeno(1,2,3-cd)pyrene	1.043	0.900	13.7	88	0.00
86 I	Perylene-d12	1.000	1.000	0.0	89	0.00
87	Benzo(b)fluoranthene	1.263	1.253	0.8	88	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF041218\
Data File : BF104446.D
Acq On : 12 Apr 2018 13:02
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 12 17:07:10 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 12 17:07:03 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.193	1.082	9.3	82	0.00
89 C	Benzo(a)pyrene	1.130	1.074	5.0	85	0.00
90	Dibenzo(a,h)anthracene	0.928	0.868	6.5	86	0.00
91	Benzo(a,h,i)perylene	0.899	0.847	5.8	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1