

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF041315\
 Data File : BF078501.D
 Acq On : 14 Apr 2015 12:46
 Operator : TP/IZ
 Sample : G1850-04
 Misc :
 ALS Vial : 40 Sample Multiplier: 0.5

Instrument :
 BNA_F
 ClientSampleId :
 SB-10DA

Manual Integrations
 APPROVED

apatel
 4/14/2015 5:27:20 PM

Quant Time: Apr 14 12:27:04 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 14 11:28:17 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.83	152	32785	10.00	ng	0.00
21) Naphthalene-d8	8.41	136	133316	10.00	ng	0.00
38) Acenaphthene-d10	10.56	164	66774	10.00	ng	-0.01
63) Phenanthrene-d10	12.39	188	131690	10.00	ng	0.00
75) Chrysene-d12	15.65	240	139641	10.00	ng	0.00
86) Perylene-d12	17.37	264	143591	10.00	ng	0.08
System Monitoring Compounds						
5) 2-Fluorophenol	5.10	112	549653	138.21	ng	0.02
7) Phenol-d6	6.43	99	716555	143.77	ng	0.00
23) Nitrobenzene-d5	7.53	82	390972	88.88	ng	0.00
41) 2,4,6-Tribromophenol	11.54	330	168973	152.56	ng	0.00
44) 2-Fluorobiphenyl	9.76	172	682234	73.02	ng	0.00
78) Terphenyl-d14	14.38	244	880039	71.20	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	8.43	128	717423	51.70	ng	99
37) 2-Methylnaphthalene	9.29	142	192345	20.41	ng	98
45) 1,1'-Biphenyl	9.87	154	95411	8.73	ng	98
48) Acenaphthylene	10.39	152	53614	3.86	ng	# 96
49) Dimethylphthalate	10.27	163	31055	3.10	ng	# 96
51) Acenaphthene	10.60	154	209173	26.77	ng	100
57) Fluorene	11.24	166	168404	17.41	ng	98
70) Phenanthrene	12.42	178	833151	54.69	ng	99
71) Anthracene	12.48	178	276440	18.41	ng	99
74) Fluoranthene	13.89	202	357552	22.25	ng	91
77) Pyrene	14.16	202	545284	31.10	ng	99
80) Benzo(a)anthracene	15.63	228	170019m	10.23	ng	
82) Chrysene	15.68	228	124195	7.95	ng	99
85) Indeno(1,2,3-cd)pyrene	18.62	276	47944m	2.71	ng	
87) Benzo(b)fluoranthene	16.94	252	103635m	5.84	ng	
88) Benzo(k)fluoranthene	16.96	252	39882m	2.53	ng	
89) Benzo(a)pyrene	17.31	252	133283	8.60	ng	97
91) Benzo(g,h,i)perylene	18.97	276	59233m	3.81	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF041315\
Data File : BF078501.D
Acq On : 14 Apr 2015 12:46
Operator : TP/IZ
Sample : G1850-04
Misc :
ALS Vial : 40 Sample Multiplier: 0.5

Instrument :
BNA_F
ClientSampleId :
SB-10DA

Manual Integrations
APPROVED

apatel
4/14/2015 5:27:20 PM

Quant Time: Apr 14 12:27:04 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 14 11:28:17 2015
Response via : Initial Calibration

