


Data Path : Z:\HPCHEM1\BNA_F\Data\BF041315\
 Data File : BF078459.D
 Acq On : 13 Apr 2015 12:37
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Mar 31 14:58:34 2015



AutoFind: Scans 1041, 1042, 1043; Background Corrected with Scan 1037

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.7	6702	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	34.9	6555	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.0	9008	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	18783	PASS
199	198	5	9	6.2	1167	PASS
275	198	10	60	25.7	4826	PASS
365	198	1	100	2.1	397	PASS
441	442	0.01	24	16.1	2553	PASS
442	198	50	100	84.5	15864	PASS
443	442	15	24	18.5	2929	PASS