

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041422\
 Data File : BF127791.D
 Acq On : 14 Apr 2022 11:13
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 15 05:05:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 13 16:40:15 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	115	0.00
2	1,4-Dioxane	40.000	39.846	0.4	115	0.00
3	Pyridine	40.000	39.953	0.1	116	0.00
4	n-Nitrosodimethylamine	40.000	34.875	12.8	100	0.00
5 S	2-Fluorophenol	80.000	80.708	-0.9	120	0.00
6	Aniline	40.000	40.253	-0.6	114	0.00
7 S	Phenol-d6	80.000	80.025	-0.0	118	0.00
8	2-Chlorophenol	40.000	41.211	-3.0	122	0.00
9	Benzaldehyde	40.000	41.514	-3.8	123	0.00
10 C	Phenol	40.000	40.525	-1.3	119	0.00
11	bis(2-Chloroethyl)ether	40.000	40.157	-0.4	118	0.00
12	1,3-Dichlorobenzene	40.000	40.810	-2.0	121	0.00
13 C	1,4-Dichlorobenzene	40.000	40.484	-1.2	121	0.00
14	1,2-Dichlorobenzene	40.000	39.223	1.9	120	0.00
15	Benzyl Alcohol	40.000	37.209	7.0	106	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.516	6.2	108	0.00
17	2-Methylphenol	40.000	40.207	-0.5	117	0.00
18	Hexachloroethane	40.000	40.487	-1.2	117	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.517	6.2	109	0.00
20	3+4-Methylphenols	40.000	39.409	1.5	115	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	117	0.00
22	Acetophenone	40.000	37.864	5.3	113	0.00
23 S	Nitrobenzene-d5	80.000	81.283	-1.6	116	0.00
24	Nitrobenzene	40.000	39.875	0.3	115	0.00
25	Isophorone	40.000	38.321	4.2	111	0.00
26 C	2-Nitrophenol	40.000	49.538	-23.8#	136	0.00
27	2,4-Dimethylphenol	40.000	38.895	2.8	116	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.176	2.1	114	0.00
29 C	2,4-Dichlorophenol	40.000	40.562	-1.4	119	0.00
30	1,2,4-Trichlorobenzene	40.000	39.991	0.0	117	0.00
31	Naphthalene	40.000	39.280	1.8	117	0.00
32	Benzoic acid	40.000	43.014	-7.5	116	0.00
33	4-Chloroaniline	40.000	40.661	-1.7	120	0.00
34 C	Hexachlorobutadiene	40.000	39.514	1.2	116	0.00
35	Caprolactam	40.000	40.482	-1.2	115	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.801	0.5	115	0.00
37	2-Methylnaphthalene	40.000	39.397	1.5	116	0.00
38	1-Methylnaphthalene	40.000	39.255	1.9	116	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	116	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.510	-3.8	121	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.520	-1.3	113	0.00
42 S	2,4,6-Tribromophenol	80.000	88.195	-10.2	130	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.716	-1.8	117	0.00
44	2,4,5-Trichlorophenol	40.000	42.485	-6.2	121	0.00
45 S	2-Fluorobiphenyl	80.000	76.410	4.5	114	0.00
46	1,1'-Biphenyl	40.000	39.799	0.5	117	0.00
47	2-Chloronaphthalene	40.000	39.843	0.4	117	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041422\
 Data File : BF127791.D
 Acq On : 14 Apr 2022 11:13
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 15 05:05:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 13 16:40:15 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.632	-11.6	121	0.00
49	Acenaphthylene	40.000	39.698	0.8	115	0.00
50	Dimethylphthalate	40.000	39.934	0.2	116	0.00
51	2,6-Dinitrotoluene	40.000	45.288	-13.2	126	0.00
52 C	Acenaphthene	40.000	40.524	-1.3	117	0.00
53	3-Nitroaniline	40.000	44.493	-11.2	125	0.00
54 P	2,4-Dinitrophenol	40.000	53.173	-32.9#	158	0.00
55	Dibenzofuran	40.000	39.990	0.0	117	0.00
56 P	4-Nitrophenol	40.000	44.662	-11.7	124	0.00
57	2,4-Dinitrotoluene	40.000	48.218	-20.5	133	0.00
58	Fluorene	40.000	40.587	-1.5	121	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.251	-5.6	123	0.00
60	Diethylphthalate	40.000	39.313	1.7	114	0.00
61	4-Chlorophenyl-phenylether	40.000	40.263	-0.7	118	0.00
62	4-Nitroaniline	40.000	46.136	-15.3	129	0.00
63	Azobenzene	40.000	37.905	5.2	110	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	120	0.00
65	4,6-Dinitro-2-methylphenol	40.000	56.321	-40.8#	155	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.524	1.2	118	0.00
67	4-Bromophenyl-phenylether	40.000	41.493	-3.7	123	0.00
68	Hexachlorobenzene	40.000	41.360	-3.4	122	0.00
69	Atrazine	40.000	39.890	0.3	123	0.00
70 C	Pentachlorophenol	40.000	41.539	-3.8	120	0.00
71	Phenanthrene	40.000	39.608	1.0	120	0.00
72	Anthracene	40.000	39.677	0.8	119	0.00
73	Carbazole	40.000	39.617	1.0	121	0.00
74	Di-n-butylphthalate	40.000	38.633	3.4	114	0.00
75 C	Fluoranthene	40.000	39.241	1.9	121	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	121	0.00
77	Benzydine	40.000	34.237	14.4	120	0.00
78	Pyrene	40.000	41.329	-3.3	121	0.00
79 S	Terphenyl-d14	80.000	79.431	0.7	118	0.00
80	Butylbenzylphthalate	40.000	41.851	-4.6	120	0.00
81	Benzo(a)anthracene	40.000	41.620	-4.0	127	0.00
82	3,3'-Dichlorobenzidine	40.000	39.948	0.1	125	0.00
83	Chrysene	40.000	39.805	0.5	122	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.644	3.4	113	0.00
85 c	Di-n-octyl phthalate	40.000	38.944	2.6	114	0.00
86 I	Perylene-d12	20.000	20.000	0.0	122	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	45.315	-13.3	127	0.00
88	Benzo(b)fluoranthene	40.000	40.115	-0.3	123	0.00
89	Benzo(k)fluoranthene	40.000	40.194	-0.5	121	0.00
90 C	Benzo(a)pyrene	40.000	41.071	-2.7	121	0.00
91	Dibenzo(a,h)anthracene	40.000	46.099	-15.2	128	0.00
92	Benzo(g,h,i)perylene	40.000	45.515	-13.8	127	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041422\
Data File : BF127791.D
Acq On : 14 Apr 2022 11:13
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 15 05:05:34 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 13 16:40:15 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1