

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041516\
 Data File : BF086223.D
 Acq On : 15 Apr 2016 18:22
 Operator : UM/SJ
 Sample : H2531-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DSN001

Quant Time: Apr 16 00:54:53 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 15 15:56:31 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	277605	20.00	ng	-0.01
21) Naphthalene-d8	8.18	136	1184979	20.00	ng	0.00
38) Acenaphthene-d10	9.93	164	498335	20.00	ng	-0.01
63) Phenanthrene-d10	11.41	188	909849	20.00	ng	0.00
75) Chrysene-d12	14.04	240	629913	20.00	ng	-0.01
86) Perylene-d12	15.48	264	463044	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.48	112	1057516	62.42	ng	0.00
7) Phenol-d6	6.51	99	981576	46.32	ng	-0.01
23) Nitrobenzene-d5	7.46	82	1779501	86.62	ng	0.00
41) 2,4,6-Tribromophenol	10.73	330	517992	129.45	ng	0.00
44) 2-Fluorobiphenyl	9.26	172	2587933	95.83	ng	0.00
78) Terphenyl-d14	13.00	244	1966805	78.51	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.65	163	182176	5.03	ng	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041516\
Data File : BF086223.D
Acq On : 15 Apr 2016 18:22
Operator : UM/SJ
Sample : H2531-01
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DSN001

Quant Time: Apr 16 00:54:53 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 15 15:56:31 2016
Response via : Initial Calibration

