

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041516\
 Data File : BF086264.D
 Acq On : 16 Apr 2016 14:25
 Operator : UM/SJ
 Sample : H2471-11
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CDMSH9

Quant Time: Apr 18 05:35:11 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 15 15:56:31 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	283856	20.00	ng	-0.01
21) Naphthalene-d8	8.18	136	1188404	20.00	ng	0.00
38) Acenaphthene-d10	9.93	164	506268	20.00	ng	-0.01
63) Phenanthrene-d10	11.41	188	888882	20.00	ng	0.00
75) Chrysene-d12	14.04	240	508282	20.00	ng	-0.01
86) Perylene-d12	15.48	264	487911	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	1618621	93.43	ng	0.01
7) Phenol-d6	6.52	99	2093900	96.63	ng	0.00
23) Nitrobenzene-d5	7.46	82	1355531	65.79	ng	0.00
41) 2,4,6-Tribromophenol	10.72	330	343467	84.49	ng	-0.01
44) 2-Fluorobiphenyl	9.25	172	1552315	56.58	ng	-0.01
78) Terphenyl-d14	13.00	244	1257127	62.19	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.65	163	380367	10.33	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041516\
Data File : BF086264.D
Acq On : 16 Apr 2016 14:25
Operator : UM/SJ
Sample : H2471-11
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CDMSH9

Quant Time: Apr 18 05:35:11 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 15 15:56:31 2016
Response via : Initial Calibration

