


Data Path : Z:\HPCHEM1\BNA F\DATA\BF041516\
 Data File : BF086244.D
 Acq On : 16 Apr 2016 5:08
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF041516.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Apr 15 15:56:31 2016



AutoFind: Scans 838, 839, 840; Background Corrected with Scan 835

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	47.0	39431	PASS
68	69	0.00	2	1.8	679	PASS
69	198	0.00	100	45.1	37819	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	57.9	48567	PASS
197	198	0.00	2	0.9	731	PASS
198	198	100	100	100.0	83832	PASS
199	198	5	9	6.9	5769	PASS
275	198	10	60	22.9	19204	PASS
365	198	1	100	2.8	2334	PASS
441	442	0.01	24	15.1	12494	PASS
442	198	50	100	98.4	82497	PASS
443	442	15	24	18.7	15392	PASS