

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041521\
 Data File : BF123562.D
 Acq On : 15 Apr 2021 11:17
 Operator : JU/CG
 Sample : SSTDICC005
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Quant Time: Apr 15 13:29:35 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 15 13:22:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.869	152	211994	20.00	ng	0.00
21) Naphthalene-d8	8.151	136	839615	20.00	ng	0.00
39) Acenaphthene-d10	9.916	164	436067	20.00	ng	0.00
64) Phenanthrene-d10	11.404	188	848845	20.00	ng	0.00
76) Chrysene-d12	14.057	240	1163344	20.00	ng	# 0.00
86) Perylene-d12	15.545	264	700286	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.475	112	139684	11.25	ng	0.00
7) Phenol-d6	6.481	99	195509	11.55	ng	-0.02
23) Nitrobenzene-d5	7.422	82	176065	10.47	ng	-0.01
42) 2,4,6-Tribromophenol	10.698	330	41498	9.75	ng	-0.01
45) 2-Fluorobiphenyl	9.233	172	320503	11.95	ng	0.00
79) Terphenyl-d14	12.992	244	700211	10.46	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.622	88	36475	3.61	ng	97
3) Pyridine	3.381	79	86151	2.96	ng	96
4) n-Nitrosodimethylamine	3.304	42	50835	4.52	ng	# 100
6) Aniline	6.522	93	113389	5.37	ng	99
8) 2-Chlorophenol	6.651	128	74017	5.56	ng	96
9) Benzaldehyde	6.410	77	68613	6.39	ng	90
10) Phenol	6.498	94	100313	5.58	ng	90
11) bis(2-Chloroethyl)ether	6.598	93	77942	5.45	ng	92
12) 1,3-Dichlorobenzene	6.810	146	80498	5.58	ng	96
13) 1,4-Dichlorobenzene	6.886	146	81536	5.59	ng	98
14) 1,2-Dichlorobenzene	7.039	146	77054	5.65	ng	95
15) Benzyl Alcohol	7.004	79	68196	5.23	ng	97
16) 2,2'-oxybis(1-Chloropr...	7.139	45	176802	6.52	ng	100
17) 2-Methylphenol	7.110	107	62912	5.57	ng	92
18) Hexachloroethane	7.386	117	29397	5.08	ng	96
19) n-Nitroso-di-n-propyla...	7.269	70	64853	5.76	ng	96
20) 3+4-Methylphenols	7.263	107	83317	6.03	ng	# 87
22) Acetophenone	7.269	105	119367	5.49	ng	# 98
24) Nitrobenzene	7.445	77	93956	5.17	ng	96
25) Isophorone	7.686	82	176286	5.21	ng	99
26) 2-Nitrophenol	7.763	139	22541	2.95	ng	# 87
27) 2,4-Dimethylphenol	7.798	122	60297	5.06	ng	90
28) bis(2-Chloroethoxy)met...	7.898	93	89505	4.94	ng	100
29) 2,4-Dichlorophenol	8.004	162	55496	4.39	ng	96
30) 1,2,4-Trichlorobenzene	8.092	180	65171	4.53	ng	99
31) Naphthalene	8.175	128	227495	5.46	ng	99
32) Benzoic acid	7.857	122	18448	2.34	ng	87
33) 4-Chloroaniline	8.216	127	91274	5.27	ng	97
34) Hexachlorobutadiene	8.298	225	40470	4.42	ng	98
35) Caprolactam	8.545	113	16650	4.09	ng	# 84
36) 4-Chloro-3-methylphenol	8.698	107	72558	5.24	ng	94
37) 2-Methylnaphthalene	8.869	142	146104	5.08	ng	100
38) 1-Methylnaphthalene	8.969	142	141189	5.22	ng	96
40) 1,2,4,5-Tetrachloroben...	9.033	216	62448	5.00	ng	98
41) Hexachlorocyclopentadiene	9.027	237	22657	3.61	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041521\
 Data File : BF123562.D
 Acq On : 15 Apr 2021 11:17
 Operator : JU/CG
 Sample : SSTDICC005
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005


Quant Time: Apr 15 13:29:35 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 15 13:22:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	9.145	196	34344	3.99	ng	94
44) 2,4,5-Trichlorophenol	9.180	196	40901	4.41	ng	92
46) 1,1'-Biphenyl	9.333	154	180011	6.04	ng	98
47) 2-Chloronaphthalene	9.357	162	134922	5.59	ng	97
48) 2-Nitroaniline	9.445	65	44074	5.19	ng	92
49) Acenaphthylene	9.774	152	219886	6.03	ng	98
50) Dimethylphthalate	9.627	163	156117	5.35	ng	97
51) 2,6-Dinitrotoluene	9.686	165	29517	4.49	ng	91
52) Acenaphthene	9.945	154	136970	5.84	ng	99
53) 3-Nitroaniline	9.857	138	36922	5.49	ng	81
54) 2,4-Dinitrophenol	9.963	184	6468	2.46	ng	92
55) Dibenzofuran	10.116	168	207583	6.06	ng	99
56) 4-Nitrophenol	10.010	139	23078	4.61	ng	91
57) 2,4-Dinitrotoluene	10.092	165	36938	4.34	ng	# 90
58) Fluorene	10.463	166	164057	6.26	ng	92
59) 2,3,4,6-Tetrachlorophenol	10.233	232	31010	4.32	ng	# 98
60) Diethylphthalate	10.333	149	167589	5.80	ng	98
61) 4-Chlorophenyl-phenyle...	10.457	204	74598	5.43	ng	90
62) 4-Nitroaniline	10.463	138	35854	5.69	ng	97
63) Azobenzene	10.616	77	190803	6.08	ng	99
65) 4,6-Dinitro-2-methylph...	10.498	198	11016	2.54	ng	82
66) n-Nitrosodiphenylamine	10.569	169	138178	5.67	ng	98
67) 4-Bromophenyl-phenylether	10.945	248	41988	4.83	ng	97
68) Hexachlorobenzene	11.016	284	49505	5.23	ng	# 92
69) Atrazine	11.092	200	42131	5.10	ng	96
70) Pentachlorophenol	11.204	266	16410	3.30	ng	97
71) Phenanthrene	11.427	178	236534	5.68	ng	97
72) Anthracene	11.480	178	235253	5.65	ng	98
73) Carbazole	11.627	167	231652	5.84	ng	98
74) Di-n-butylphthalate	11.968	149	261241	4.93	ng	96
75) Fluoranthene	12.621	202	261290	5.87	ng	97
77) Benzidine	12.739	184	146089	4.66	ng	98
78) Pyrene	12.851	202	468441	5.06	ng	95
80) Butylbenzylphthalate	13.474	149	167748	3.98	ng	94
81) Benzo(a)anthracene	14.045	228	366357	5.04	ng	98
82) 3,3'-Dichlorobenzidine	14.004	252	118497	5.11	ng	# 97
83) Chrysene	14.080	228	371799	5.11	ng	98
84) Bis(2-ethylhexyl)phtha...	14.039	149	229020	4.12	ng	# 98
85) Di-n-octyl phthalate	14.656	149	363032	4.49	ng	98
87) Indeno(1,2,3-cd)pyrene	17.027	276	207590	4.20	ng	97
88) Benzo(b)fluoranthene	15.104	252	352041	7.45	ng	# 97
89) Benzo(k)fluoranthene	15.133	252	359854	7.70	ng	# 97
90) Benzo(a)pyrene	15.474	252	187843	4.52	ng	97
91) Dibenzo(a,h)anthracene	17.050	278	174598	4.12	ng	95
92) Benzo(g,h,i)perylene	17.474	276	175392	4.47	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 15 13:29:35 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF041521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 15 13:22:17 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 6.869 min Scan# 813
 Delta R.T. -0.000 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17




Instrument :
 BNA_F
 ClientSampleID :
 SSTDICC005


Tgt Ion:	152	Resp:	211994
Ion	Ratio	Lower	Upper
152	100		
150	155.4	127.0	190.6
115	65.7	54.3	81.5



1,4-Dioxane
 Concen: 3.61 ng
 RT: 2.622 min Scan# 91
 Delta R.T. 0.006 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17

Tgt Ion:	88	Resp:	36475
Ion	Ratio	Lower	Upper
88	100		
58	101.6	83.8	125.8
43	39.5	32.2	48.4









Pyridine
 Concen: 2.96 ng
 RT: 3.381 min Scan# 220
 Delta R.T. 0.017 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005


Tgt Ion:	79	Resp:	86151
Ion	Ratio	Lower	Upper
79	100		
52	104.6	87.4	131.0
51	53.4	41.4	62.0



n-Nitrosodimethylamine
 Concen: 4.52 ng
 RT: 3.304 min Scan# 207
 Delta R.T. -0.018 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17

Tgt Ion:	42	Resp:	50835
Ion	Ratio	Lower	Upper
42	100		
74	104.4	83.5	125.3
44	8.7	5.7	8.5#









2-Fluorophenol
Concen: 11.25 ng
RT: 5.475 min Scan# 576
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17




Instrument :
BNA_F
ClientSampleId :
SSTDIC005


Tgt Ion:	112	Resp:	139684
Ion	Ratio	Lower	Upper
112	100		
64	76.2	61.4	92.0
63	40.8	33.0	49.4



Aniline
Concen: 5.37 ng
RT: 6.522 min Scan# 754
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	93	Resp:	113389
Ion	Ratio	Lower	Upper
93	100		
66	45.1	36.4	54.6
65	25.0	19.4	29.0









Phenol-d6
 Concen: 11.55 ng
 RT: 6.481 min Scan# 747
 Delta R.T. -0.018 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005


Tgt Ion:	99	Resp:	195509
Ion	Ratio	Lower	Upper
99	100		
42	24.4	19.8	29.8
71	40.2	31.1	46.7



2-Chlorophenol
 Concen: 5.56 ng
 RT: 6.651 min Scan# 776
 Delta R.T. -0.006 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17

Tgt Ion:	128	Resp:	74017
Ion	Ratio	Lower	Upper
128	100		
130	31.9	12.6	52.6
64	57.5	41.5	81.5









Benzaldehyde
Concen: 6.39 ng
RT: 6.410 min Scan# 735
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 77 Resp: 68613
Ion Ratio Lower Upper
77 100
105 85.5 75.2 115.2
106 82.7 71.6 111.6



Phenol
Concen: 5.58 ng
RT: 6.498 min Scan# 750
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion: 94 Resp: 100313
Ion Ratio Lower Upper
94 100
65 31.5 18.1 58.1
66 42.1 29.0 69.0









#1
 bis(2-Chloroethyl)ether
 Concen: 5.45 ng
 RT: 6.598 min Scan# 767
 Delta R.T. -0.006 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Tgt Ion:	93	Resp:	77942
Ion	Ratio	Lower	Upper
93	100		
63	87.7	76.1	116.1
95	28.7	11.5	51.5



#12
 1,3-Dichlorobenzene
 Concen: 5.58 ng
 RT: 6.810 min Scan# 803
 Delta R.T. -0.000 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17


Tgt Ion:	146	Resp:	80498
Ion	Ratio	Lower	Upper
146	100		
148	63.7	50.9	76.3
75	36.8	34.2	51.4





Abundance Scan 816 (6.887 min): BF123565.D\data.ms (-811) #13

1,4-Dichlorobenzene
Concen: 5.59 ng
RT: 6.886 min Scan# 816
Delta R.T. -0.000 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17


Instrument :
BNA_F
ClientSampleId :
SSTDICC005



Tgt Ion:	146	Resp:	81536
Ion	Ratio	Lower	Upper
146	100		
148	62.8	49.9	74.9
111	46.7	39.5	59.3





Abundance Scan 816 (6.886 min): BF123562.D\data.ms (-765) (-)




Abundance Scan 843 (7.045 min): BF123565.D\data.ms (-837) #14


1,2-Dichlorobenzene
Concen: 5.65 ng
RT: 7.039 min Scan# 842
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	146	Resp:	77054
Ion	Ratio	Lower	Upper
146	100		
148	60.0	51.6	77.4
111	45.5	38.4	57.6



Abundance Scan 842 (7.039 min): BF123562.D\data.ms (-792) (-)









Benzyl Alcohol
 Concen: 5.23 ng
 RT: 7.004 min Scan# 836
 Delta R.T. -0.006 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Tgt Ion:	79	Resp:	68196
Ion	Ratio	Lower	Upper
79	100		
108	66.6	55.8	83.8
77	61.4	49.9	74.9



2,2'-oxybis(1-Chloropropane)
 Concen: 6.52 ng
 RT: 7.139 min Scan# 859
 Delta R.T. -0.006 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17


Tgt Ion:	45	Resp:	176802
Ion	Ratio	Lower	Upper
45	100		
77	9.9	0.0	30.0
79	7.3	0.0	27.3






Abundance Scan 856 (7.122 min): BF123565.D\data.ms (-850) #17

2-Methylphenol
Concen: 5.57 ng
RT: 7.110 min Scan# 854
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampled :
SSTDIC005







Tgt Ion:	107	Resp:	62912
Ion	Ratio	Lower	Upper
107	100		
108	125.4	89.8	134.8
77	53.8	40.9	61.3
79	51.2	40.2	60.4




Abundance Scan 901 (7.387 min): BF123565.D\data.ms (-896) #18

Hexachloroethane
Concen: 5.08 ng
RT: 7.386 min Scan# 901
Delta R.T. -0.000 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17


Tgt Ion:	117	Resp:	29397
Ion	Ratio	Lower	Upper
117	100		
119	100.9	77.8	116.8
201	88.4	67.8	101.8








#19
n-Nitroso-di-n-propylamine
Concen: 5.76 ng
RT: 7.269 min Scan# 881
Delta R.T. -0.018 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17


Instrument :
BNA_F
ClientSampleId :
SSTDIC005





Tgt Ion:	70	Resp:	64853
Ion	Ratio	Lower	Upper
70	100		
42	79.3	60.1	90.1
101	9.2	7.4	11.2
130	17.0	14.2	21.2




#20
3+4-Methylphenols
Concen: 6.03 ng
RT: 7.263 min Scan# 880
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17



Tgt Ion:	107	Resp:	83317
Ion	Ratio	Lower	Upper
107	100		
108	83.9	71.5	111.5
77	81.4	82.0	122.0#
79	30.7	14.1	54.1






Abundance Scan 1032 (8.157 min): BF123565.D\data.ms (-1021)




Naphthalene-d8
Concen: 20.00 ng
RT: 8.151 min Scan# 1031
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005

Tgt Ion:	136	Resp:	839615
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	13.2	9.8	14.6
68	5.7	4.1	6.1






Abundance Scan 883 (7.281 min): BF123565.D\data.ms (-876) #2




Acetophenone
Concen: 5.49 ng
RT: 7.269 min Scan# 881
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	105	Resp:	119367
Ion	Ratio	Lower	Upper
105	100		
71	8.7	4.2	6.4#
51	42.4	33.6	50.4
120	19.5	16.3	24.5





Abundance Scan 909 (7.434 min): BF123565.D\data.ms (-904) #23




Nitrobenzene-d5
Concen: 10.47 ng
RT: 7.422 min Scan# 907
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005


Tgt Ion:	82	Resp:	176065
Ion	Ratio	Lower	Upper
82	100		
128	32.4	29.4	44.2
54	68.4	54.1	81.1



Abundance Scan 907 (7.422 min): BF123562.D\data.ms (-853) (-)





Abundance Scan 913 (7.457 min): BF123565.D\data.ms (-905) #24




Nitrobenzene
Concen: 5.17 ng
RT: 7.445 min Scan# 911
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	77	Resp:	93956
Ion	Ratio	Lower	Upper
77	100		
123	40.2	34.5	51.7
65	14.2	12.2	18.2




Abundance Scan 911 (7.445 min): BF123562.D\data.ms (-862) (-)






Abundance Scan 953 (7.692 min): BF123565.D\data.ms (-947) #25

Isophorone
Concen: 5.21 ng
RT: 7.686 min Scan# 952
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005







Tgt Ion:	82	Resp:	176286
Ion	Ratio	Lower	Upper
82	100		
95	7.4	5.8	8.6
138	15.0	11.8	17.8




Abundance Scan 966 (7.769 min): BF123565.D\data.ms (-961) #26

2-Nitrophenol
Concen: 2.95 ng
RT: 7.763 min Scan# 965
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	139	Resp:	22541
Ion	Ratio	Lower	Upper
139	100		
109	33.8	28.6	43.0
65	85.3	56.8	85.2#






Abundance Scan 972 (7.804 min): BF123565.D\data.ms (-968) #27




2,4-Dimethylphenol
Concen: 5.06 ng
RT: 7.798 min Scan# 971
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:	122	Resp:	60297
Ion	Ratio	Lower	Upper
122	100		
107	122.2	106.3	159.5
121	48.1	45.4	68.0






Abundance Scan 989 (7.904 min): BF123565.D\data.ms (-980) #28




bis(2-Chloroethoxy)methane
Concen: 4.94 ng
RT: 7.898 min Scan# 988
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	93	Resp:	89505
Ion	Ratio	Lower	Upper
93	100		
95	33.0	26.5	39.7
123	10.9	8.4	12.6






Abundance Scan 1007 (8.010 min): BF123565.D\data.ms (-10629)




2,4-Dichlorophenol
Concen: 4.39 ng
RT: 8.004 min Scan# 1006
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005

Tgt Ion:	162	Resp:	55496
Ion	Ratio	Lower	Upper
162	100		
164	61.5	44.1	84.1
98	43.9	27.3	67.3






Abundance Scan 1022 (8.098 min): BF123565.D\data.ms (-10130)




1,2,4-Trichlorobenzene
Concen: 4.53 ng
RT: 8.092 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	180	Resp:	65171
Ion	Ratio	Lower	Upper
180	100		
182	95.7	77.2	115.8
145	34.1	26.0	39.0







Abundance Scan 1036 (8.181 min): BF123565.D\data.ms (-10281-)



Naphthalene
Concen: 5.46 ng
RT: 8.175 min Scan# 1035
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17




Instrument :
BNA_F
ClientSampleId :
SSTDIC005

Tgt Ion:	128	Resp:	227495
Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.9	13.3
127	12.5	10.8	16.2




Benzoic acid
Concen: 2.34 ng
RT: 7.857 min Scan# 981
Delta R.T. -0.065 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	122	Resp:	18448
Ion	Ratio	Lower	Upper
122	100		
105	136.1	107.8	147.8
77	121.3	81.6	121.6






Abundance Scan 1043 (8.222 min): BF123565.D\data.ms (-1033)




4-Chloroaniline
Concen: 5.27 ng
RT: 8.216 min Scan# 1042
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005

Tgt Ion:	127	Resp:	91274
Ion	Ratio	Lower	Upper
127	100		
129	31.7	25.7	38.5
65	40.5	34.5	51.7
92	19.5	17.4	26.0






Abundance Scan 1056 (8.298 min): BF123565.D\data.ms (-1034)




Hexachlorobutadiene
Concen: 4.42 ng
RT: 8.298 min Scan# 1056
Delta R.T. -0.000 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	225	Resp:	40470
Ion	Ratio	Lower	Upper
225	100		
223	62.8	50.4	75.6
227	61.7	51.4	77.0






Abundance Scan 1107 (8.598 min): BF123565.D\data.ms (-11035)




Caprolactam
Concen: 4.09 ng
RT: 8.545 min Scan# 1098
Delta R.T. -0.053 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005

Tgt Ion:113 Resp: 16650
Ion Ratio Lower Upper
113 100
55 301.2 254.9 294.9#
56 227.4 179.4 219.4#






Abundance Scan 1125 (8.704 min): BF123565.D\data.ms (-11136)



4-Chloro-3-methylphenol
Concen: 5.24 ng
RT: 8.698 min Scan# 1124
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:107 Resp: 72558
Ion Ratio Lower Upper
107 100
144 19.8 16.3 24.5
142 68.2 50.1 75.1



Abundance Scan 1153 (8.869 min): BF123565.D\data.ms (-11437-)

2-Methylnaphthalene

Concen: 5.08 ng

RT: 8.869 min Scan# 1153

Delta R.T. -0.000 min

Lab File: BF123562.D

Acq: 15 Apr 2021 11:17

Tgt Ion:142 Resp: 146104

Ion Ratio Lower Upper

142 100

141 89.9 71.8 107.6

Instrument :

BNA_F

ClientSampleId :

SSTDIC005

Abundance Scan 1153 (8.869 min): BF123562.D\data.ms

Ref 50

39.0 63.0 89.0 115.0 142.1

m/z-->

Abundance Scan 1153 (8.869 min): BF123562.D\data.ms

Raw 50

39.0 63.0 89.0 115.0 142.1


m/z-->

Abundance Scan 1153 (8.869 min): BF123562.D\data.ms (-1102) (-)

Sub 50

39.0 63.0 89.0 115.0 142.1

m/z-->



Abundance Scan 1171 (8.975 min): BF123565.D\data.ms (-11438-)

1-Methylnaphthalene

Concen: 5.22 ng

RT: 8.969 min Scan# 1170

Delta R.T. -0.006 min

Lab File: BF123562.D

Acq: 15 Apr 2021 11:17

Tgt Ion:142 Resp: 141189

Ion Ratio Lower Upper

142 100

141 87.3 72.8 109.2

Abundance Scan 1170 (8.969 min): BF123562.D\data.ms

Ref 50

35.9 63.0 115.0 142.1 202.8 235.8

m/z-->

Abundance Scan 1170 (8.969 min): BF123562.D\data.ms

Raw 50

63.0 115.0 142.1 202.8


m/z-->

Abundance Scan 1170 (8.969 min): BF123562.D\data.ms (-1120) (-)


Sub 50

63.0 115.0 142.1 202.8

m/z-->






Abundance Scan 1331 (9.916 min): BF123565.D\data.ms (-1329)




Acenaphthene-d10
Concen: 20.00 ng
RT: 9.916 min Scan# 1331
Delta R.T. -0.000 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005

Tgt Ion:	164	Resp:	436067
Ion	Ratio	Lower	Upper
164	100		
162	102.2	80.7	121.1
160	47.4	37.3	55.9






Abundance Scan 1182 (9.039 min): BF123565.D\data.ms (-1179)



1,2,4,5-Tetrachlorobenzene
Concen: 5.00 ng
RT: 9.033 min Scan# 1181
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17


Tgt Ion:	216	Resp:	62448
Ion	Ratio	Lower	Upper
216	100		
214	77.5	63.8	95.8
179	22.4	17.9	26.9
108	22.0	18.6	28.0





Abundance Scan 1180 (9.028 min): BF123565.D\data.ms (-11741-)

Hexachlorocyclopentadiene
Concen: 3.61 ng
RT: 9.027 min Scan# 1180
Delta R.T. -0.000 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17


Instrument :
BNA_F
ClientSampleId :
SSTDIC005



Tgt Ion:	237	Resp:	22657
Ion	Ratio	Lower	Upper
237	100		
235	61.2	38.8	78.8
272	13.2	0.0	32.4





Abundance Scan 1180 (9.027 min): BF123562.D\data.ms (-1129-)




Abundance Scan 1466 (10.710 min): BF123565.D\data.ms (-1460-)

2,4,6-Tribromophenol
Concen: 9.75 ng
RT: 10.698 min Scan# 1464
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17


Tgt Ion:	330	Resp:	41498
Ion	Ratio	Lower	Upper
330	100		
332	95.9	77.1	115.7
141	38.9	27.9	41.9



Abundance Scan 1464 (10.698 min): BF123562.D\data.ms (-1415-)






Abundance Scan 1200 (9.145 min): BF123565.D\data.ms (-1195.3)




2,4,6-Trichlorophenol
Concen: 3.99 ng
RT: 9.145 min Scan# 1200
Delta R.T. -0.000 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005

Tgt Ion:	196	Resp:	34344
Ion	Ratio	Lower	Upper
196	100		
198	104.1	78.2	117.2
200	33.4	25.0	37.6






Abundance Scan 1207 (9.186 min): BF123565.D\data.ms (-1204.4)




2,4,5-Trichlorophenol
Concen: 4.41 ng
RT: 9.180 min Scan# 1206
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	196	Resp:	40901
Ion	Ratio	Lower	Upper
196	100		
198	88.3	76.5	114.7
97	57.9	52.9	79.3
132	30.6	25.3	37.9






Abundance Scan 1216 (9.239 min): BF123565.D\data.ms (-12145)




2-Fluorobiphenyl
Concen: 11.95 ng
RT: 9.233 min Scan# 1215
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005

Tgt Ion:	172	Resp:	320503
Ion	Ratio	Lower	Upper
172	100		
171	33.9	28.1	42.1
170	22.6	19.2	28.8






Abundance Scan 1233 (9.339 min): BF123565.D\data.ms (-12274)




1,1'-Biphenyl
Concen: 6.04 ng
RT: 9.333 min Scan# 1232
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	154	Resp:	180011
Ion	Ratio	Lower	Upper
154	100		
153	39.7	20.7	60.7
76	11.1	0.0	31.7






Abundance Scan 1237 (9.363 min): BF123565.D\data.ms (-1237-)




2-Chloronaphthalene
Concen: 5.59 ng
RT: 9.357 min Scan# 1236
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005

Tgt Ion:	162	Resp:	134922
Ion	Ratio	Lower	Upper
162	100		
127	43.5	36.2	54.2
164	33.1	25.4	38.2






Abundance Scan 1253 (9.457 min): BF123565.D\data.ms (-1248-)




2-Nitroaniline
Concen: 5.19 ng
RT: 9.445 min Scan# 1251
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	65	Resp:	44074
Ion	Ratio	Lower	Upper
65	100		
92	57.0	47.1	70.7
138	76.3	70.2	105.2




Abundance Scan 1308 (9.781 min): BF123565.D\data.ms (-1299)





Acenaphthylene
Concen: 6.03 ng
RT: 9.774 min Scan# 1307
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17


Instrument :
BNA_F
ClientSampleId :
SSTDIC005




Tgt Ion:	152	Resp:	219886
Ion	Ratio	Lower	Upper
152	100		
151	21.5	16.0	24.0
153	12.5	10.4	15.6





Abundance Scan 1284 (9.639 min): BF123565.D\data.ms (-1275)




Dimethylphthalate
Concen: 5.35 ng
RT: 9.627 min Scan# 1282
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17



Tgt Ion:	163	Resp:	156117
Ion	Ratio	Lower	Upper
163	100		
194	3.6	3.1	4.7
164	8.8	8.2	12.2





Abundance Scan 1294 (9.698 min): BF123565.D\data.ms (-12851-)




2,6-Dinitrotoluene
Concen: 4.49 ng
RT: 9.686 min Scan# 1292
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005


Tgt Ion:165 Resp: 29517
Ion Ratio Lower Upper
165 100
63 91.1 67.7 101.5
89 83.8 58.8 88.2



Abundance Scan 1292 (9.686 min): BF123562.D\data.ms (-1243-)





Abundance Scan 1337 (9.951 min): BF123565.D\data.ms (-13252-)




Acenaphthene
Concen: 5.84 ng
RT: 9.945 min Scan# 1336
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17


Tgt Ion:154 Resp: 136970
Ion Ratio Lower Upper
154 100
153 109.2 88.1 132.1
152 55.5 43.4 65.0



Abundance Scan 1336 (9.945 min): BF123562.D\data.ms (-1286-)






Abundance Scan 1323 (9.869 min): BF123565.D\data.ms (-1315)




3-Nitroaniline
Concen: 5.49 ng
RT: 9.857 min Scan# 1321
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005

Tgt Ion:138 Resp: 36922
Ion Ratio Lower Upper
138 100
108 11.5 10.2 15.4
92 105.2 103.3 154.9






Abundance Scan 1340 (9.969 min): BF123565.D\data.ms (-1335)



2,4-Dinitrophenol
Concen: 2.46 ng
RT: 9.963 min Scan# 1339
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17


Tgt Ion:184 Resp: 6468
Ion Ratio Lower Upper
184 100
63 84.6 73.5 110.3
154 72.4 52.5 78.7






Abundance Scan 1366 (10.122 min): BF123565.D\data.ms (-1366) (-)

Dibenzofuran
Concen: 6.06 ng
RT: 10.116 min Scan# 1365
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDICC005




Tgt Ion:	168	Resp:	207583
Ion	Ratio	Lower	Upper
168	100		
139	41.9	33.2	49.8
169	13.4	10.8	16.2







Abundance Scan 1349 (10.022 min): BF123565.D\data.ms (-1349) (-)

4-Nitrophenol
Concen: 4.61 ng
RT: 10.010 min Scan# 1347
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17



Tgt Ion:	139	Resp:	23078
Ion	Ratio	Lower	Upper
139	100		
109	83.6	63.2	103.2
65	113.9	110.0	150.0









Scan 1363 (10.104 min): BF123565.D\data.ms (-1356) (-)
 2,4-Dinitrotoluene
 Concen: 4.34 ng
 RT: 10.092 min Scan# 1361
 Delta R.T. -0.012 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Tgt Ion:	165	Resp:	36938
Ion	Ratio	Lower	Upper
165	100		
63	62.0	43.5	65.3
89	93.8	67.7	101.5
182	2.1	2.1	3.1#



Scan 1425 (10.469 min): BF123565.D\data.ms (-1458) (-)
 Fluorene
 Concen: 6.26 ng
 RT: 10.463 min Scan# 1424
 Delta R.T. -0.006 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17


Tgt Ion:	166	Resp:	164057
Ion	Ratio	Lower	Upper
166	100		
165	84.0	73.5	110.3
167	11.8	10.7	16.1






Abundance Scan 1386 (10.239 min): BF123565.D\data.ms (-1358) (-)

2,3,4,6-Tetrachlorophenol
Concen: 4.32 ng
RT: 10.233 min Scan# 1385
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005







Tgt Ion:	232	Resp:	31010
Ion	Ratio	Lower	Upper
232	100		
131	53.3	43.2	64.8
130	2.1	2.2	3.2#
166	31.7	27.3	40.9




Abundance Scan 1403 (10.339 min): BF123565.D\data.ms (-1359) (-)

Diethylphthalate
Concen: 5.80 ng
RT: 10.333 min Scan# 1402
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	149	Resp:	167589
Ion	Ratio	Lower	Upper
149	100		
177	20.2	15.4	23.2
150	11.5	9.9	14.9









Scan 1423 (10.457 min): BF123565.D\data.ms (-1418) (-)
 4-Chlorophenyl-phenylether
 Concen: 5.43 ng
 RT: 10.457 min Scan# 1423
 Delta R.T. -0.000 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005


Tgt Ion:	204	Resp:	74598
Ion	Ratio	Lower	Upper
204	100		
206	29.0	25.7	38.5
141	60.1	55.5	83.3



Scan 1424 (10.463 min): BF123565.D\data.ms (-1419) (-)
 4-Nitroaniline
 Concen: 5.69 ng
 RT: 10.463 min Scan# 1424
 Delta R.T. -0.024 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17

Tgt Ion:	138	Resp:	35854
Ion	Ratio	Lower	Upper
138	100		
92	52.2	35.8	75.8
108	85.7	63.6	103.6









Azobenzene
Concen: 6.08 ng
RT: 10.616 min Scan# 1450
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17




Instrument :
BNA_F
ClientSampleId :
SSTDIC005


Tgt Ion:	77	Resp:	190803
Ion	Ratio	Lower	Upper
77	100		
182	21.8	1.7	41.7
105	21.9	1.7	41.7
51	34.4	15.9	55.9



Phenanthrene-d10
Concen: 20.00 ng
RT: 11.404 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	188	Resp:	848845
Ion	Ratio	Lower	Upper
188	100		
94	8.9	6.9	10.3
80	10.4	8.2	12.2









4,6-Dinitro-2-methylphenol
Concen: 2.54 ng
RT: 10.498 min Scan# 1430
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17




Instrument :
BNA_F
ClientSampleId :
SSTDIC005


Tgt Ion:	198	Resp:	11016
Ion	Ratio	Lower	Upper
198	100		
51	100.0	64.8	104.8
105	78.7	42.6	82.6



n-Nitrosodiphenylamine
Concen: 5.67 ng
RT: 10.569 min Scan# 1442
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	169	Resp:	138178
Ion	Ratio	Lower	Upper
169	100		
168	64.8	53.3	79.9
167	35.6	29.9	44.9









4-Bromophenyl-phenylether
Concen: 4.83 ng
RT: 10.945 min Scan# 1506
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17




Instrument :
BNA_F
ClientSampleId :
SSTDIC005


Tgt Ion:	248	Resp:	41988
Ion	Ratio	Lower	Upper
248	100		
250	96.6	77.4	116.2
141	85.5	63.7	95.5



Hexachlorobenzene
Concen: 5.23 ng
RT: 11.016 min Scan# 1518
Delta R.T. -0.000 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	284	Resp:	49505
Ion	Ratio	Lower	Upper
284	100		
142	27.4	28.3	42.5#
249	28.2	23.5	35.3









#59 (-)
Atrazine
 Concen: 5.10 ng
 RT: 11.092 min Scan# 1531
 Delta R.T. -0.012 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17




Instrument :
 BNA_F
ClientSampleId :
 SSTDICC005


Tgt Ion:	200	Resp:	42131
Ion	Ratio	Lower	Upper
200	100		
173	32.3	6.9	46.9
215	46.9	26.5	66.5



#70 (-)
Pentachlorophenol
 Concen: 3.30 ng
 RT: 11.204 min Scan# 1550
 Delta R.T. -0.006 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17

Tgt Ion:	266	Resp:	16410
Ion	Ratio	Lower	Upper
266	100		
268	67.1	51.6	77.4
264	64.1	52.6	78.8









#72 (-)
Phenanthrene
Concen: 5.68 ng
RT: 11.427 min Scan# 1588
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17




Instrument :
BNA_F
ClientSampleId :
SSTDIC005


Tgt Ion:	178	Resp:	236534
Ion	Ratio	Lower	Upper
178	100		
176	17.8	15.6	23.4
179	15.0	12.5	18.7



#73 (-)
Anthracene
Concen: 5.65 ng
RT: 11.480 min Scan# 1597
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	178	Resp:	235253
Ion	Ratio	Lower	Upper
178	100		
176	18.1	15.4	23.0
179	14.8	12.6	18.8









Scan 1624 (11.639 min): BF123565.D\data.ms (-1673) (-)
 Carbazole
 Concen: 5.84 ng
 RT: 11.627 min Scan# 1622
 Delta R.T. -0.012 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Tgt Ion:	167	Resp:	231652
Ion	Ratio	Lower	Upper
167	100		
166	20.2	17.0	25.6
139	13.8	10.4	15.6




Scan 1680 (11.969 min): BF123565.D\data.ms (-1673) (-)
 Di-n-butylphthalate
 Concen: 4.93 ng
 RT: 11.968 min Scan# 1680
 Delta R.T. -0.000 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17

Tgt Ion:	149	Resp:	261241
Ion	Ratio	Lower	Upper
149	100		
150	7.9	7.8	11.8
104	5.4	4.6	6.8






Abundance Scan 1791 (12.621 min): BF123565.D\data.ms (-1740) (-)




Fluoranthene
Concen: 5.87 ng
RT: 12.621 min Scan# 1791
Delta R.T. -0.000 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005

Tgt Ion:	202	Resp:	261290
Ion	Ratio	Lower	Upper
202	100		
101	8.9	0.0	30.9
203	16.9	0.0	37.5







Abundance Scan 2035 (14.057 min): BF123565.D\data.ms (-1984) (-)



Chrysene-d12
Concen: 20.00 ng
RT: 14.057 min Scan# 2035
Delta R.T. -0.000 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	240	Resp:	1163344
Ion	Ratio	Lower	Upper
240	100		
120	10.9	7.2	10.8#
236	24.2	19.0	28.4









Scan 1811 (12.739 min): BF123565.D\data.ms (-1807) (-)
 Benzidine
 Concen: 4.66 ng
 RT: 12.739 min Scan# 1811
 Delta R.T. -0.000 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Tgt Ion:	184	Resp:	146089
Ion	Ratio	Lower	Upper
184	100		
185	15.9	12.1	18.1
183	10.4	7.8	11.8



Scan 1830 (12.851 min): BF123565.D\data.ms (-1873) (-)
 Pyrene
 Concen: 5.06 ng
 RT: 12.851 min Scan# 1830
 Delta R.T. -0.000 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17


Tgt Ion:	202	Resp:	468441
Ion	Ratio	Lower	Upper
202	100		
200	17.7	16.1	24.1
203	16.8	14.8	22.2





Abundance Scan 1855 (12.998 min): BF123565.D\data.ms (-1879) (-)

Terphenyl-d14
Concen: 10.46 ng
RT: 12.992 min Scan# 1854
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17


Instrument :
BNA_F
ClientSampleId :
SSTDIC005



Tgt Ion:	244	Resp:	700211
Ion	Ratio	Lower	Upper
244	100		
212	6.5	5.7	8.5
122	12.8	9.7	14.5





Abundance Scan 1854 (12.992 min): BF123562.D\data.ms (-1804) (-)




Abundance Scan 1936 (13.474 min): BF123565.D\data.ms (-1930) (-)

Butylbenzylphthalate
Concen: 3.98 ng
RT: 13.474 min Scan# 1936
Delta R.T. -0.000 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17


Tgt Ion:	149	Resp:	167748
Ion	Ratio	Lower	Upper
149	100		
91	73.9	63.6	95.4
206	19.2	14.5	21.7



Abundance Scan 1936 (13.474 min): BF123562.D\data.ms (-1885) (-)






Abundance Scan 2033 (14.045 min): BF123565.D\data.ms (-2033) (-)




Benzo(a)anthracene
Concen: 5.04 ng
RT: 14.045 min Scan# 2033
Delta R.T. -0.000 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005

Tgt Ion:	228	Resp:	366357
Ion	Ratio	Lower	Upper
228	100		
226	26.8	22.4	33.6
229	19.1	16.0	24.0







Abundance Scan 2027 (14.010 min): BF123565.D\data.ms (-2027) (-)



3,3'-Dichlorobenzidine
Concen: 5.11 ng
RT: 14.004 min Scan# 2026
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	252	Resp:	118497
Ion	Ratio	Lower	Upper
252	100		
254	62.1	51.1	76.7
126	13.1	7.7	11.5#









#33 (-)
Chrysene
 Concen: 5.11 ng
 RT: 14.080 min Scan# 2039
 Delta R.T. -0.006 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17




Instrument :
 BNA_F
ClientSampleId :
 SSTDICC005

Tgt Ion:	228	Resp:	371799
Ion	Ratio	Lower	Upper
228	100		
226	29.9	24.3	36.5
229	18.2	16.2	24.2



#34 (-)
Bis(2-ethylhexyl)phthalate
 Concen: 4.12 ng
 RT: 14.039 min Scan# 2032
 Delta R.T. -0.000 min
 Lab File: BF123562.D
 Acq: 15 Apr 2021 11:17


Tgt Ion:	149	Resp:	229020
Ion	Ratio	Lower	Upper
149	100		
167	28.0	21.8	32.6
279	2.9	3.3	4.9#






Abundance Scan 2137 (14.657 min): BF123565.D\data.ms (-2135) (-)

Di-n-octyl phthalate
Concen: 4.49 ng
RT: 14.656 min Scan# 2137
Delta R.T. -0.000 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005







Tgt Ion:	149	Resp:	363032
Ion	Ratio	Lower	Upper
149	100		
167	1.4	1.2	1.8
43	13.9	11.8	17.6




Abundance Scan 2287 (15.539 min): BF123565.D\data.ms (-2286) (-)

Perylene-d12
Concen: 20.00 ng
RT: 15.545 min Scan# 2288
Delta R.T. 0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Tgt Ion:	264	Resp:	700286
Ion	Ratio	Lower	Upper
264	100		
260	23.1	18.7	28.1
265	21.1	17.8	26.8






Abundance Scan 2542 (17.039 min): BF123565.D\data.ms (-2542) (-)




Indeno(1,2,3-cd)pyrene
Concen: 4.20 ng
RT: 17.027 min Scan# 2540
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005

Tgt Ion:276 Resp: 207590
Ion Ratio Lower Upper
276 100
138 22.3 16.2 24.2
277 23.9 20.1 30.1






Abundance Scan 2214 (15.110 min): BF123565.D\data.ms (-2214) (-)



Benzo(b)fluoranthene
Concen: 7.45 ng
RT: 15.104 min Scan# 2213
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17


Tgt Ion:252 Resp: 352041
Ion Ratio Lower Upper
252 100
253 22.4 18.1 27.1
125 11.5 6.2 9.4#






Abundance Scan 2219 (15.139 min): BF123565.D\data.ms (-2168) (-)

Benzo(k)fluoranthene
Concen: 7.70 ng
RT: 15.133 min Scan# 2218
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005




Tgt Ion:	252	Resp:	359854
Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.9	26.9
125	11.3	7.0	10.6#






Abundance Scan 2277 (15.480 min): BF123565.D\data.ms (-2168) (-)

Benzo(a)pyrene
Concen: 4.52 ng
RT: 15.474 min Scan# 2276
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17




Tgt Ion:	252	Resp:	187843
Ion	Ratio	Lower	Upper
252	100		
253	21.1	18.4	27.6
125	10.0	8.0	12.0






Abundance Scan 2545 (17.057 min): BF123565.D\data.ms (-2545) (-)

Dibenzo(a,h)anthracene
Concen: 4.12 ng
RT: 17.050 min Scan# 2544
Delta R.T. -0.006 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17

Instrument :
BNA_F
ClientSampleId :
SSTDIC005




Tgt Ion:	278	Resp:	174598
Ion	Ratio	Lower	Upper
278	100		
139	13.8	10.2	15.2
279	21.2	19.4	29.0



Abundance Scan 2618 (17.486 min): BF123565.D\data.ms (-2618) (-)

Benzo(g,h,i)perylene
Concen: 4.47 ng
RT: 17.474 min Scan# 2616
Delta R.T. -0.012 min
Lab File: BF123562.D
Acq: 15 Apr 2021 11:17



Tgt Ion:	276	Resp:	175392
Ion	Ratio	Lower	Upper
276	100		
277	25.9	18.2	27.2
138	17.5	13.8	20.6

