

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041521\
 Data File : BF123563.D
 Acq On : 15 Apr 2021 11:49
 Operator : JU/CG
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Quant Time: Apr 15 13:27:24 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 15 13:22:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.869	152	207816	20.00	ng	0.00
21) Naphthalene-d8	8.151	136	801512	20.00	ng	0.00
39) Acenaphthene-d10	9.916	164	418401	20.00	ng	0.00
64) Phenanthrene-d10	11.404	188	807744	20.00	ng	0.00
76) Chrysene-d12	14.051	240	691221	20.00	ng	0.00
86) Perylene-d12	15.539	264	609063	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.475	112	287225	23.60	ng	0.00
7) Phenol-d6	6.487	99	394222	23.76	ng	-0.01
23) Nitrobenzene-d5	7.428	82	362688	22.60	ng	0.00
42) 2,4,6-Tribromophenol	10.698	330	93222	22.82	ng	-0.01
45) 2-Fluorobiphenyl	9.233	172	633885	24.64	ng	0.00
79) Terphenyl-d14	12.992	244	787101	19.80	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.616	88	73327	7.40	ng	99
3) Pyridine	3.369	79	177606	6.22	ng	98
4) n-Nitrosodimethylamine	3.299	42	105851	9.60	ng	93
6) Aniline	6.522	93	226564	10.94	ng	98
8) 2-Chlorophenol	6.651	128	153473	11.77	ng	96
9) Benzaldehyde	6.410	77	131778	12.52	ng	99
10) Phenol	6.498	94	197512	11.20	ng	94
11) bis(2-Chloroethyl)ether	6.598	93	151793	10.83	ng	96
12) 1,3-Dichlorobenzene	6.810	146	154889	10.94	ng	94
13) 1,4-Dichlorobenzene	6.887	146	156542	10.95	ng	95
14) 1,2-Dichlorobenzene	7.039	146	151803	11.36	ng	99
15) Benzyl Alcohol	7.004	79	145345	11.38	ng	97
16) 2,2'-oxybis(1-Chloropr...	7.145	45	343453	12.92	ng	99
17) 2-Methylphenol	7.110	107	125681	11.36	ng	95
18) Hexachloroethane	7.386	117	62354	10.98	ng	98
19) n-Nitroso-di-n-propyla...	7.269	70	126767	11.49	ng	99
20) 3+4-Methylphenols	7.263	107	165804	12.24	ng	91
22) Acetophenone	7.269	105	239801	11.55	ng	# 94
24) Nitrobenzene	7.445	77	192915	11.11	ng	95
25) Isophorone	7.681	82	348064	10.77	ng	96
26) 2-Nitrophenol	7.763	139	52705	7.22	ng	93
27) 2,4-Dimethylphenol	7.798	122	119801	10.53	ng	97
28) bis(2-Chloroethoxy)met...	7.898	93	177477	10.25	ng	99
29) 2,4-Dichlorophenol	8.004	162	116381	9.64	ng	100
30) 1,2,4-Trichlorobenzene	8.092	180	127436	9.29	ng	99
31) Naphthalene	8.175	128	451820	11.37	ng	100
32) Benzoic acid	7.869	122	45589	6.06	ng	85
33) 4-Chloroaniline	8.216	127	176947	10.71	ng	98
34) Hexachlorobutadiene	8.298	225	79082	9.04	ng	98
35) Caprolactam	8.551	113	38807	9.99	ng	95
36) 4-Chloro-3-methylphenol	8.698	107	147645	11.18	ng	90
37) 2-Methylnaphthalene	8.869	142	291999	10.63	ng	96
38) 1-Methylnaphthalene	8.969	142	275272	10.66	ng	100
40) 1,2,4,5-Tetrachloroben...	9.033	216	122437	10.22	ng	97
41) Hexachlorocyclopentadiene	9.028	237	54980	9.14	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041521\
 Data File : BF123563.D
 Acq On : 15 Apr 2021 11:49
 Operator : JU/CG
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Quant Time: Apr 15 13:27:24 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 15 13:22:17 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	9.145	196	80690	9.76	ng	99
44) 2,4,5-Trichlorophenol	9.181	196	90467	10.16	ng	96
46) 1,1'-Biphenyl	9.333	154	359856	12.58	ng	99
47) 2-Chloronaphthalene	9.357	162	262003	11.32	ng	98
48) 2-Nitroaniline	9.445	65	101056	12.40	ng	88
49) Acenaphthylene	9.775	152	438873	12.55	ng	99
50) Dimethylphthalate	9.628	163	316071	11.28	ng	99
51) 2,6-Dinitrotoluene	9.686	165	65053	10.31	ng	95
52) Acenaphthene	9.945	154	272417	12.10	ng	97
53) 3-Nitroaniline	9.857	138	76481	11.86	ng	97
54) 2,4-Dinitrophenol	9.957	184	18938	7.49	ng	# 5
55) Dibenzofuran	10.116	168	411325	12.52	ng	99
56) 4-Nitrophenol	10.010	139	54787	11.40	ng	89
57) 2,4-Dinitrotoluene	10.092	165	86082	10.55	ng	98
58) Fluorene	10.463	166	332788	13.23	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.233	232	67537	9.81	ng	98
60) Diethylphthalate	10.333	149	345422	12.47	ng	98
61) 4-Chlorophenyl-phenyle...	10.457	204	150366	11.41	ng	95
62) 4-Nitroaniline	10.463	138	78195	12.94	ng	99
63) Azobenzene	10.616	77	392538	13.04	ng	99
65) 4,6-Dinitro-2-methylph...	10.498	198	31882	7.72	ng	93
66) n-Nitrosodiphenylamine	10.569	169	280555	12.10	ng	97
67) 4-Bromophenyl-phenylether	10.945	248	91470	11.05	ng	98
68) Hexachlorobenzene	11.010	284	100046	11.10	ng	97
69) Atrazine	11.092	200	86977	11.06	ng	98
70) Pentachlorophenol	11.204	266	42036	8.89	ng	91
71) Phenanthrene	11.427	178	479774	12.10	ng	99
72) Anthracene	11.474	178	474892	11.99	ng	99
73) Carbazole	11.627	167	475371	12.58	ng	99
74) Di-n-butylphthalate	11.969	149	573067	11.37	ng	99
75) Fluoranthene	12.616	202	534182	12.62	ng	100
77) Benzidine	12.739	184	229818	12.35	ng	98
78) Pyrene	12.845	202	547643	9.95	ng	98
80) Butylbenzylphthalate	13.468	149	220158	8.79	ng	98
81) Benzo(a)anthracene	14.039	228	451487	10.46	ng	97
82) 3,3'-Dichlorobenzidine	14.004	252	150648	10.94	ng	96
83) Chrysene	14.074	228	449863	10.40	ng	99
84) Bis(2-ethylhexyl)phtha...	14.033	149	323798	9.80	ng	# 97
85) Di-n-octyl phthalate	14.651	149	501946	10.45	ng	99
87) Indeno(1,2,3-cd)pyrene	17.027	276	357364	8.31	ng	98
88) Benzo(b)fluoranthene	15.098	252	417012	10.14	ng	98
89) Benzo(k)fluoranthene	15.127	252	398458	9.81	ng	98
90) Benzo(a)pyrene	15.474	252	367587	10.16	ng	98
91) Dibenzo(a,h)anthracene	17.045	278	301522	8.19	ng	99
92) Benzo(g,h,i)perylene	17.474	276	288431	8.44	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041521\
 Data File : BF123563.D
 Acq On : 15 Apr 2021 11:49
 Operator : JU/CG
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Quant Time: Apr 15 13:27:24 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 15 13:22:17 2021
 Response via : Initial Calibration









1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.869 min Scan# 813
Delta R.T. -0.000 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:152 Resp: 207816
Ion Ratio Lower Upper
152 100
150 152.0 127.0 190.6
115 62.1 54.3 81.5



1,4-Dioxane
Concen: 7.40 ng
RT: 2.616 min Scan# 90
Delta R.T. -0.000 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Tgt Ion: 88 Resp: 73327
Ion Ratio Lower Upper
88 100
58 103.7 83.8 125.8
43 39.3 32.2 48.4









Pyridine
Concen: 6.22 ng
RT: 3.369 min Scan# 218
Delta R.T. 0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion: 79 Resp: 177606
Ion Ratio Lower Upper
79 100
52 107.5 87.4 131.0
51 50.4 41.4 62.0



n-Nitrosodimethylamine
Concen: 9.60 ng
RT: 3.299 min Scan# 206
Delta R.T. -0.024 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Tgt Ion: 42 Resp: 105851
Ion Ratio Lower Upper
42 100
74 96.7 83.5 125.3
44 5.8 5.7 8.5









2-Fluorophenol
 Concen: 23.60 ng
 RT: 5.475 min Scan# 576
 Delta R.T. -0.006 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion:	112	Resp:	287225
Ion	Ratio	Lower	Upper
112	100		
64	73.7	61.4	92.0
63	40.7	33.0	49.4



Aniline
 Concen: 10.94 ng
 RT: 6.522 min Scan# 754
 Delta R.T. -0.012 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49

Tgt Ion:	93	Resp:	226564
Ion	Ratio	Lower	Upper
93	100		
66	44.6	36.4	54.6
65	22.7	19.4	29.0









Phenol-d6
 Concen: 23.76 ng
 RT: 6.487 min Scan# 748
 Delta R.T. -0.012 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion:	99	Resp:	394222
Ion	Ratio	Lower	Upper
99	100		
42	22.6	19.8	29.8
71	35.8	31.1	46.7



2-Chlorophenol
 Concen: 11.77 ng
 RT: 6.651 min Scan# 776
 Delta R.T. -0.006 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49

Tgt Ion:	128	Resp:	153473
Ion	Ratio	Lower	Upper
128	100		
130	32.1	12.6	52.6
64	57.2	41.5	81.5









Benzaldehyde
 Concen: 12.52 ng
 RT: 6.410 min Scan# 735
 Delta R.T. -0.006 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion:	77	Resp:	131778
Ion	Ratio	Lower	Upper
77	100		
105	95.0	75.2	115.2
106	90.6	71.6	111.6



Phenol
 Concen: 11.20 ng
 RT: 6.498 min Scan# 750
 Delta R.T. -0.012 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49

Tgt Ion:	94	Resp:	197512
Ion	Ratio	Lower	Upper
94	100		
65	35.8	18.1	58.1
66	43.6	29.0	69.0









#1
 bis(2-Chloroethyl)ether
 Concen: 10.83 ng
 RT: 6.598 min Scan# 767
 Delta R.T. -0.006 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Tgt Ion:	93	Resp:	151793
Ion	Ratio	Lower	Upper
93	100		
63	90.4	76.1	116.1
95	31.7	11.5	51.5



#12
 1,3-Dichlorobenzene
 Concen: 10.94 ng
 RT: 6.810 min Scan# 803
 Delta R.T. -0.000 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49


Tgt Ion:	146	Resp:	154889
Ion	Ratio	Lower	Upper
146	100		
148	65.1	50.9	76.3
75	35.8	34.2	51.4





Abundance Scan 816 (6.887 min): BF123565.D\data.ms (-811) #13

1,4-Dichlorobenzene
Concen: 10.95 ng
RT: 6.887 min Scan# 816
Delta R.T. -0.000 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49


Instrument :
BNA_F
ClientSampleId :
SSTDIC010



Tgt Ion:	146	Resp:	156542
Ion	Ratio	Lower	Upper
146	100		
148	64.1	49.9	74.9
111	43.8	39.5	59.3




Abundance Scan 816 (6.887 min): BF123563.D\data.ms (-765) (-)





Abundance Scan 843 (7.045 min): BF123565.D\data.ms (-837) #14

1,2-Dichlorobenzene
Concen: 11.36 ng
RT: 7.039 min Scan# 842
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Tgt Ion:	146	Resp:	151803
Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.6	77.4
111	46.4	38.4	57.6




Abundance Scan 842 (7.039 min): BF123563.D\data.ms (-792) (-)






Abundance Scan 837 (7.010 min): BF123565.D\data.ms (-831) #15

Benzyl Alcohol
Concen: 11.38 ng
RT: 7.004 min Scan# 836
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010







Tgt Ion:	79	Resp:	145345
Ion	Ratio	Lower	Upper
79	100		
108	72.8	55.8	83.8
77	64.0	49.9	74.9



Abundance Scan 860 (7.145 min): BF123565.D\data.ms (-852) #16

2,2'-oxybis(1-Chloropropane)
Concen: 12.92 ng
RT: 7.145 min Scan# 860
Delta R.T. -0.000 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49


Tgt Ion:	45	Resp:	343453
Ion	Ratio	Lower	Upper
45	100		
77	9.4	0.0	30.0
79	7.5	0.0	27.3






Abundance Scan 856 (7.122 min): BF123565.D\data.ms (-850) #17

2-Methylphenol
Concen: 11.36 ng
RT: 7.110 min Scan# 854
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010







Tgt Ion:	107	Resp:	125681
Ion	Ratio	Lower	Upper
107	100		
108	116.8	89.8	134.8
77	54.8	40.9	61.3
79	53.9	40.2	60.4




Abundance Scan 901 (7.387 min): BF123565.D\data.ms (-896) #18

Hexachloroethane
Concen: 10.98 ng
RT: 7.386 min Scan# 901
Delta R.T. -0.000 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49


Tgt Ion:	117	Resp:	62354
Ion	Ratio	Lower	Upper
117	100		
119	98.1	77.8	116.8
201	88.8	67.8	101.8








#19
n-Nitroso-di-n-propylamine
Concen: 11.49 ng
RT: 7.269 min Scan# 881
Delta R.T. -0.018 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49


Instrument :
BNA_F
ClientSampleId :
SSTDICC010





Tgt Ion:	70	Resp:	126767
Ion	Ratio	Lower	Upper
70	100		
42	75.8	60.1	90.1
101	8.7	7.4	11.2
130	19.2	14.2	21.2




#20
3+4-Methylphenols
Concen: 12.24 ng
RT: 7.263 min Scan# 880
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49



Tgt Ion:	107	Resp:	165804
Ion	Ratio	Lower	Upper
107	100		
108	91.1	71.5	111.5
77	83.2	82.0	122.0
79	33.0	14.1	54.1







Abundance Scan 1032 (8.157 min): BF123565.D\data.ms (-1021)



Naphthalene-d8
Concen: 20.00 ng
RT: 8.151 min Scan# 1031
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49




Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	136	Resp:	801512
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.7	13.1
54	13.3	9.8	14.6
68	5.6	4.1	6.1



Acetophenone
Concen: 11.55 ng
RT: 7.269 min Scan# 881
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49


Tgt Ion:	105	Resp:	239801
Ion	Ratio	Lower	Upper
105	100		
71	8.2	4.2	6.4#
51	46.6	33.6	50.4
120	19.4	16.3	24.5






Abundance Scan 909 (7.434 min): BF123565.D\data.ms (-904) #23

Nitrobenzene-d5
Concen: 22.60 ng
RT: 7.428 min Scan# 908
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	82	Resp:	362688
Ion	Ratio	Lower	Upper
82	100		
128	37.8	29.4	44.2
54	64.7	54.1	81.1






Abundance Scan 913 (7.457 min): BF123565.D\data.ms (-905) #24

Nitrobenzene
Concen: 11.11 ng
RT: 7.445 min Scan# 911
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49




Tgt Ion:	77	Resp:	192915
Ion	Ratio	Lower	Upper
77	100		
123	39.3	34.5	51.7
65	15.7	12.2	18.2






Abundance Scan 953 (7.692 min): BF123565.D\data.ms (-947) #25

Isophorone
Concen: 10.77 ng
RT: 7.681 min Scan# 951
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010






Tgt Ion:	82	Resp:	348064
Ion	Ratio	Lower	Upper
82	100		
95	6.6	5.8	8.6
138	12.9	11.8	17.8




Abundance Scan 966 (7.769 min): BF123565.D\data.ms (-961) #26

2-Nitrophenol
Concen: 7.22 ng
RT: 7.763 min Scan# 965
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Tgt Ion:	139	Resp:	52705
Ion	Ratio	Lower	Upper
139	100		
109	37.8	28.6	43.0
65	78.4	56.8	85.2





Abundance Scan 972 (7.804 min): BF123565.D\data.ms (-968) #27




2,4-Dimethylphenol
Concen: 10.53 ng
RT: 7.798 min Scan# 971
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:	122	Resp:	119801
Ion	Ratio	Lower	Upper
122	100		
107	134.8	106.3	159.5
121	60.8	45.4	68.0



Abundance Scan 971 (7.798 min): BF123563.D\data.ms (-921) (-)





Abundance Scan 989 (7.904 min): BF123565.D\data.ms (-980) #28




bis(2-Chloroethoxy)methane
Concen: 10.25 ng
RT: 7.898 min Scan# 988
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49


Tgt Ion:	93	Resp:	177477
Ion	Ratio	Lower	Upper
93	100		
95	32.5	26.5	39.7
123	11.3	8.4	12.6



Abundance Scan 988 (7.898 min): BF123563.D\data.ms (-938) (-)





Abundance Scan 1007 (8.010 min): BF123565.D\data.ms (-10629)




2,4-Dichlorophenol
Concen: 9.64 ng
RT: 8.004 min Scan# 1006
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:	162	Resp:	116381
Ion	Ratio	Lower	Upper
162	100		
164	63.8	44.1	84.1
98	47.5	27.3	67.3



Abundance Scan 1006 (8.004 min): BF123563.D\data.ms (-956) (-)





Abundance Scan 1022 (8.098 min): BF123565.D\data.ms (-10630)




1,2,4-Trichlorobenzene
Concen: 9.29 ng
RT: 8.092 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49


Tgt Ion:	180	Resp:	127436
Ion	Ratio	Lower	Upper
180	100		
182	95.1	77.2	115.8
145	33.4	26.0	39.0



Abundance Scan 1021 (8.092 min): BF123563.D\data.ms (-971) (-)






Abundance Scan 1036 (8.181 min): BF123565.D\data.ms (-1028) (-)




Naphthalene
 Concen: 11.37 ng
 RT: 8.175 min Scan# 1035
 Delta R.T. -0.006 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49

Instrument :
 BNA_F
ClientSampleId :
 SSTDICC010

Tgt Ion:	128	Resp:	451820
Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.9	13.3
127	13.5	10.8	16.2






Abundance Scan 992 (7.922 min): BF123565.D\data.ms (-977) (-)




Benzoic acid
 Concen: 6.06 ng
 RT: 7.869 min Scan# 983
 Delta R.T. -0.053 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49

Tgt Ion:	122	Resp:	45589
Ion	Ratio	Lower	Upper
122	100		
105	142.8	107.8	147.8
77	118.7	81.6	121.6






Abundance Scan 1043 (8.222 min): BF123565.D\data.ms (-1033)




4-Chloroaniline
Concen: 10.71 ng
RT: 8.216 min Scan# 1042
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	127	Resp:	176947
Ion	Ratio	Lower	Upper
127	100		
129	32.7	25.7	38.5
65	41.5	34.5	51.7
92	23.1	17.4	26.0






Abundance Scan 1056 (8.298 min): BF123565.D\data.ms (-1056)




Hexachlorobutadiene
Concen: 9.04 ng
RT: 8.298 min Scan# 1056
Delta R.T. -0.000 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Tgt Ion:	225	Resp:	79082
Ion	Ratio	Lower	Upper
225	100		
223	63.2	50.4	75.6
227	67.0	51.4	77.0





Abundance Scan 1107 (8.598 min): BF123565.D\data.ms (-11035)




Caprolactam
Concen: 9.99 ng
RT: 8.551 min Scan# 1099
Delta R.T. -0.047 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:113 Resp: 38807
Ion Ratio Lower Upper
113 100
55 267.1 254.9 294.9
56 188.8 179.4 219.4



Abundance Scan 1099 (8.551 min): BF123563.D\data.ms (-1056) (-)





Abundance Scan 1125 (8.704 min): BF123565.D\data.ms (-11136)




4-Chloro-3-methylphenol
Concen: 11.18 ng
RT: 8.698 min Scan# 1124
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Tgt Ion:107 Resp: 147645
Ion Ratio Lower Upper
107 100
144 23.2 16.3 24.5
142 71.0 50.1 75.1




Abundance Scan 1124 (8.698 min): BF123563.D\data.ms (-1074) (-)






Abundance Scan 1153 (8.869 min): BF123565.D\data.ms (-11437-)

2-Methylnaphthalene
Concen: 10.63 ng
RT: 8.869 min Scan# 1153
Delta R.T. -0.000 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampled :
SSTDIC010




Tgt Ion:142 Resp: 291999
Ion Ratio Lower Upper
142 100
141 86.4 71.8 107.6






Abundance Scan 1171 (8.975 min): BF123565.D\data.ms (-11438-)

1-Methylnaphthalene
Concen: 10.66 ng
RT: 8.969 min Scan# 1170
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49



Tgt Ion:142 Resp: 275272
Ion Ratio Lower Upper
142 100
141 90.8 72.8 109.2



Abundance Scan 1331 (9.916 min): BF123565.D\data.ms (-1329)

Acenaphthene-d10

Concen: 20.00 ng

RT: 9.916 min Scan# 1331

Delta R.T. -0.000 min

Lab File: BF123563.D


Acq: 15 Apr 2021 11:49

Instrument :

BNA_F

ClientSampleId :

SSTDIC010






Tgt Ion:164 Resp: 418401

Ion Ratio Lower Upper

164 100

162 102.7 80.7 121.1

160 45.2 37.3 55.9



Abundance Scan 1182 (9.039 min): BF123565.D\data.ms (-1179)

1,2,4,5-Tetrachlorobenzene


Concen: 10.22 ng

RT: 9.033 min Scan# 1181

Delta R.T. -0.006 min

Lab File: BF123563.D

Acq: 15 Apr 2021 11:49



Tgt Ion:216 Resp: 122437




Ion Ratio Lower Upper

216 100

214 82.8 63.8 95.8

179 22.3 17.9 26.9


108 22.0 18.6 28.0






Abundance Scan 1180 (9.028 min): BF123565.D\data.ms (-11741-)

Hexachlorocyclopentadiene
Concen: 9.14 ng
RT: 9.028 min Scan# 1180
Delta R.T. -0.000 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	237	Resp:	54980
Ion	Ratio	Lower	Upper
237	100		
235	57.0	38.8	78.8
272	12.6	0.0	32.4






Abundance Scan 1466 (10.710 min): BF123565.D\data.ms (-1460-)


2,4,6-Tribromophenol
Concen: 22.82 ng
RT: 10.698 min Scan# 1464
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49



Tgt Ion:	330	Resp:	93222
Ion	Ratio	Lower	Upper
330	100		
332	99.3	77.1	115.7
141	36.6	27.9	41.9





Abundance Scan 1200 (9.145 min): BF123563.D\data.ms (-1195) (-)




2,4,6-Trichlorophenol
Concen: 9.76 ng
RT: 9.145 min Scan# 1200
Delta R.T. -0.000 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:	196	Resp:	80690
Ion	Ratio	Lower	Upper
196	100		
198	97.2	78.2	117.2
200	29.2	25.0	37.6



Abundance Scan 1200 (9.145 min): BF123563.D\data.ms (-1149) (-)





Abundance Scan 1207 (9.186 min): BF123563.D\data.ms (-1204) (-)




2,4,5-Trichlorophenol
Concen: 10.16 ng
RT: 9.181 min Scan# 1206
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49


Tgt Ion:	196	Resp:	90467
Ion	Ratio	Lower	Upper
196	100		
198	95.8	76.5	114.7
97	58.6	52.9	79.3
132	29.5	25.3	37.9



Abundance Scan 1206 (9.181 min): BF123563.D\data.ms (-1156) (-)






Abundance Scan 1216 (9.239 min): BF123565.D\data.ms (-12145)




2-Fluorobiphenyl
Concen: 24.64 ng
RT: 9.233 min Scan# 1215
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	172	Resp:	633885
Ion	Ratio	Lower	Upper
172	100		
171	34.5	28.1	42.1
170	22.9	19.2	28.8






Abundance Scan 1233 (9.339 min): BF123565.D\data.ms (-122746)




1,1'-Biphenyl
Concen: 12.58 ng
RT: 9.333 min Scan# 1232
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Tgt Ion:	154	Resp:	359856
Ion	Ratio	Lower	Upper
154	100		
153	40.5	20.7	60.7
76	11.1	0.0	31.7






Abundance Scan 1237 (9.363 min): BF123565.D\data.ms (-1237-)




2-Chloronaphthalene
Concen: 11.32 ng
RT: 9.357 min Scan# 1236
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	162	Resp:	262003
Ion	Ratio	Lower	Upper
162	100		
127	43.8	36.2	54.2
164	32.5	25.4	38.2






Abundance Scan 1253 (9.457 min): BF123565.D\data.ms (-1248-)




2-Nitroaniline
Concen: 12.40 ng
RT: 9.445 min Scan# 1251
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Tgt Ion:	65	Resp:	101056
Ion	Ratio	Lower	Upper
65	100		
92	55.2	47.1	70.7
138	72.1	70.2	105.2






Abundance Scan 1308 (9.781 min): BF123565.D\data.ms (-1299)




Acenaphthylene
Concen: 12.55 ng
RT: 9.775 min Scan# 1307
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	152	Resp:	438873
Ion	Ratio	Lower	Upper
152	100		
151	20.4	16.0	24.0
153	13.1	10.4	15.6






Abundance Scan 1284 (9.639 min): BF123565.D\data.ms (-1275)




Dimethylphthalate
Concen: 11.28 ng
RT: 9.628 min Scan# 1282
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Tgt Ion:	163	Resp:	316071
Ion	Ratio	Lower	Upper
163	100		
194	3.6	3.1	4.7
164	10.7	8.2	12.2





Abundance Scan 1294 (9.698 min): BF123565.D\data.ms (-12851-)




2,6-Dinitrotoluene
Concen: 10.31 ng
RT: 9.686 min Scan# 1292
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:165 Resp: 65053
Ion Ratio Lower Upper
165 100
63 87.8 67.7 101.5
89 79.6 58.8 88.2



Abundance Scan 1292 (9.686 min): BF123563.D\data.ms (-1243-)





Abundance Scan 1337 (9.951 min): BF123565.D\data.ms (-13252-)




Acenaphthene
Concen: 12.10 ng
RT: 9.945 min Scan# 1336
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Tgt Ion:154 Resp: 272417
Ion Ratio Lower Upper
154 100
153 114.1 88.1 132.1
152 53.8 43.4 65.0



Abundance Scan 1336 (9.945 min): BF123563.D\data.ms (-1286-)



Abundance Scan 1323 (9.869 min): BF123565.D\data.ms (-1315)

3-Nitroaniline

Concen: 11.86 ng

RT: 9.857 min Scan# 1321

Delta R.T. -0.012 min

Lab File: BF123563.D


Acq: 15 Apr 2021 11:49

Instrument :

BNA_F

ClientSampleId :

SSTDIC010





Tgt Ion:138 Resp: 76481

Ion Ratio Lower Upper


138 100

108 12.2 10.2 15.4

92 133.2 103.3 154.9



Abundance Scan 1321 (9.857 min): BF123563.D\data.ms (-1272) (-)



Abundance Scan 1340 (9.969 min): BF123565.D\data.ms (-1335)

2,4-Dinitrophenol

Concen: 7.49 ng

RT: 9.957 min Scan# 1338

Delta R.T. -0.012 min

Lab File: BF123563.D

Acq: 15 Apr 2021 11:49




Tgt Ion:184 Resp: 18938

Ion Ratio Lower Upper

184 100

63 131.5 73.5 110.3#

154 199.9 52.5 78.7#




Abundance Scan 1338 (9.957 min): BF123563.D\data.ms (-1289) (-)






Abundance Scan 1366 (10.122 min): BF123565.D\data.ms (-1366) (-)

Dibenzofuran
Concen: 12.52 ng
RT: 10.116 min Scan# 1365
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	168	Resp:	411325
Ion	Ratio	Lower	Upper
168	100		
139	42.4	33.2	49.8
169	13.5	10.8	16.2







Abundance Scan 1349 (10.022 min): BF123565.D\data.ms (-1349) (-)

4-Nitrophenol
Concen: 11.40 ng
RT: 10.010 min Scan# 1347
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49



Tgt Ion:	139	Resp:	54787
Ion	Ratio	Lower	Upper
139	100		
109	77.0	63.2	103.2
65	114.6	110.0	150.0









Scan 1363 (10.104 min): BF123565.D\data.ms (-1357) (-)
 2,4-Dinitrotoluene
 Concen: 10.55 ng
 RT: 10.092 min Scan# 1361
 Delta R.T. -0.012 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Tgt Ion:	165	Resp:	86082
Ion	Ratio	Lower	Upper
165	100		
63	54.8	43.5	65.3
89	87.5	67.7	101.5
182	2.6	2.1	3.1



Scan 1425 (10.469 min): BF123565.D\data.ms (-1458) (-)
 Fluorene
 Concen: 13.23 ng
 RT: 10.463 min Scan# 1424
 Delta R.T. -0.006 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49


Tgt Ion:	166	Resp:	332788
Ion	Ratio	Lower	Upper
166	100		
165	91.4	73.5	110.3
167	12.4	10.7	16.1






Abundance Scan 1386 (10.239 min): BF123565.D\data.ms (-1358) (-)

2,3,4,6-Tetrachlorophenol
Concen: 9.81 ng
RT: 10.233 min Scan# 1385
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	232	Resp:	67537
Ion	Ratio	Lower	Upper
232	100		
131	52.5	43.2	64.8
130	2.6	2.2	3.2
166	35.1	27.3	40.9






Abundance Scan 1403 (10.339 min): BF123565.D\data.ms (-1359) (-)

Diethylphthalate
Concen: 12.47 ng
RT: 10.333 min Scan# 1402
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49




Tgt Ion:	149	Resp:	345422
Ion	Ratio	Lower	Upper
149	100		
177	20.6	15.4	23.2
150	12.2	9.9	14.9






Abundance Scan 1423 (10.457 min): BF123565.D\data.ms (-1418) (-)

4-Chlorophenyl-phenylether
Concen: 11.41 ng
RT: 10.457 min Scan# 1423
Delta R.T. -0.000 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	204	Resp:	150366
Ion	Ratio	Lower	Upper
204	100		
206	32.5	25.7	38.5
141	63.7	55.5	83.3







Abundance Scan 1428 (10.486 min): BF123565.D\data.ms (-1418) (-)

4-Nitroaniline
Concen: 12.94 ng
RT: 10.463 min Scan# 1424
Delta R.T. -0.024 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49



Tgt Ion:	138	Resp:	78195
Ion	Ratio	Lower	Upper
138	100		
92	56.5	35.8	75.8
108	82.8	63.6	103.6









Azobenzene
Concen: 13.04 ng
RT: 10.616 min Scan# 1450
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:	77	Resp:	392538
Ion	Ratio	Lower	Upper
77	100		
182	22.3	1.7	41.7
105	20.6	1.7	41.7
51	36.0	15.9	55.9



Phenanthrene-d10
Concen: 20.00 ng
RT: 11.404 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Tgt Ion:	188	Resp:	807744
Ion	Ratio	Lower	Upper
188	100		
94	8.8	6.9	10.3
80	10.6	8.2	12.2









4,6-Dinitro-2-methylphenol
Concen: 7.72 ng
RT: 10.498 min Scan# 1430
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:	198	Resp:	31882
Ion	Ratio	Lower	Upper
198	100		
51	76.8	64.8	104.8
105	59.8	42.6	82.6



n-Nitrosodiphenylamine
Concen: 12.10 ng
RT: 10.569 min Scan# 1442
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Tgt Ion:	169	Resp:	280555
Ion	Ratio	Lower	Upper
169	100		
168	64.2	53.3	79.9
167	36.3	29.9	44.9









4-Bromophenyl-phenylether
Concen: 11.05 ng
RT: 10.945 min Scan# 1506
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:	248	Resp:	91470
Ion	Ratio	Lower	Upper
248	100		
250	95.1	77.4	116.2
141	80.7	63.7	95.5



Hexachlorobenzene
Concen: 11.10 ng
RT: 11.010 min Scan# 1517
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Tgt Ion:	284	Resp:	100046
Ion	Ratio	Lower	Upper
284	100		
142	35.7	28.3	42.5
249	33.2	23.5	35.3









#59 (-)
Atrazine
 Concen: 11.06 ng
 RT: 11.092 min Scan# 1531
 Delta R.T. -0.012 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49




Instrument :
 BNA_F
ClientSampleId :
 SSTDICC010


Tgt Ion:	200	Resp:	86977
Ion	Ratio	Lower	Upper
200	100		
173	27.3	6.9	46.9
215	44.7	26.5	66.5



#70 (-)
Pentachlorophenol
 Concen: 8.89 ng
 RT: 11.204 min Scan# 1550
 Delta R.T. -0.006 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49

Tgt Ion:	266	Resp:	42036
Ion	Ratio	Lower	Upper
266	100		
268	55.3	51.6	77.4
264	61.4	52.6	78.8









#72 (-)
Phenanthrene
Concen: 12.10 ng
RT: 11.427 min Scan# 1588
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:	178	Resp:	479774
Ion	Ratio	Lower	Upper
178	100		
176	19.6	15.6	23.4
179	14.9	12.5	18.7



#73 (-)
Anthracene
Concen: 11.99 ng
RT: 11.474 min Scan# 1596
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Tgt Ion:	178	Resp:	474892
Ion	Ratio	Lower	Upper
178	100		
176	18.2	15.4	23.0
179	15.6	12.6	18.8









Carbazole
Concen: 12.58 ng
RT: 11.627 min Scan# 1622
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49




Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	167	Resp:	475371
Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.0	25.6
139	13.5	10.4	15.6



Di-n-butylphthalate
Concen: 11.37 ng
RT: 11.969 min Scan# 1680
Delta R.T. -0.000 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49


Tgt Ion:	149	Resp:	573067
Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.8	11.8
104	4.9	4.6	6.8






Abundance Scan 1791 (12.621 min): BF123565.D\data.ms (-1791) (-)

Fluoranthene
Concen: 12.62 ng
RT: 12.616 min Scan# 1790
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	202	Resp:	534182
Ion	Ratio	Lower	Upper
202	100		
101	10.7	0.0	30.9
203	17.7	0.0	37.5







Abundance Scan 2035 (14.057 min): BF123565.D\data.ms (-2035) (-)

Chrysene-d12
Concen: 20.00 ng
RT: 14.051 min Scan# 2034
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49



Tgt Ion:	240	Resp:	691221
Ion	Ratio	Lower	Upper
240	100		
120	10.4	7.2	10.8
236	24.3	19.0	28.4









Benzidine
 Concen: 12.35 ng
 RT: 12.739 min Scan# 1811
 Delta R.T. -0.000 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49




Instrument :
 BNA_F
ClientSampleId :
 SSTDICC010

Tgt Ion:	184	Resp:	229818
Ion	Ratio	Lower	Upper
184	100		
185	14.0	12.1	18.1
183	10.1	7.8	11.8



Pyrene
 Concen: 9.95 ng
 RT: 12.845 min Scan# 1829
 Delta R.T. -0.006 min
 Lab File: BF123563.D
 Acq: 15 Apr 2021 11:49


Tgt Ion:	202	Resp:	547643
Ion	Ratio	Lower	Upper
202	100		
200	19.5	16.1	24.1
203	17.2	14.8	22.2






Abundance Scan 1855 (12.998 min): BF123565.D\data.ms (-1879) (-)

Terphenyl-d14
Concen: 19.80 ng
RT: 12.992 min Scan# 1854
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010







Tgt Ion:	244	Resp:	787101
Ion	Ratio	Lower	Upper
244	100		
212	6.7	5.7	8.5
122	10.2	9.7	14.5



Abundance Scan 1936 (13.474 min): BF123565.D\data.ms (-1880) (-)

Butylbenzylphthalate
Concen: 8.79 ng
RT: 13.468 min Scan# 1935
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49


Tgt Ion:	149	Resp:	220158
Ion	Ratio	Lower	Upper
149	100		
91	82.1	63.6	95.4
206	18.2	14.5	21.7






Abundance Scan 2033 (14.045 min): BF123565.D\data.ms (-2033) (-)

Benzo(a)anthracene
Concen: 10.46 ng
RT: 14.039 min Scan# 2032
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	228	Resp:	451487
Ion	Ratio	Lower	Upper
228	100		
226	26.4	22.4	33.6
229	18.5	16.0	24.0







Abundance Scan 2027 (14.010 min): BF123565.D\data.ms (-2027) (-)

3,3'-Dichlorobenzidine
Concen: 10.94 ng
RT: 14.004 min Scan# 2026
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49



Tgt Ion:	252	Resp:	150648
Ion	Ratio	Lower	Upper
252	100		
254	67.2	51.1	76.7
126	8.8	7.7	11.5









Chrysene
Concen: 10.40 ng
RT: 14.074 min Scan# 2038
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49




Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	228	Resp:	449863
Ion	Ratio	Lower	Upper
228	100		
226	29.6	24.3	36.5
229	19.9	16.2	24.2



Bis(2-ethylhexyl)phthalate
Concen: 9.80 ng
RT: 14.033 min Scan# 2031
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49


Tgt Ion:	149	Resp:	323798
Ion	Ratio	Lower	Upper
149	100		
167	25.4	21.8	32.6
279	2.9	3.3	4.9#






Abundance Scan 2137 (14.657 min): BF123565.D\data.ms (-2137) (-)

Di-n-octyl phthalate
Concen: 10.45 ng
RT: 14.651 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010






Tgt Ion:	149	Resp:	501946
Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	15.0	11.8	17.6



Abundance Scan 2287 (15.539 min): BF123565.D\data.ms (-2287) (-)

Perylene-d12
Concen: 20.00 ng
RT: 15.539 min Scan# 2287
Delta R.T. -0.000 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49


Tgt Ion:	264	Resp:	609063
Ion	Ratio	Lower	Upper
264	100		
260	24.9	18.7	28.1
265	22.6	17.8	26.8






Abundance Scan 2542 (17.039 min): BF123565.D\data.ms (-2537) (-)

Indeno(1,2,3-cd)pyrene
Concen: 8.31 ng
RT: 17.027 min Scan# 2540
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	276	Resp:	357364
Ion	Ratio	Lower	Upper
276	100		
138	21.3	16.2	24.2
277	24.6	20.1	30.1






Abundance Scan 2214 (15.110 min): BF123565.D\data.ms (-2188) (-)

Benzo(b)fluoranthene
Concen: 10.14 ng
RT: 15.098 min Scan# 2212
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49




Tgt Ion:	252	Resp:	417012
Ion	Ratio	Lower	Upper
252	100		
253	21.9	18.1	27.1
125	8.9	6.2	9.4






Abundance Scan 2219 (15.139 min): BF123565.D\data.ms (-2168) (-)

Benzo(k)fluoranthene
Concen: 9.81 ng
RT: 15.127 min Scan# 2217
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	252	Resp:	398458
Ion	Ratio	Lower	Upper
252	100		
253	21.2	17.9	26.9
125	9.0	7.0	10.6






Abundance Scan 2277 (15.480 min): BF123565.D\data.ms (-2168) (-)

Benzo(a)pyrene
Concen: 10.16 ng
RT: 15.474 min Scan# 2276
Delta R.T. -0.006 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49




Tgt Ion:	252	Resp:	367587
Ion	Ratio	Lower	Upper
252	100		
253	21.8	18.4	27.6
125	9.9	8.0	12.0






Abundance Scan 2545 (17.057 min): BF123565.D\data.ms (-2545) (-)

Dibenzo(a,h)anthracene
Concen: 8.19 ng
RT: 17.045 min Scan# 2543
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	278	Resp:	301522
Ion	Ratio	Lower	Upper
278	100		
139	13.8	10.2	15.2
279	24.2	19.4	29.0



Abundance Scan 2618 (17.486 min): BF123565.D\data.ms (-2618) (-)

Benzo(g,h,i)perylene
Concen: 8.44 ng
RT: 17.474 min Scan# 2616
Delta R.T. -0.012 min
Lab File: BF123563.D
Acq: 15 Apr 2021 11:49



Tgt Ion:	276	Resp:	288431
Ion	Ratio	Lower	Upper
276	100		
277	24.5	18.2	27.2
138	18.2	13.8	20.6

