

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041521\
 Data File : BF123566.D
 Acq On : 15 Apr 2021 13:27
 Operator : JU/CG
 Sample : SSTDICC050
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Quant Time: Apr 15 13:50:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 15 13:22:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.869	152	241870	20.00	ng	0.00
21) Naphthalene-d8	8.157	136	859048	20.00	ng	0.00
39) Acenaphthene-d10	9.916	164	454219	20.00	ng	0.00
64) Phenanthrene-d10	11.404	188	846328	20.00	ng	0.00
76) Chrysene-d12	14.057	240	607830	20.00	ng	0.00
86) Perylene-d12	15.539	264	457990	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	1358749	95.93	ng	0.00
7) Phenol-d6	6.504	99	1887391	97.76	ng	0.00
23) Nitrobenzene-d5	7.439	82	1809666	105.21	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	510631	115.14	ng	0.00
45) 2-Fluorobiphenyl	9.239	172	2737044	98.00	ng	0.00
79) Terphenyl-d14	12.998	244	3092681	88.46	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.616	88	367358	31.84	ng	97
3) Pyridine	3.369	79	886276	26.67	ng	98
4) n-Nitrosodimethylamine	3.328	42	524815	40.90	ng	100
6) Aniline	6.534	93	1128516	46.84	ng	99
8) 2-Chlorophenol	6.657	128	737079	48.56	ng	97
9) Benzaldehyde	6.422	77	502130	41.00	ng	98
10) Phenol	6.516	94	1031230	50.24	ng	98
11) bis(2-Chloroethyl)ether	6.610	93	745851	45.72	ng	97
12) 1,3-Dichlorobenzene	6.816	146	743583	45.14	ng	96
13) 1,4-Dichlorobenzene	6.892	146	756914	45.48	ng	97
14) 1,2-Dichlorobenzene	7.045	146	710970	45.70	ng	98
15) Benzyl Alcohol	7.016	79	787390	52.96	ng	96
16) 2,2'-oxybis(1-Chloropr...	7.151	45	1700935	54.99	ng	99
17) 2-Methylphenol	7.122	107	644391	50.04	ng	98
18) Hexachloroethane	7.386	117	313266	47.42	ng	96
19) n-Nitroso-di-n-propyla...	7.292	70	627770	48.87	ng	98
20) 3+4-Methylphenols	7.281	107	805809	51.11	ng	# 86
22) Acetophenone	7.286	105	1130473	50.82	ng	# 93
24) Nitrobenzene	7.457	77	947440	50.91	ng	95
25) Isophorone	7.698	82	1736804	50.13	ng	99
26) 2-Nitrophenol	7.769	139	369362	47.18	ng	95
27) 2,4-Dimethylphenol	7.810	122	588018	48.24	ng	98
28) bis(2-Chloroethoxy)met...	7.904	93	874457	47.14	ng	99
29) 2,4-Dichlorophenol	8.016	162	588517	45.48	ng	98
30) 1,2,4-Trichlorobenzene	8.098	180	615314	41.84	ng	100
31) Naphthalene	8.181	128	2078360	48.79	ng	100
32) Benzoic acid	7.939	122	387519	48.03	ng	90
33) 4-Chloroaniline	8.228	127	872550	49.28	ng	98
34) Hexachlorobutadiene	8.298	225	384685	41.03	ng	98
35) Caprolactam	8.610	113	210236	50.51	ng	96
36) 4-Chloro-3-methylphenol	8.710	107	769298	54.33	ng	94
37) 2-Methylnaphthalene	8.875	142	1366130	46.41	ng	99
38) 1-Methylnaphthalene	8.975	142	1281036	46.29	ng	100
40) 1,2,4,5-Tetrachloroben...	9.039	216	592204	45.52	ng	99
41) Hexachlorocyclopentadiene	9.028	237	338223	51.80	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041521\
 Data File : BF123566.D
 Acq On : 15 Apr 2021 13:27
 Operator : JU/CG
 Sample : SSTDICC050
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050


Quant Time: Apr 15 13:50:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 15 13:22:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	9.151	196	416844	46.46	ng	99
44) 2,4,5-Trichlorophenol	9.186	196	441314	45.64	ng	98
46) 1,1'-Biphenyl	9.339	154	1646662	53.02	ng	98
47) 2-Chloronaphthalene	9.363	162	1242741	49.46	ng	98
48) 2-Nitroaniline	9.457	65	551659	62.33	ng	93
49) Acenaphthylene	9.780	152	2020795	53.22	ng	98
50) Dimethylphthalate	9.645	163	1450749	47.69	ng	100
51) 2,6-Dinitrotoluene	9.704	165	342645	50.03	ng	# 82
52) Acenaphthene	9.957	154	1329358	54.37	ng	98
53) 3-Nitroaniline	9.869	138	378627	54.08	ng	90
54) 2,4-Dinitrophenol	9.975	184	163920	59.74	ng	88
55) Dibenzofuran	10.127	168	1846230	51.76	ng	100
56) 4-Nitrophenol	10.027	139	312675	59.92	ng	94
57) 2,4-Dinitrotoluene	10.104	165	455650	51.44	ng	87
58) Fluorene	10.469	166	1422849	52.11	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.239	232	371718	49.72	ng	99
60) Diethylphthalate	10.345	149	1594857	53.02	ng	99
61) 4-Chlorophenyl-phenyle...	10.463	204	674339	47.15	ng	95
62) 4-Nitroaniline	10.486	138	390520	59.55	ng	94
63) Azobenzene	10.622	77	1819851	55.70	ng	97
65) 4,6-Dinitro-2-methylph...	10.516	198	223905	51.73	ng	95
66) n-Nitrosodiphenylamine	10.580	169	1283063	52.81	ng	99
67) 4-Bromophenyl-phenylether	10.951	248	426223	49.13	ng	98
68) Hexachlorobenzene	11.022	284	478999	50.71	ng	# 91
69) Atrazine	11.110	200	395533	48.00	ng	96
70) Pentachlorophenol	11.210	266	282781	57.08	ng	98
71) Phenanthrene	11.433	178	2031304	48.91	ng	99
72) Anthracene	11.486	178	2037093	49.08	ng	99
73) Carbazole	11.639	167	2035040	51.41	ng	98
74) Di-n-butylphthalate	11.974	149	2498752	47.32	ng	99
75) Fluoranthene	12.627	202	2230304	50.28	ng	99
77) Benzidine	12.745	184	907766	55.48	ng	99
78) Pyrene	12.857	202	2261470	46.74	ng	99
80) Butylbenzylphthalate	13.474	149	1041459	47.31	ng	99
81) Benzo(a)anthracene	14.045	228	1759085	46.33	ng	99
82) 3,3'-Dichlorobenzidine	14.010	252	547894	45.24	ng	96
83) Chrysene	14.086	228	1787959	47.03	ng	99
84) Bis(2-ethylhexyl)phtha...	14.039	149	1331127	45.84	ng	99
85) Di-n-octyl phthalate	14.657	149	2116383	50.09	ng	98
87) Indeno(1,2,3-cd)pyrene	17.039	276	1222226	37.81	ng	99
88) Benzo(b)fluoranthene	15.109	252	1438478	46.54	ng	99
89) Benzo(k)fluoranthene	15.139	252	1442183	47.21	ng	99
90) Benzo(a)pyrene	15.480	252	1256457	46.20	ng	97
91) Dibenzo(a,h)anthracene	17.056	278	1047328	37.83	ng	98
92) Benzo(g,h,i)perylene	17.492	276	1047344	40.77	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Apr 15 13:50:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF041521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 15 13:22:17 2021
Response via : Initial Calibration








1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.869 min Scan# 813
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27


Instrument :
BNA_F
ClientSampleId :
SSTDIC050





Tgt Ion:152 Resp: 241870
Ion Ratio Lower Upper
152 100
150 152.5 127.0 190.6
115 65.0 54.3 81.5




1,4-Dioxane
Concen: 31.84 ng
RT: 2.616 min Scan# 90
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27



Tgt Ion: 88 Resp: 367358
Ion Ratio Lower Upper
88 100
58 101.5 83.8 125.8
43 39.7 32.2 48.4









Pyridine
Concen: 26.67 ng
RT: 3.369 min Scan# 218
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27




Instrument :
BNA_F
ClientSampleId :
SSTDIC050


Tgt Ion: 79 Resp: 886276
Ion Ratio Lower Upper
79 100
52 106.7 87.4 131.0
51 51.8 41.4 62.0



n-Nitrosodimethylamine
Concen: 40.90 ng
RT: 3.328 min Scan# 211
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion: 42 Resp: 524815
Ion Ratio Lower Upper
42 100
74 104.1 83.5 125.3
44 7.7 5.7 8.5









2-Fluorophenol
Concen: 95.93 ng
RT: 5.481 min Scan# 577
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tgt Ion: 112 Resp: 1358749




Ion	Ratio	Lower	Upper
112	100		
64	81.1	61.4	92.0
63	43.2	33.0	49.4




Aniline
Concen: 46.84 ng
RT: 6.534 min Scan# 756
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion: 93 Resp: 1128516

Ion	Ratio	Lower	Upper
93	100		
66	46.2	36.4	54.6
65	25.0	19.4	29.0









Phenol-d6
 Concen: 97.76 ng
 RT: 6.504 min Scan# 751
 Delta R.T. 0.006 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Tgt Ion: 99 Resp: 1887391




Ion	Ratio	Lower	Upper
99	100		
42	24.7	19.8	29.8
71	40.0	31.1	46.7




2-Chlorophenol
 Concen: 48.56 ng
 RT: 6.657 min Scan# 777
 Delta R.T. -0.000 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27

Tgt Ion: 128 Resp: 737079

Ion	Ratio	Lower	Upper
128	100		
130	33.4	12.6	52.6
64	64.9	41.5	81.5









Benzaldehyde
Concen: 41.00 ng
RT: 6.422 min Scan# 737
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27




Instrument :
BNA_F
ClientSampleId :
SSTDIC050


Tgt Ion:	77	Resp:	502130
Ion	Ratio	Lower	Upper
77	100		
105	97.6	75.2	115.2
106	92.9	71.6	111.6



Phenol
Concen: 50.24 ng
RT: 6.516 min Scan# 753
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	94	Resp:	1031230
Ion	Ratio	Lower	Upper
94	100		
65	37.2	18.1	58.1
66	50.1	29.0	69.0









#1
 bis(2-Chloroethyl)ether
 Concen: 45.72 ng
 RT: 6.610 min Scan# 769
 Delta R.T. 0.006 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27




Instrument :
 BNA_F
 ClientSampled :
 SSTDICC050

Tgt Ion:	93	Resp:	745851
Ion	Ratio	Lower	Upper
93	100		
63	93.6	76.1	116.1
95	34.0	11.5	51.5



#12
 1,3-Dichlorobenzene
 Concen: 45.14 ng
 RT: 6.816 min Scan# 804
 Delta R.T. 0.006 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27


Tgt Ion:	146	Resp:	743583
Ion	Ratio	Lower	Upper
146	100		
148	65.2	50.9	76.3
75	37.9	34.2	51.4





Abundance Scan 816 (6.887 min): BF123565.D\data.ms (-811) #13

1,4-Dichlorobenzene
Concen: 45.48 ng
RT: 6.892 min Scan# 817
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27


Instrument :
BNA_F
ClientSampled :
SSTDICC050



Tgt Ion:	146	Resp:	756914
Ion	Ratio	Lower	Upper
146	100		
148	64.6	49.9	74.9
111	46.5	39.5	59.3





Abundance Scan 817 (6.892 min): BF123566.D\data.ms (-765) (-)




Abundance Scan 843 (7.045 min): BF123565.D\data.ms (-837) #14


1,2-Dichlorobenzene
Concen: 45.70 ng
RT: 7.045 min Scan# 843
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	146	Resp:	710970
Ion	Ratio	Lower	Upper
146	100		
148	65.0	51.6	77.4
111	51.2	38.4	57.6



Abundance Scan 843 (7.045 min): BF123566.D\data.ms (-792) (-)









Benzyl Alcohol
 Concen: 52.96 ng
 RT: 7.016 min Scan# 838
 Delta R.T. 0.006 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Tgt Ion:	79	Resp:	787390
Ion	Ratio	Lower	Upper
79	100		
108	74.5	55.8	83.8
77	63.4	49.9	74.9



2,2'-oxybis(1-Chloropropane)
 Concen: 54.99 ng
 RT: 7.151 min Scan# 861
 Delta R.T. 0.006 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27



Tgt Ion:	45	Resp:	1700935
Ion	Ratio	Lower	Upper
45	100		
77	9.7	0.0	30.0
79	7.7	0.0	27.3





Abundance Scan 856 (7.122 min): BF123565.D\data.ms (-850) #17

2-Methylphenol
Concen: 50.04 ng
RT: 7.122 min Scan# 856
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampled :
SSTDIC050





Tgt Ion:	107	Resp:	644391
Ion	Ratio	Lower	Upper
107	100		
108	113.6	89.8	134.8
77	52.4	40.9	61.3
79	52.8	40.2	60.4






Abundance Scan 901 (7.387 min): BF123565.D\data.ms (-896) #18

Hexachloroethane
Concen: 47.42 ng
RT: 7.386 min Scan# 901
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27



Tgt Ion:	117	Resp:	313266
Ion	Ratio	Lower	Upper
117	100		
119	98.0	77.8	116.8
201	77.2	67.8	101.8









#19
 n-Nitroso-di-n-propylamine
 Concen: 48.87 ng
 RT: 7.292 min Scan# 885
 Delta R.T. 0.006 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Tgt Ion:	70	Resp:	627770
Ion	Ratio	Lower	Upper
70	100		
42	72.9	60.1	90.1
101	10.0	7.4	11.2
130	18.6	14.2	21.2




#20
 3+4-Methylphenols
 Concen: 51.11 ng
 RT: 7.281 min Scan# 883
 Delta R.T. 0.006 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27

Tgt Ion:	107	Resp:	805809
Ion	Ratio	Lower	Upper
107	100		
108	93.7	71.5	111.5
77	129.7	82.0	122.0#
79	35.6	14.1	54.1






Abundance Scan 1032 (8.157 min): BF123565.D\data.ms (-1021)




Naphthalene-d8
Concen: 20.00 ng
RT: 8.157 min Scan# 1032
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tgt Ion:	136	Resp:	859048
Ion	Ratio	Lower	Upper
136	100		
137	11.6	8.7	13.1
54	12.7	9.8	14.6
68	5.7	4.1	6.1






Abundance Scan 883 (7.281 min): BF123565.D\data.ms (-876) #2




Acetophenone
Concen: 50.82 ng
RT: 7.286 min Scan# 884
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	105	Resp:	1130473
Ion	Ratio	Lower	Upper
105	100		
71	7.5	4.2	6.4#
51	36.0	33.6	50.4
120	19.9	16.3	24.5






Abundance Scan 909 (7.434 min): BF123565.D\data.ms (-904) #23




Nitrobenzene-d5
Concen: 105.21 ng
RT: 7.439 min Scan# 910
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:	82	Resp:	1809666
Ion	Ratio	Lower	Upper
82	100		
128	37.0	29.4	44.2
54	66.6	54.1	81.1






Abundance Scan 913 (7.457 min): BF123565.D\data.ms (-905) #24



Nitrobenzene
Concen: 50.91 ng
RT: 7.457 min Scan# 913
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27


Tgt Ion:	77	Resp:	947440
Ion	Ratio	Lower	Upper
77	100		
123	39.3	34.5	51.7
65	16.3	12.2	18.2






Abundance Scan 953 (7.692 min): BF123565.D\data.ms (-947) #25

Isophorone
Concen: 50.13 ng
RT: 7.698 min Scan# 954
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050




Tgt Ion:	82	Resp:	1736804
Ion	Ratio	Lower	Upper
82	100		
95	7.4	5.8	8.6
138	15.4	11.8	17.8






Abundance Scan 966 (7.769 min): BF123565.D\data.ms (-961) #26

2-Nitrophenol
Concen: 47.18 ng
RT: 7.769 min Scan# 966
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27




Tgt Ion:	139	Resp:	369362
Ion	Ratio	Lower	Upper
139	100		
109	37.4	28.6	43.0
65	76.6	56.8	85.2






Abundance Scan 972 (7.804 min): BF123565.D\data.ms (-968) #27

2,4-Dimethylphenol
Concen: 48.24 ng
RT: 7.810 min Scan# 973
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDICC050




Tgt Ion:	122	Resp:	588018
Ion	Ratio	Lower	Upper
122	100		
107	130.9	106.3	159.5
121	58.6	45.4	68.0






Abundance Scan 989 (7.904 min): BF123565.D\data.ms (-980) #28


bis(2-Chloroethoxy)methane
Concen: 47.14 ng
RT: 7.904 min Scan# 989
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27



Tgt Ion:	93	Resp:	874457
Ion	Ratio	Lower	Upper
93	100		
95	32.6	26.5	39.7
123	10.9	8.4	12.6






Abundance Scan 1007 (8.010 min): BF123565.D\data.ms (-1029)




2,4-Dichlorophenol
Concen: 45.48 ng
RT: 8.016 min Scan# 1008
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tgt Ion:	162	Resp:	588517
Ion	Ratio	Lower	Upper
162	100		
164	64.7	44.1	84.1
98	44.5	27.3	67.3






Abundance Scan 1022 (8.098 min): BF123565.D\data.ms (-1030)




1,2,4-Trichlorobenzene
Concen: 41.84 ng
RT: 8.098 min Scan# 1022
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	180	Resp:	615314
Ion	Ratio	Lower	Upper
180	100		
182	96.4	77.2	115.8
145	33.2	26.0	39.0






Abundance Scan 1036 (8.181 min): BF123565.D\data.ms (-10281)




Naphthalene
Concen: 48.79 ng
RT: 8.181 min Scan# 1036
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tgt Ion:	128	Resp:	2078360
Ion	Ratio	Lower	Upper
128	100		
129	11.2	8.9	13.3
127	13.7	10.8	16.2






Abundance Scan 992 (7.922 min): BF123565.D\data.ms (-9777)




Benzoic acid
Concen: 48.03 ng
RT: 7.939 min Scan# 995
Delta R.T. 0.018 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	122	Resp:	387519
Ion	Ratio	Lower	Upper
122	100		
105	142.4	107.8	147.8
77	108.2	81.6	121.6






Abundance Scan 1043 (8.222 min): BF123565.D\data.ms (-1033)




4-Chloroaniline
Concen: 49.28 ng
RT: 8.228 min Scan# 1044
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tgt Ion:	127	Resp:	872550
Ion	Ratio	Lower	Upper
127	100		
129	33.1	25.7	38.5
65	41.7	34.5	51.7
92	22.2	17.4	26.0






Abundance Scan 1056 (8.298 min): BF123565.D\data.ms (-1034)




Hexachlorobutadiene
Concen: 41.03 ng
RT: 8.298 min Scan# 1056
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	225	Resp:	384685
Ion	Ratio	Lower	Upper
225	100		
223	64.4	50.4	75.6
227	65.9	51.4	77.0






Abundance Scan 1107 (8.598 min): BF123565.D\data.ms (-11035)




Caprolactam
Concen: 50.51 ng
RT: 8.610 min Scan# 1109
Delta R.T. 0.012 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tgt Ion:113 Resp: 210236
Ion Ratio Lower Upper
113 100
55 269.5 254.9 294.9
56 208.3 179.4 219.4






Abundance Scan 1125 (8.704 min): BF123565.D\data.ms (-11036)



4-Chloro-3-methylphenol
Concen: 54.33 ng
RT: 8.710 min Scan# 1126
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:107 Resp: 769298
Ion Ratio Lower Upper
107 100
144 21.0 16.3 24.5
142 68.5 50.1 75.1



Abundance Scan 1153 (8.869 min): BF123565.D\data.ms (-1102107)

2-Methylnaphthalene

Concen: 46.41 ng

RT: 8.875 min Scan# 1154

Delta R.T. 0.006 min

Lab File: BF123566.D

Acq: 15 Apr 2021 13:27

Tgt Ion:142 Resp: 1366130

Ion Ratio Lower Upper

142 100

141 90.6 71.8 107.6

Instrument :

BNA_F

ClientSampled :

SSTDIC050

Abundance Scan 1154 (8.875 min): BF123566.D\data.ms

Ref 50

m/z-->

Abundance Scan 1154 (8.875 min): BF123566.D\data.ms

Raw 50

m/z-->

Abundance Scan 1154 (8.875 min): BF123566.D\data.ms (-1102107)

Sub 50

m/z-->

Abundance

1500000

8.875

1000000

500000

0

Time-->

8.80 8.85 8.90 8.95

Abundance Scan 1171 (8.975 min): BF123565.D\data.ms (-1102108)

1-Methylnaphthalene

Concen: 46.29 ng

RT: 8.975 min Scan# 1171

Delta R.T. -0.000 min

Lab File: BF123566.D

Acq: 15 Apr 2021 13:27

Tgt Ion:142 Resp: 1281036

Ion Ratio Lower Upper

142 100

141 91.4 72.8 109.2

Abundance Scan 1171 (8.975 min): BF123566.D\data.ms

Ref 50

m/z-->

Abundance Scan 1171 (8.975 min): BF123566.D\data.ms

Raw 50

m/z-->

Abundance Scan 1171 (8.975 min): BF123566.D\data.ms (-1102109)

Sub 50

m/z-->

Abundance

1500000

8.975

1000000


500000

0

Time-->

8.90 8.95 9.00




Abundance Scan 1331 (9.916 min): BF123565.D\data.ms (-1329)




Acenaphthene-d10
Concen: 20.00 ng
RT: 9.916 min Scan# 1331
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tgt Ion:	164	Resp:	454219
Ion	Ratio	Lower	Upper
164	100		
162	101.2	80.7	121.1
160	45.5	37.3	55.9






Abundance Scan 1182 (9.039 min): BF123565.D\data.ms (-1179)




1,2,4,5-Tetrachlorobenzene
Concen: 45.52 ng
RT: 9.039 min Scan# 1182
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	216	Resp:	592204
Ion	Ratio	Lower	Upper
216	100		
214	79.6	63.8	95.8
179	22.5	17.9	26.9
108	22.2	18.6	28.0






Abundance Scan 1180 (9.028 min): BF123565.D\data.ms (-11741-)




Hexachlorocyclopentadiene
Concen: 51.80 ng
RT: 9.028 min Scan# 1180
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tgt Ion:	237	Resp:	338223
Ion	Ratio	Lower	Upper
237	100		
235	56.4	38.8	78.8
272	12.0	0.0	32.4






Abundance Scan 1466 (10.710 min): BF123565.D\data.ms (-1460-)




2,4,6-Tribromophenol
Concen: 115.14 ng
RT: 10.710 min Scan# 1466
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	330	Resp:	510631
Ion	Ratio	Lower	Upper
330	100		
332	95.5	77.1	115.7
141	33.8	27.9	41.9






Abundance Scan 1200 (9.145 min): BF123565.D\data.ms (-1195.1)




2,4,6-Trichlorophenol
Concen: 46.46 ng
RT: 9.151 min Scan# 1201
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tgt Ion:	196	Resp:	416844
Ion	Ratio	Lower	Upper
196	100		
198	96.9	78.2	117.2
200	32.3	25.0	37.6






Abundance Scan 1207 (9.186 min): BF123565.D\data.ms (-1206.1)




2,4,5-Trichlorophenol
Concen: 45.64 ng
RT: 9.186 min Scan# 1207
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	196	Resp:	441314
Ion	Ratio	Lower	Upper
196	100		
198	96.2	76.5	114.7
97	68.2	52.9	79.3
132	33.1	25.3	37.9






Abundance Scan 1216 (9.239 min): BF123565.D\data.ms (-12145)




2-Fluorobiphenyl
Concen: 98.00 ng
RT: 9.239 min Scan# 1216
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tgt Ion:172 Resp: 2737044
Ion Ratio Lower Upper
172 100
171 36.0 28.1 42.1
170 24.8 19.2 28.8






Abundance Scan 1233 (9.339 min): BF123565.D\data.ms (-122746)




1,1'-Biphenyl
Concen: 53.02 ng
RT: 9.339 min Scan# 1233
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:154 Resp: 1646662
Ion Ratio Lower Upper
154 100
153 39.8 20.7 60.7
76 12.5 0.0 31.7






Abundance Scan 1237 (9.363 min): BF123565.D\data.ms (-1237)




2-Chloronaphthalene
Concen: 49.46 ng
RT: 9.363 min Scan# 1237
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tgt Ion:162 Resp: 1242741
Ion Ratio Lower Upper
162 100
127 46.1 36.2 54.2
164 33.3 25.4 38.2






Abundance Scan 1253 (9.457 min): BF123565.D\data.ms (-1248)




2-Nitroaniline
Concen: 62.33 ng
RT: 9.457 min Scan# 1253
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion: 65 Resp: 551659
Ion Ratio Lower Upper
65 100
92 57.8 47.1 70.7
138 77.9 70.2 105.2






Abundance Scan 1308 (9.781 min): BF123565.D\data.ms (-123565)




Acenaphthylene
Concen: 53.22 ng
RT: 9.780 min Scan# 1308
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tgt Ion:	152	Resp:	2020795
Ion	Ratio	Lower	Upper
152	100		
151	21.0	16.0	24.0
153	13.4	10.4	15.6






Abundance Scan 1284 (9.639 min): BF123565.D\data.ms (-123565)




Dimethylphthalate
Concen: 47.69 ng
RT: 9.645 min Scan# 1285
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	163	Resp:	1450749
Ion	Ratio	Lower	Upper
163	100		
194	3.8	3.1	4.7
164	10.1	8.2	12.2






Abundance Scan 1294 (9.698 min): BF123565.D\data.ms (-12851-)




2,6-Dinitrotoluene
Concen: 50.03 ng
RT: 9.704 min Scan# 1295
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampled :
SSTDIC050

Tgt Ion:165 Resp: 342645
Ion Ratio Lower Upper
165 100
63 64.9 67.7 101.5#
89 61.9 58.8 88.2






Abundance Scan 1337 (9.951 min): BF123565.D\data.ms (-13252-)




Acenaphthene
Concen: 54.37 ng
RT: 9.957 min Scan# 1338
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:154 Resp: 1329358
Ion Ratio Lower Upper
154 100
153 112.5 88.1 132.1
152 54.4 43.4 65.0




Abundance Scan 1323 (9.869 min): BF123565.D\data.ms (-1333)





3-Nitroaniline
Concen: 54.08 ng
RT: 9.869 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27


Instrument :
BNA_F
ClientSampleId :
SSTDIC050




Tgt Ion:	138	Resp:	378627
Ion	Ratio	Lower	Upper
138	100		
108	12.9	10.2	15.4
92	142.0	103.3	154.9





Abundance Scan 1340 (9.969 min): BF123565.D\data.ms (-1335)



2,4-Dinitrophenol
Concen: 59.74 ng
RT: 9.975 min Scan# 1341
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27




Tgt Ion:	184	Resp:	163920
Ion	Ratio	Lower	Upper
184	100		
63	75.3	73.5	110.3
154	68.5	52.5	78.7





Abundance Scan 1366 (10.122 min): BF123565.D\data.ms (-1315) (-)

Dibenzofuran
Concen: 51.76 ng
RT: 10.127 min Scan# 1367
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27


Instrument :
BNA_F
ClientSampleId :
SSTDIC050



Tgt Ion:	168	Resp:	1846230
Ion	Ratio	Lower	Upper
168	100		
139	41.5	33.2	49.8
169	13.8	10.8	16.2




Abundance Scan 1367 (10.127 min): BF123566.D\data.ms (-1315) (-)





Abundance Scan 1349 (10.022 min): BF123565.D\data.ms (-1356) (-)


4-Nitrophenol
Concen: 59.92 ng
RT: 10.027 min Scan# 1350
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27




Tgt Ion:	139	Resp:	312675
Ion	Ratio	Lower	Upper
139	100		
109	84.0	63.2	103.2
65	118.6	110.0	150.0



Abundance Scan 1350 (10.027 min): BF123566.D\data.ms (-1298) (-)









2,4-Dinitrotoluene
 Concen: 51.44 ng
 RT: 10.104 min Scan# 1363
 Delta R.T. -0.000 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Tgt Ion:	165	Resp:	455650
Ion	Ratio	Lower	Upper
165	100		
63	65.3	43.5	65.3
89	94.9	67.7	101.5
182	2.1	2.1	3.1



Fluorene
 Concen: 52.11 ng
 RT: 10.469 min Scan# 1425
 Delta R.T. -0.000 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27


Tgt Ion:	166	Resp:	1422849
Ion	Ratio	Lower	Upper
166	100		
165	92.1	73.5	110.3
167	13.8	10.7	16.1






Abundance Scan 1386 (10.239 min): BF123565.D\data.ms (-1335) (-)

2,3,4,6-Tetrachlorophenol
Concen: 49.72 ng
RT: 10.239 min Scan# 1386
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050




Tgt Ion:	232	Resp:	371718
Ion	Ratio	Lower	Upper
232	100		
131	53.3	43.2	64.8
130	2.7	2.2	3.2
166	33.7	27.3	40.9






Abundance Scan 1403 (10.339 min): BF123565.D\data.ms (-1350) (-)

Diethylphthalate
Concen: 53.02 ng
RT: 10.345 min Scan# 1404
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27




Tgt Ion:	149	Resp:	1594857
Ion	Ratio	Lower	Upper
149	100		
177	20.0	15.4	23.2
150	12.6	9.9	14.9






Abundance Scan 1423 (10.457 min): BF123565.D\data.ms (-1458) (-)

4-Chlorophenyl-phenylether
Concen: 47.15 ng
RT: 10.463 min Scan# 1424
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050




Tgt Ion:	204	Resp:	674339
Ion	Ratio	Lower	Upper
204	100		
206	33.0	25.7	38.5
141	63.7	55.5	83.3







Abundance Scan 1428 (10.486 min): BF123565.D\data.ms (-1459) (-)

4-Nitroaniline
Concen: 59.55 ng
RT: 10.486 min Scan# 1428
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27



Tgt Ion:	138	Resp:	390520
Ion	Ratio	Lower	Upper
138	100		
92	59.1	35.8	75.8
108	90.4	63.6	103.6









Azobenzene
Concen: 55.70 ng
RT: 10.622 min Scan# 1451
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tgt Ion: 77 Resp: 1819851




Ion	Ratio	Lower	Upper
77	100		
182	18.9	1.7	41.7
105	20.8	1.7	41.7
51	36.7	15.9	55.9




Phenanthrene-d10
Concen: 20.00 ng
RT: 11.404 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion: 188 Resp: 846328

Ion	Ratio	Lower	Upper
188	100		
94	9.8	6.9	10.3
80	11.0	8.2	12.2









4,6-Dinitro-2-methylphenol
Concen: 51.73 ng
RT: 10.516 min Scan# 1433
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27




Instrument :
BNA_F
ClientSampleId :
SSTDIC050


Tgt Ion:	198	Resp:	223905
Ion	Ratio	Lower	Upper
198	100		
51	78.1	64.8	104.8
105	60.9	42.6	82.6



n-Nitrosodiphenylamine
Concen: 52.81 ng
RT: 10.580 min Scan# 1444
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	169	Resp:	1283063
Ion	Ratio	Lower	Upper
169	100		
168	67.1	53.3	79.9
167	36.2	29.9	44.9









4-Bromophenyl-phenylether
Concen: 49.13 ng
RT: 10.951 min Scan# 1507
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27




Instrument :
BNA_F
ClientSampleId :
SSTDIC050


Tgt Ion:	248	Resp:	426223
Ion	Ratio	Lower	Upper
248	100		
250	96.8	77.4	116.2
141	82.9	63.7	95.5



Hexachlorobenzene
Concen: 50.71 ng
RT: 11.022 min Scan# 1519
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	284	Resp:	478999
Ion	Ratio	Lower	Upper
284	100		
142	27.1	28.3	42.5#
249	28.3	23.5	35.3









Scan 1533 (11.104 min): BF123565.D\data.ms (-1559) (-)
Atrazine
 Concen: 48.00 ng
 RT: 11.110 min Scan# 1534
 Delta R.T. 0.006 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27




Instrument :
 BNA_F
ClientSampleId :
 SSTDICC050

Tgt Ion:	200	Resp:	395533
Ion	Ratio	Lower	Upper
200	100		
173	27.3	6.9	46.9
215	50.9	26.5	66.5



Scan 1551 (11.210 min): BF123565.D\data.ms (-1570) (-)
Pentachlorophenol
 Concen: 57.08 ng
 RT: 11.210 min Scan# 1551
 Delta R.T. -0.000 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27


Tgt Ion:	266	Resp:	282781
Ion	Ratio	Lower	Upper
266	100		
268	64.4	51.6	77.4
264	63.0	52.6	78.8






Abundance Scan 1589 (11.433 min): BF123565.D\data.ms (-1572) (-)

Phenanthrene
Concen: 48.91 ng
RT: 11.433 min Scan# 1589
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050




Tgt Ion:	178	Resp:	2031304
Ion	Ratio	Lower	Upper
178	100		
176	20.4	15.6	23.4
179	15.7	12.5	18.7







Abundance Scan 1598 (11.486 min): BF123565.D\data.ms (-1573) (-)

Anthracene
Concen: 49.08 ng
RT: 11.486 min Scan# 1598
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27



Tgt Ion:	178	Resp:	2037093
Ion	Ratio	Lower	Upper
178	100		
176	19.7	15.4	23.0
179	15.9	12.6	18.8









#173 (-)
 Carbazole
 Concen: 51.41 ng
 RT: 11.639 min Scan# 1624
 Delta R.T. -0.000 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Tgt Ion:	167	Resp:	2035040
Ion	Ratio	Lower	Upper
167	100		
166	22.0	17.0	25.6
139	14.0	10.4	15.6



#174 (-)
 Di-n-butylphthalate
 Concen: 47.32 ng
 RT: 11.974 min Scan# 1681
 Delta R.T. 0.006 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27


Tgt Ion:	149	Resp:	2498752
Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.8	11.8
104	5.6	4.6	6.8






Abundance Scan 1791 (12.621 min): BF123565.D\data.ms (-1791) (-)

Fluoranthene
Concen: 50.28 ng
RT: 12.627 min Scan# 1792
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050




Tgt Ion:	202	Resp:	2230304
Ion	Ratio	Lower	Upper
202	100		
101	9.8	0.0	30.9
203	17.3	0.0	37.5







Abundance Scan 2035 (14.057 min): BF123565.D\data.ms (-2035) (-)

Chrysene-d12
Concen: 20.00 ng
RT: 14.057 min Scan# 2035
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27



Tgt Ion:	240	Resp:	607830
Ion	Ratio	Lower	Upper
240	100		
120	10.2	7.2	10.8
236	23.0	19.0	28.4









Benzidine
 Concen: 55.48 ng
 RT: 12.745 min Scan# 1812
 Delta R.T. 0.006 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27




Instrument :
 BNA_F
ClientSampleId :
 SSTDICC050

Tgt Ion:	184	Resp:	907766
Ion	Ratio	Lower	Upper
184	100		
185	14.6	12.1	18.1
183	9.9	7.8	11.8



Pyrene
 Concen: 46.74 ng
 RT: 12.857 min Scan# 1831
 Delta R.T. 0.006 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27


Tgt Ion:	202	Resp:	2261470
Ion	Ratio	Lower	Upper
202	100		
200	20.0	16.1	24.1
203	17.7	14.8	22.2






Abundance Scan 1855 (12.998 min): BF123565.D\data.ms (-1879) (-)

Terphenyl-d14
Concen: 88.46 ng
RT: 12.998 min Scan# 1855
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050







Tgt Ion:	244	Resp:	3092681
Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.7	8.5
122	12.4	9.7	14.5




Abundance Scan 1936 (13.474 min): BF123565.D\data.ms (-1880) (-)

Butylbenzylphthalate
Concen: 47.31 ng
RT: 13.474 min Scan# 1936
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	149	Resp:	1041459
Ion	Ratio	Lower	Upper
149	100		
91	80.5	63.6	95.4
206	17.9	14.5	21.7









Benzo(a)anthracene
Concen: 46.33 ng
RT: 14.045 min Scan# 2033
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27




Instrument :
BNA_F
ClientSampleId :
SSTDIC050


Tgt Ion:	228	Resp:	1759085
Ion	Ratio	Lower	Upper
228	100		
226	27.2	22.4	33.6
229	20.1	16.0	24.0



3,3'-Dichlorobenzidine
Concen: 45.24 ng
RT: 14.010 min Scan# 2027
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	252	Resp:	547894
Ion	Ratio	Lower	Upper
252	100		
254	67.0	51.1	76.7
126	10.9	7.7	11.5









#33 (-)
Chrysene
 Concen: 47.03 ng
 RT: 14.086 min Scan# 2040
 Delta R.T. -0.000 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27

Instrument :
 BNA_F
ClientSampleId :
 SSTDICC050

Tgt Ion:228 Resp: 1787959




Ion	Ratio	Lower	Upper
228	100		
226	30.7	24.3	36.5
229	19.8	16.2	24.2




#34 (-)
Bis(2-ethylhexyl)phthalate
 Concen: 45.84 ng
 RT: 14.039 min Scan# 2032
 Delta R.T. -0.000 min
 Lab File: BF123566.D
 Acq: 15 Apr 2021 13:27

Tgt Ion:149 Resp: 1331127

Ion	Ratio	Lower	Upper
149	100		
167	27.0	21.8	32.6
279	3.8	3.3	4.9





Abundance Scan 2137 (14.657 min): BF123565.D\data.ms (-2137) (-)




Di-n-octyl phthalate
Concen: 50.09 ng
RT: 14.657 min Scan# 2137
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050


Tgt Ion:149 Resp: 2116383
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 14.0 11.8 17.6



Abundance Scan 2137 (14.657 min): BF123566.D\data.ms (-2086) (-)





Abundance Scan 2287 (15.539 min): BF123565.D\data.ms (-2287) (-)




Perylene-d12
Concen: 20.00 ng
RT: 15.539 min Scan# 2287
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:264 Resp: 457990
Ion Ratio Lower Upper
264 100
260 23.4 18.7 28.1
265 20.5 17.8 26.8




Abundance Scan 2287 (15.539 min): BF123566.D\data.ms (-2236) (-)






Abundance Scan 2542 (17.039 min): BF123565.D\data.ms (-2537) (-)

Indeno(1,2,3-cd)pyrene
Concen: 37.81 ng
RT: 17.039 min Scan# 2542
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050







Tgt Ion:	276	Resp:	1222226
Ion	Ratio	Lower	Upper
276	100		
138	21.1	16.2	24.2
277	25.5	20.1	30.1




Abundance Scan 2214 (15.110 min): BF123565.D\data.ms (-2488) (-)

Benzo(b)fluoranthene
Concen: 46.54 ng
RT: 15.109 min Scan# 2214
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	252	Resp:	1438478
Ion	Ratio	Lower	Upper
252	100		
253	22.3	18.1	27.1
125	8.5	6.2	9.4





Abundance Scan 2219 (15.139 min): BF123565.D\data.ms (-2168) (-)




Benzo(k)fluoranthene
Concen: 47.21 ng
RT: 15.139 min Scan# 2219
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050


Tgt Ion:252 Resp: 1442183
Ion Ratio Lower Upper
252 100
253 22.0 17.9 26.9
125 9.1 7.0 10.6



Abundance Scan 2219 (15.139 min): BF123566.D\data.ms (-2168) (-)





Abundance Scan 2277 (15.480 min): BF123565.D\data.ms (-2168) (-)




Benzo(a)pyrene
Concen: 46.20 ng
RT: 15.480 min Scan# 2277
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27


Tgt Ion:252 Resp: 1256457
Ion Ratio Lower Upper
252 100
253 21.5 18.4 27.6
125 9.2 8.0 12.0



Abundance Scan 2277 (15.480 min): BF123566.D\data.ms (-2226) (-)






Abundance Scan 2545 (17.057 min): BF123565.D\data.ms (-2545) (-)




Dibenzo(a,h)anthracene
Concen: 37.83 ng
RT: 17.056 min Scan# 2545
Delta R.T. -0.000 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tgt Ion:	278	Resp:	1047328
Ion	Ratio	Lower	Upper
278	100		
139	12.2	10.2	15.2
279	23.4	19.4	29.0



Abundance Scan 2618 (17.486 min): BF123565.D\data.ms (-2618) (-)



Benzo(g,h,i)perylene
Concen: 40.77 ng
RT: 17.492 min Scan# 2619
Delta R.T. 0.006 min
Lab File: BF123566.D
Acq: 15 Apr 2021 13:27

Tgt Ion:	276	Resp:	1047344
Ion	Ratio	Lower	Upper
276	100		
277	25.8	18.2	27.2
138	18.2	13.8	20.6

