

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041521\
 Data File : BF123567.D
 Acq On : 15 Apr 2021 13:59
 Operator : JU/CG
 Sample : SSTDICC060
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Quant Time: Apr 15 14:24:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 15 13:53:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	256952	20.00	ng	0.00
21) Naphthalene-d8	8.157	136	944197	20.00	ng	0.00
39) Acenaphthene-d10	9.922	164	492724	20.00	ng	0.00
64) Phenanthrene-d10	11.410	188	931337	20.00	ng	0.00
76) Chrysene-d12	14.063	240	613988	20.00	ng	0.00
86) Perylene-d12	15.539	264	455606	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	1712569	113.81	ng	0.00
7) Phenol-d6	6.504	99	2399235	116.97	ng	0.00
23) Nitrobenzene-d5	7.445	82	2330561	123.27	ng	0.01
42) 2,4,6-Tribromophenol	10.716	330	672780	139.85	ng	0.00
45) 2-Fluorobiphenyl	9.239	172	3550258	117.18	ng	0.00
79) Terphenyl-d14	13.004	244	3866519	109.48	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.616	88	453616	37.01	ng	99
3) Pyridine	3.375	79	1123835	31.84	ng	99
4) n-Nitrosodimethylamine	3.340	42	674354	49.47	ng	99
6) Aniline	6.534	93	1435908	56.10	ng	98
8) 2-Chlorophenol	6.663	128	934764	57.97	ng	98
9) Benzaldehyde	6.422	77	604499	46.46	ng	99
10) Phenol	6.522	94	1294325	59.36	ng	91
11) bis(2-Chloroethyl)ether	6.610	93	952667	54.97	ng	97
12) 1,3-Dichlorobenzene	6.816	146	944143	53.95	ng	97
13) 1,4-Dichlorobenzene	6.892	146	963365	54.49	ng	97
14) 1,2-Dichlorobenzene	7.045	146	891584	53.95	ng	97
15) Benzyl Alcohol	7.016	79	994917	62.99	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.151	45	2146233	65.31	ng	99
17) 2-Methylphenol	7.128	107	807512	59.02	ng	97
18) Hexachloroethane	7.386	117	405096	57.72	ng	94
19) n-Nitroso-di-n-propyla...	7.298	70	804097	58.92	ng	98
20) 3+4-Methylphenols	7.281	107	1015754	60.65	ng	97
22) Acetophenone	7.286	105	1427202	58.38	ng	# 95
24) Nitrobenzene	7.463	77	1215480	59.42	ng	97
25) Isophorone	7.698	82	2266398	59.52	ng	99
26) 2-Nitrophenol	7.775	139	482569	56.09	ng	99
27) 2,4-Dimethylphenol	7.810	122	766331	57.20	ng	99
28) bis(2-Chloroethoxy)met...	7.910	93	1121214	54.99	ng	100
29) 2,4-Dichlorophenol	8.016	162	761233	53.52	ng	99
30) 1,2,4-Trichlorobenzene	8.098	180	793711	49.10	ng	98
31) Naphthalene	8.181	128	2632900	56.23	ng	99
32) Benzoic acid	7.957	122	568520	64.11	ng	89
33) 4-Chloroaniline	8.228	127	1103015	56.67	ng	98
34) Hexachlorobutadiene	8.298	225	506567	49.16	ng	98
35) Caprolactam	8.622	113	215887	47.19	ng	94
36) 4-Chloro-3-methylphenol	8.716	107	976729	62.76	ng	95
37) 2-Methylnaphthalene	8.875	142	1756792	54.30	ng	99
38) 1-Methylnaphthalene	8.975	142	1637020	53.82	ng	98
40) 1,2,4,5-Tetrachloroben...	9.039	216	750061	53.15	ng	100
41) Hexachlorocyclopentadiene	9.028	237	453556	64.04	ng	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041521\
 Data File : BF123567.D
 Acq On : 15 Apr 2021 13:59
 Operator : JU/CG
 Sample : SSTDICC060
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Quant Time: Apr 15 14:24:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 15 13:53:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	9.151	196	546823	56.18	ng	97
44) 2,4,5-Trichlorophenol	9.192	196	586834	55.95	ng	99
46) 1,1'-Biphenyl	9.345	154	2085137	61.89	ng	99
47) 2-Chloronaphthalene	9.369	162	1592045	58.41	ng	99
48) 2-Nitroaniline	9.457	65	727385	75.76	ng	91
49) Acenaphthylene	9.786	152	2589758	62.88	ng	100
50) Dimethylphthalate	9.645	163	1892685	57.36	ng	100
51) 2,6-Dinitrotoluene	9.704	165	435074	58.56	ng	97
52) Acenaphthene	9.957	154	1705114	64.29	ng	99
53) 3-Nitroaniline	9.875	138	498979	65.70	ng	94
54) 2,4-Dinitrophenol	9.975	184	228651	76.82	ng	90
55) Dibenzofuran	10.127	168	2361476	61.03	ng	98
56) 4-Nitrophenol	10.033	139	402675	71.13	ng	95
57) 2,4-Dinitrotoluene	10.110	165	591048	61.51	ng	# 88
58) Fluorene	10.469	166	1852790	62.55	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.245	232	474928	58.56	ng	99
60) Diethylphthalate	10.345	149	2074892	63.59	ng	99
61) 4-Chlorophenyl-phenyle...	10.463	204	856172	55.18	ng	98
62) 4-Nitroaniline	10.498	138	496500	69.79	ng	96
63) Azobenzene	10.622	77	2346042	66.19	ng	96
65) 4,6-Dinitro-2-methylph...	10.522	198	312296	65.56	ng	94
66) n-Nitrosodiphenylamine	10.586	169	1646697	61.60	ng	100
67) 4-Bromophenyl-phenylether	10.951	248	554292	58.06	ng	93
68) Hexachlorobenzene	11.022	284	625689	60.19	ng	98
69) Atrazine	11.110	200	521124	57.47	ng	97
70) Pentachlorophenol	11.210	266	374036	68.61	ng	96
71) Phenanthrene	11.439	178	2591310	56.70	ng	100
72) Anthracene	11.486	178	2588440	56.67	ng	99
73) Carbazole	11.639	167	2583776	59.32	ng	97
74) Di-n-butylphthalate	11.974	149	3244381	55.83	ng	99
75) Fluoranthene	12.627	202	2783361	57.02	ng	98
77) Benzidine	12.745	184	901021	54.51	ng	99
78) Pyrene	12.857	202	2786038	57.00	ng	98
80) Butylbenzylphthalate	13.474	149	1321862	59.44	ng	100
81) Benzo(a)anthracene	14.051	228	2098469	54.72	ng	99
82) 3,3'-Dichlorobenzidine	14.010	252	632176	51.67	ng	# 97
83) Chrysene	14.086	228	2067594	53.84	ng	99
84) Bis(2-ethylhexyl)phtha...	14.039	149	1719343	58.61	ng	99
85) Di-n-octyl phthalate	14.657	149	2656497	62.24	ng	99
87) Indeno(1,2,3-cd)pyrene	17.045	276	1603224	49.86	ng	100
88) Benzo(b)fluoranthene	15.110	252	1843807	59.96	ng	98
89) Benzo(k)fluoranthene	15.145	252	1599058	52.62	ng	98
90) Benzo(a)pyrene	15.486	252	1517549	56.09	ng	98
91) Dibenzo(a,h)anthracene	17.068	278	1334299	48.45	ng	99
92) Benzo(g,h,i)perylene	17.503	276	1344327	52.60	ng	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 6.875 min Scan# 814
 Delta R.T. 0.006 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Tgt Ion:152 Resp: 256952
 Ion Ratio Lower Upper
 152 100
 150 163.6 127.0 190.6
 115 61.2 54.3 81.5




1,4-Dioxane
 Concen: 37.01 ng
 RT: 2.616 min Scan# 90
 Delta R.T. -0.000 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59

Tgt Ion: 88 Resp: 453616
 Ion Ratio Lower Upper
 88 100
 58 104.0 83.8 125.8
 43 39.6 32.2 48.4






Abundance Scan 217 (3.363 min): BF123565.D\data.ms (-213)#3




Pyridine
Concen: 31.84 ng
RT: 3.375 min Scan# 219
Delta R.T. 0.012 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060

Tgt Ion: 79 Resp: 1123835
Ion Ratio Lower Upper
79 100
52 108.1 87.4 131.0
51 51.3 41.4 62.0







Abundance Scan 210 (3.322 min): BF123565.D\data.ms (-203)#4



n-Nitrosodimethylamine
Concen: 49.47 ng
RT: 3.340 min Scan# 213
Delta R.T. 0.018 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion: 42 Resp: 674354
Ion Ratio Lower Upper
42 100
74 103.8 83.5 125.3
44 7.0 5.7 8.5









2-Fluorophenol
 Concen: 113.81 ng
 RT: 5.487 min Scan# 578
 Delta R.T. 0.006 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Tgt Ion:112 Resp: 1712569




Ion	Ratio	Lower	Upper
112	100		
64	75.5	61.4	92.0
63	40.5	33.0	49.4




Aniline
 Concen: 56.10 ng
 RT: 6.534 min Scan# 756
 Delta R.T. -0.000 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59

Tgt Ion: 93 Resp: 1435908

Ion	Ratio	Lower	Upper
93	100		
66	46.7	36.4	54.6
65	24.6	19.4	29.0









Phenol-d6
 Concen: 116.97 ng
 RT: 6.504 min Scan# 751
 Delta R.T. 0.006 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Tgt Ion: 99 Resp: 2399235




Ion	Ratio	Lower	Upper
99	100		
42	24.9	19.8	29.8
71	40.5	31.1	46.7




2-Chlorophenol
 Concen: 57.97 ng
 RT: 6.663 min Scan# 778
 Delta R.T. 0.006 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59

Tgt Ion: 128 Resp: 934764

Ion	Ratio	Lower	Upper
128	100		
130	33.4	12.6	52.6
64	59.3	41.5	81.5









Benzaldehyde
Concen: 46.46 ng
RT: 6.422 min Scan# 737
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59




Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:	77	Resp:	604499
Ion	Ratio	Lower	Upper
77	100		
105	96.3	75.2	115.2
106	92.9	71.6	111.6




Phenol
Concen: 59.36 ng
RT: 6.522 min Scan# 754
Delta R.T. 0.012 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:	94	Resp:	1294325
Ion	Ratio	Lower	Upper
94	100		
65	40.9	18.1	58.1
66	57.4	29.0	69.0






Abundance Scan 768 (6.604 min): BF123565.D\data.ms (-764) #1




bis(2-Chloroethyl)ether
Concen: 54.97 ng
RT: 6.610 min Scan# 769
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060

Tgt Ion:	93	Resp:	952667
Ion	Ratio	Lower	Upper
93	100		
63	99.2	76.1	116.1
95	33.3	11.5	51.5






Abundance Scan 803 (6.810 min): BF123565.D\data.ms (-798) #1



1,3-Dichlorobenzene
Concen: 53.95 ng
RT: 6.816 min Scan# 804
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59


Tgt Ion:	146	Resp:	944143
Ion	Ratio	Lower	Upper
146	100		
148	65.0	50.9	76.3
75	40.0	34.2	51.4





Abundance Scan 816 (6.887 min): BF123565.D\data.ms (-811) #13

1,4-Dichlorobenzene
Concen: 54.49 ng
RT: 6.892 min Scan# 817
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59


Instrument :
BNA_F
ClientSampled :
SSTDIC060



Tgt Ion:	146	Resp:	963365
Ion	Ratio	Lower	Upper
146	100		
148	64.6	49.9	74.9
111	47.5	39.5	59.3





Abundance Scan 817 (6.892 min): BF123567.D\data.ms (-765) (-)




Abundance Scan 843 (7.045 min): BF123565.D\data.ms (-837) #14

1,2-Dichlorobenzene
Concen: 53.95 ng
RT: 7.045 min Scan# 843
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59


Tgt Ion:	146	Resp:	891584
Ion	Ratio	Lower	Upper
146	100		
148	63.3	51.6	77.4
111	51.9	38.4	57.6



Abundance Scan 843 (7.045 min): BF123567.D\data.ms (-792) (-)






Abundance Scan 837 (7.010 min): BF123565.D\data.ms (-831) #15




Benzyl Alcohol
Concen: 62.99 ng
RT: 7.016 min Scan# 838
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060

Tgt Ion:	79	Resp:	994917
Ion	Ratio	Lower	Upper
79	100		
108	70.7	55.8	83.8
77	63.9	49.9	74.9






Abundance Scan 860 (7.145 min): BF123565.D\data.ms (-852) #16




2,2'-oxybis(1-Chloropropane)
Concen: 65.31 ng
RT: 7.151 min Scan# 861
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:	45	Resp:	2146233
Ion	Ratio	Lower	Upper
45	100		
77	10.3	0.0	30.0
79	7.7	0.0	27.3





Abundance Scan 856 (7.122 min): BF123565.D\data.ms (-850) #17




2-Methylphenol
Concen: 59.02 ng
RT: 7.128 min Scan# 857
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060


Tgt Ion:	107	Resp:	807512
Ion	Ratio	Lower	Upper
107	100		
108	116.0	89.8	134.8
77	53.2	40.9	61.3
79	52.1	40.2	60.4



Abundance Scan 857 (7.128 min): BF123567.D\data.ms (-805) (-)





Abundance Scan 901 (7.387 min): BF123565.D\data.ms (-896) #18





Hexachloroethane
Concen: 57.72 ng
RT: 7.386 min Scan# 901
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:	117	Resp:	405096
Ion	Ratio	Lower	Upper
117	100		
119	93.7	77.8	116.8
201	76.0	67.8	101.8




Abundance Scan 901 (7.386 min): BF123567.D\data.ms (-850) (-)








#19
 n-Nitroso-di-n-propylamine
 Concen: 58.92 ng
 RT: 7.298 min Scan# 886
 Delta R.T. 0.012 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59


Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060





Tgt Ion:	70	Resp:	804097
Ion	Ratio	Lower	Upper
70	100		
42	73.6	60.1	90.1
101	9.4	7.4	11.2
130	19.0	14.2	21.2




#20
 3+4-Methylphenols
 Concen: 60.65 ng
 RT: 7.281 min Scan# 883
 Delta R.T. 0.006 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59



Tgt Ion:	107	Resp:	1015754
Ion	Ratio	Lower	Upper
107	100		
108	91.3	71.5	111.5
77	109.0	82.0	122.0
79	33.8	14.1	54.1






Abundance Scan 1032 (8.157 min): BF123565.D\data.ms (-1021)




Naphthalene-d8
Concen: 20.00 ng
RT: 8.157 min Scan# 1032
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060

Tgt Ion:	136	Resp:	944197
Ion	Ratio	Lower	Upper
136	100		
137	10.6	8.7	13.1
54	12.5	9.8	14.6
68	5.4	4.1	6.1






Abundance Scan 883 (7.281 min): BF123565.D\data.ms (-876) #2




Acetophenone
Concen: 58.38 ng
RT: 7.286 min Scan# 884
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:	105	Resp:	1427202
Ion	Ratio	Lower	Upper
105	100		
71	4.2	4.2	6.4#
51	46.4	33.6	50.4
120	19.6	16.3	24.5






Abundance Scan 909 (7.434 min): BF123565.D\data.ms (-904) #23




Nitrobenzene-d5
Concen: 123.27 ng
RT: 7.445 min Scan# 911
Delta R.T. 0.012 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060

Tgt Ion: 82 Resp: 2330561
Ion Ratio Lower Upper
82 100
128 38.9 29.4 44.2
54 65.2 54.1 81.1






Abundance Scan 913 (7.457 min): BF123565.D\data.ms (-905) #24




Nitrobenzene
Concen: 59.42 ng
RT: 7.463 min Scan# 914
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion: 77 Resp: 1215480
Ion Ratio Lower Upper
77 100
123 40.5 34.5 51.7
65 15.8 12.2 18.2






Abundance Scan 953 (7.692 min): BF123565.D\data.ms (-947) #25




Isophorone
Concen: 59.52 ng
RT: 7.698 min Scan# 954
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060

Tgt Ion:	82	Resp:	2266398
Ion	Ratio	Lower	Upper
82	100		
95	7.6	5.8	8.6
138	14.6	11.8	17.8






Abundance Scan 966 (7.769 min): BF123565.D\data.ms (-961) #26




2-Nitrophenol
Concen: 56.09 ng
RT: 7.775 min Scan# 967
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:	139	Resp:	482569
Ion	Ratio	Lower	Upper
139	100		
109	36.0	28.6	43.0
65	69.8	56.8	85.2






Abundance Scan 972 (7.804 min): BF123565.D\data.ms (-968) #27




2,4-Dimethylphenol
Concen: 57.20 ng
RT: 7.810 min Scan# 973
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:	122	Resp:	766331
Ion	Ratio	Lower	Upper
122	100		
107	132.8	106.3	159.5
121	58.0	45.4	68.0






Abundance Scan 989 (7.904 min): BF123565.D\data.ms (-980) #28




bis(2-Chloroethoxy)methane
Concen: 54.99 ng
RT: 7.910 min Scan# 990
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:	93	Resp:	1121214
Ion	Ratio	Lower	Upper
93	100		
95	33.0	26.5	39.7
123	11.1	8.4	12.6





Abundance Scan 1007 (8.010 min): BF123565.D\data.ms (-10629)




2,4-Dichlorophenol
Concen: 53.52 ng
RT: 8.016 min Scan# 1008
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060


Tgt Ion:	162	Resp:	761233
Ion	Ratio	Lower	Upper
162	100		
164	64.9	44.1	84.1
98	46.9	27.3	67.3



Abundance Scan 1008 (8.016 min): BF123567.D\data.ms (-956) (-)





Abundance Scan 1022 (8.098 min): BF123565.D\data.ms (-10630)




1,2,4-Trichlorobenzene
Concen: 49.10 ng
RT: 8.098 min Scan# 1022
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59


Tgt Ion:	180	Resp:	793711
Ion	Ratio	Lower	Upper
180	100		
182	97.8	77.2	115.8
145	34.1	26.0	39.0



Abundance Scan 1022 (8.098 min): BF123567.D\data.ms (-971) (-)






Abundance Scan 1036 (8.181 min): BF123565.D\data.ms (-10281-)




Naphthalene
Concen: 56.23 ng
RT: 8.181 min Scan# 1036
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:128 Resp: 2632900
Ion Ratio Lower Upper
128 100
129 11.6 8.9 13.3
127 13.5 10.8 16.2






Abundance Scan 992 (7.922 min): BF123565.D\data.ms (-977#32)




Benzoic acid
Concen: 64.11 ng
RT: 7.957 min Scan# 998
Delta R.T. 0.035 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:122 Resp: 568520
Ion Ratio Lower Upper
122 100
105 141.6 107.8 147.8
77 111.8 81.6 121.6






Abundance Scan 1043 (8.222 min): BF123565.D\data.ms (-1033)




4-Chloroaniline
Concen: 56.67 ng
RT: 8.228 min Scan# 1044
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampled :
SSTDIC060

Tgt Ion:	127	Resp:	1103015
Ion	Ratio	Lower	Upper
127	100		
129	32.3	25.7	38.5
65	41.0	34.5	51.7
92	22.7	17.4	26.0






Abundance Scan 1056 (8.298 min): BF123565.D\data.ms (-1034)




Hexachlorobutadiene
Concen: 49.16 ng
RT: 8.298 min Scan# 1056
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:	225	Resp:	506567
Ion	Ratio	Lower	Upper
225	100		
223	61.7	50.4	75.6
227	62.4	51.4	77.0





Abundance Scan 1107 (8.598 min): BF123565.D\data.ms (-11035)




Caprolactam
Concen: 47.19 ng
RT: 8.622 min Scan# 1111
Delta R.T. 0.023 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060


Tgt Ion:	113	Resp:	215887
Ion	Ratio	Lower	Upper
113	100		
55	284.4	254.9	294.9
56	209.0	179.4	219.4



Abundance Scan 1111 (8.622 min): BF123567.D\data.ms (-1056) (-)





Abundance Scan 1125 (8.704 min): BF123565.D\data.ms (-11136)




4-Chloro-3-methylphenol
Concen: 62.76 ng
RT: 8.716 min Scan# 1127
Delta R.T. 0.012 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:	107	Resp:	976729
Ion	Ratio	Lower	Upper
107	100		
144	21.2	16.3	24.5
142	67.1	50.1	75.1




Abundance Scan 1127 (8.716 min): BF123567.D\data.ms (-1074) (-)





Abundance Scan 1153 (8.869 min): BF123565.D\data.ms (-11437-)

2-Methylnaphthalene
Concen: 54.30 ng
RT: 8.875 min Scan# 1154
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59


Instrument :
BNA_F
ClientSampled :
SSTDIC060



Tgt Ion:142 Resp: 1756792
Ion Ratio Lower Upper
142 100
141 89.1 71.8 107.6






Abundance Scan 1154 (8.875 min): BF123567.D\data.ms (-1102) (-)




Abundance Scan 1171 (8.975 min): BF123565.D\data.ms (-11438-)

1-Methylnaphthalene
Concen: 53.82 ng
RT: 8.975 min Scan# 1171
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59


Tgt Ion:142 Resp: 1637020
Ion Ratio Lower Upper
142 100
141 93.0 72.8 109.2



Abundance Scan 1171 (8.975 min): BF123567.D\data.ms (-1120) (-)






Abundance Scan 1331 (9.916 min): BF123565.D\data.ms (-1329)




Acenaphthene-d10
Concen: 20.00 ng
RT: 9.922 min Scan# 1332
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060

Tgt Ion:	164	Resp:	492724
Ion	Ratio	Lower	Upper
164	100		
162	103.5	80.7	121.1
160	46.7	37.3	55.9






Abundance Scan 1182 (9.039 min): BF123565.D\data.ms (-1179)




1,2,4,5-Tetrachlorobenzene
Concen: 53.15 ng
RT: 9.039 min Scan# 1182
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:	216	Resp:	750061
Ion	Ratio	Lower	Upper
216	100		
214	79.5	63.8	95.8
179	22.6	17.9	26.9
108	23.0	18.6	28.0






Abundance Scan 1180 (9.028 min): BF123565.D\data.ms (-11741)




Hexachlorocyclopentadiene
Concen: 64.04 ng
RT: 9.028 min Scan# 1180
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060

Tgt Ion:	237	Resp:	453556
Ion	Ratio	Lower	Upper
237	100		
235	58.2	38.8	78.8
272	11.8	0.0	32.4






Abundance Scan 1466 (10.710 min): BF123565.D\data.ms (-1460)




2,4,6-Tribromophenol
Concen: 139.85 ng
RT: 10.716 min Scan# 1467
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:	330	Resp:	672780
Ion	Ratio	Lower	Upper
330	100		
332	95.6	77.1	115.7
141	33.9	27.9	41.9






Abundance Scan 1200 (9.145 min): BF123565.D\data.ms (-11953)




2,4,6-Trichlorophenol
Concen: 56.18 ng
RT: 9.151 min Scan# 1201
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampled :
SSTDIC060

Tgt Ion:196 Resp: 546823
Ion Ratio Lower Upper
196 100
198 94.2 78.2 117.2
200 31.6 25.0 37.6






Abundance Scan 1207 (9.186 min): BF123565.D\data.ms (-1204)




2,4,5-Trichlorophenol
Concen: 55.95 ng
RT: 9.192 min Scan# 1208
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:196 Resp: 586834
Ion Ratio Lower Upper
196 100
198 96.4 76.5 114.7
97 64.6 52.9 79.3
132 32.8 25.3 37.9






Abundance Scan 1216 (9.239 min): BF123565.D\data.ms (-121145)




2-Fluorobiphenyl
Concen: 117.18 ng
RT: 9.239 min Scan# 1216
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060

Tgt Ion:172 Resp: 3550258
Ion Ratio Lower Upper
172 100
171 35.7 28.1 42.1
170 24.9 19.2 28.8







Abundance Scan 1233 (9.339 min): BF123565.D\data.ms (-122746)



1,1'-Biphenyl
Concen: 61.89 ng
RT: 9.345 min Scan# 1234
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:154 Resp: 2085137
Ion Ratio Lower Upper
154 100
153 41.3 20.7 60.7
76 11.0 0.0 31.7









2-Chloronaphthalene
 Concen: 58.41 ng
 RT: 9.369 min Scan# 1238
 Delta R.T. 0.006 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Tgt Ion:162 Resp: 1592045




Ion	Ratio	Lower	Upper
162	100		
127	45.3	36.2	54.2
164	33.1	25.4	38.2




2-Nitroaniline
 Concen: 75.76 ng
 RT: 9.457 min Scan# 1253
 Delta R.T. -0.000 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59

Tgt Ion: 65 Resp: 727385

Ion	Ratio	Lower	Upper
65	100		
92	56.8	47.1	70.7
138	75.5	70.2	105.2






Abundance Scan 1308 (9.781 min): BF123565.D\data.ms (-123565) (-)




Acenaphthylene
Concen: 62.88 ng
RT: 9.786 min Scan# 1309
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060

Tgt Ion:152 Resp: 2589758
Ion Ratio Lower Upper
152 100
151 20.0 16.0 24.0
153 12.8 10.4 15.6






Abundance Scan 1284 (9.639 min): BF123565.D\data.ms (-123565) (-)




Dimethylphthalate
Concen: 57.36 ng
RT: 9.645 min Scan# 1285
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:163 Resp: 1892685
Ion Ratio Lower Upper
163 100
194 3.7 3.1 4.7
164 10.2 8.2 12.2






Abundance Scan 1294 (9.698 min): BF123565.D\data.ms (-12851-)




2,6-Dinitrotoluene
Concen: 58.56 ng
RT: 9.704 min Scan# 1295
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 435074
Ion Ratio Lower Upper
165 100
63 86.7 67.7 101.5
89 77.4 58.8 88.2






Abundance Scan 1337 (9.951 min): BF123565.D\data.ms (-13252-)




Acenaphthene
Concen: 64.29 ng
RT: 9.957 min Scan# 1338
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:154 Resp: 1705114
Ion Ratio Lower Upper
154 100
153 110.8 88.1 132.1
152 53.4 43.4 65.0






Abundance Scan 1323 (9.869 min): BF123565.D\data.ms (-1333)




3-Nitroaniline
Concen: 65.70 ng
RT: 9.875 min Scan# 1324
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampled :
SSTDIC060

Tgt Ion:138 Resp: 498979
Ion Ratio Lower Upper
138 100
108 11.7 10.2 15.4
92 136.1 103.3 154.9






Abundance Scan 1340 (9.969 min): BF123565.D\data.ms (-1335)



2,4-Dinitrophenol
Concen: 76.82 ng
RT: 9.975 min Scan# 1341
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59


Tgt Ion:184 Resp: 228651
Ion Ratio Lower Upper
184 100
63 104.2 73.5 110.3
154 70.8 52.5 78.7






Abundance Scan 1366 (10.122 min): BF123565.D\data.ms (-1315) (-)

Dibenzofuran
Concen: 61.03 ng
RT: 10.127 min Scan# 1367
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060




Tgt Ion:168 Resp: 2361476
Ion Ratio Lower Upper
168 100
139 43.1 33.2 49.8
169 13.6 10.8 16.2







Abundance Scan 1349 (10.022 min): BF123565.D\data.ms (-1356) (-)

4-Nitrophenol
Concen: 71.13 ng
RT: 10.033 min Scan# 1351
Delta R.T. 0.012 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59



Tgt Ion:139 Resp: 402675
Ion Ratio Lower Upper
139 100
109 84.3 63.2 103.2
65 122.0 110.0 150.0









2,4-Dinitrotoluene
 Concen: 61.51 ng
 RT: 10.110 min Scan# 1364
 Delta R.T. 0.006 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Tgt Ion:	165	Resp:	591048
Ion	Ratio	Lower	Upper
165	100		
63	65.7	43.5	65.3#
89	94.0	67.7	101.5
182	2.4	2.1	3.1



Fluorene
 Concen: 62.55 ng
 RT: 10.469 min Scan# 1425
 Delta R.T. -0.000 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59


Tgt Ion:	166	Resp:	1852790
Ion	Ratio	Lower	Upper
166	100		
165	90.7	73.5	110.3
167	13.5	10.7	16.1






Abundance Scan 1386 (10.239 min): BF123565.D\data.ms (-1358) (-)

2,3,4,6-Tetrachlorophenol
Concen: 58.56 ng
RT: 10.245 min Scan# 1387
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060




Tgt Ion:	232	Resp:	474928
Ion	Ratio	Lower	Upper
232	100		
131	54.6	43.2	64.8
130	2.7	2.2	3.2
166	34.0	27.3	40.9







Abundance Scan 1403 (10.339 min): BF123565.D\data.ms (-1359) (-)

Diethylphthalate
Concen: 63.59 ng
RT: 10.345 min Scan# 1404
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59



Tgt Ion:	149	Resp:	2074892
Ion	Ratio	Lower	Upper
149	100		
177	20.2	15.4	23.2
150	12.7	9.9	14.9









4-Chlorophenyl-phenylether
Concen: 55.18 ng
RT: 10.463 min Scan# 1424
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59




Instrument :
BNA_F
ClientSampleId :
SSTDIC060


Tgt Ion:	204	Resp:	856172
Ion	Ratio	Lower	Upper
204	100		
206	34.2	25.7	38.5
141	68.2	55.5	83.3



4-Nitroaniline
Concen: 69.79 ng
RT: 10.498 min Scan# 1430
Delta R.T. 0.012 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:	138	Resp:	496500
Ion	Ratio	Lower	Upper
138	100		
92	57.1	35.8	75.8
108	88.2	63.6	103.6









Azobenzene
Concen: 66.19 ng
RT: 10.622 min Scan# 1451
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59




Instrument :
BNA_F
ClientSampleId :
SSTDIC060


Tgt Ion: 77 Resp: 2346042
Ion Ratio Lower Upper
77 100
182 18.5 1.7 41.7
105 21.0 1.7 41.7
51 38.1 15.9 55.9



Phenanthrene-d10
Concen: 20.00 ng
RT: 11.410 min Scan# 1585
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion: 188 Resp: 931337
Ion Ratio Lower Upper
188 100
94 8.1 6.9 10.3
80 9.3 8.2 12.2









4,6-Dinitro-2-methylphenol
 Concen: 65.56 ng
 RT: 10.522 min Scan# 1434
 Delta R.T. 0.012 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060


Tgt Ion:	198	Resp:	312296
Ion	Ratio	Lower	Upper
198	100		
51	78.6	64.8	104.8
105	59.4	42.6	82.6



n-Nitrosodiphenylamine
 Concen: 61.60 ng
 RT: 10.586 min Scan# 1445
 Delta R.T. 0.006 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59

Tgt Ion:	169	Resp:	1646697
Ion	Ratio	Lower	Upper
169	100		
168	67.2	53.3	79.9
167	37.3	29.9	44.9









Scan 1507 (10.951 min): BF123565.D\data.ms (-1567) (-)
 4-Bromophenyl-phenylether
 Concen: 58.06 ng
 RT: 10.951 min Scan# 1507
 Delta R.T. -0.000 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Tgt Ion:	248	Resp:	554292
Ion	Ratio	Lower	Upper
248	100		
250	97.8	77.4	116.2
141	92.3	63.7	95.5



Scan 1518 (11.016 min): BF123565.D\data.ms (-1568) (-)
 Hexachlorobenzene
 Concen: 60.19 ng
 RT: 11.022 min Scan# 1519
 Delta R.T. 0.006 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59


Tgt Ion:	284	Resp:	625689
Ion	Ratio	Lower	Upper
284	100		
142	32.9	28.3	42.5
249	29.3	23.5	35.3






Abundance Scan 1533 (11.104 min): BF123565.D\data.ms (-1559) (-)

Atrazine
Concen: 57.47 ng
RT: 11.110 min Scan# 1534
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampled :
SSTDIC060




Tgt Ion:	200	Resp:	521124
Ion	Ratio	Lower	Upper
200	100		
173	28.2	6.9	46.9
215	48.2	26.5	66.5






Abundance Scan 1551 (11.210 min): BF123565.D\data.ms (-1570) (-)


Pentachlorophenol
Concen: 68.61 ng
RT: 11.210 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59



Tgt Ion:	266	Resp:	374036
Ion	Ratio	Lower	Upper
266	100		
268	62.2	51.6	77.4
264	62.2	52.6	78.8






Abundance Scan 1589 (11.433 min): BF123565.D\data.ms (-1572) (-)




Phenanthrene
Concen: 56.70 ng
RT: 11.439 min Scan# 1590
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060

Tgt Ion:178 Resp: 2591310
Ion Ratio Lower Upper
178 100
176 19.6 15.6 23.4
179 15.8 12.5 18.7







Abundance Scan 1598 (11.486 min): BF123565.D\data.ms (-1573) (-)



Anthracene
Concen: 56.67 ng
RT: 11.486 min Scan# 1598
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:178 Resp: 2588440
Ion Ratio Lower Upper
178 100
176 19.5 15.4 23.0
179 16.1 12.6 18.8









Scan 1624 (11.639 min): BF123565.D\data.ms (-1573) (-)
 Carbazole
 Concen: 59.32 ng
 RT: 11.639 min Scan# 1624
 Delta R.T. -0.000 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Tgt Ion:	167	Resp:	2583776
Ion	Ratio	Lower	Upper
167	100		
166	22.3	17.0	25.6
139	14.7	10.4	15.6



Scan 1680 (11.969 min): BF123565.D\data.ms (-1675) (-)
 Di-n-butylphthalate
 Concen: 55.83 ng
 RT: 11.974 min Scan# 1681
 Delta R.T. 0.006 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59


Tgt Ion:	149	Resp:	3244381
Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.8	11.8
104	5.6	4.6	6.8






Abundance Scan 1791 (12.621 min): BF123565.D\data.ms (-1791) (-)

Fluoranthene
Concen: 57.02 ng
RT: 12.627 min Scan# 1792
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060







Tgt Ion:	202	Resp:	2783361
Ion	Ratio	Lower	Upper
202	100		
101	10.7	0.0	30.9
203	19.0	0.0	37.5




Abundance Scan 2035 (14.057 min): BF123565.D\data.ms (-2035) (-)

Chrysene-d12
Concen: 20.00 ng
RT: 14.063 min Scan# 2036
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:	240	Resp:	613988
Ion	Ratio	Lower	Upper
240	100		
120	7.7	7.2	10.8
236	25.1	19.0	28.4









Benzidine
Concen: 54.51 ng
RT: 12.745 min Scan# 1812
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59




Instrument :
BNA_F
ClientSampleId :
SSTDIC060


Tgt Ion:	184	Resp:	901021
Ion	Ratio	Lower	Upper
184	100		
185	14.7	12.1	18.1
183	10.3	7.8	11.8



Pyrene
Concen: 57.00 ng
RT: 12.857 min Scan# 1831
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:	202	Resp:	2786038
Ion	Ratio	Lower	Upper
202	100		
200	21.2	16.1	24.1
203	18.2	14.8	22.2









Terphenyl-d14
 Concen: 109.48 ng
 RT: 13.004 min Scan# 1856
 Delta R.T. 0.006 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Tgt Ion:244 Resp: 3866519




Ion	Ratio	Lower	Upper
244	100		
212	7.2	5.7	8.5
122	12.1	9.7	14.5




Butylbenzylphthalate
 Concen: 59.44 ng
 RT: 13.474 min Scan# 1936
 Delta R.T. -0.000 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59

Tgt Ion:149 Resp: 1321862

Ion	Ratio	Lower	Upper
149	100		
91	79.8	63.6	95.4
206	18.5	14.5	21.7






Abundance Scan 2033 (14.045 min): BF123565.D\data.ms (-2033) (-)




Benzo(a)anthracene
Concen: 54.72 ng
RT: 14.051 min Scan# 2034
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:228 Resp: 2098469
Ion Ratio Lower Upper
228 100
226 27.7 22.4 33.6
229 19.6 16.0 24.0







Abundance Scan 2027 (14.010 min): BF123565.D\data.ms (-2027) (-)



3,3'-Dichlorobenzidine
Concen: 51.67 ng
RT: 14.010 min Scan# 2027
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:252 Resp: 632176
Ion Ratio Lower Upper
252 100
254 65.8 51.1 76.7
126 11.9 7.7 11.5#









#33 (-)
Chrysene
 Concen: 53.84 ng
 RT: 14.086 min Scan# 2040
 Delta R.T. -0.000 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59

Instrument :
 BNA_F
ClientSampleId :
 SSTDICC060

Tgt Ion:228 Resp: 2067594




Ion	Ratio	Lower	Upper
228	100		
226	31.5	24.3	36.5
229	20.0	16.2	24.2



#34 (-)
Bis(2-ethylhexyl)phthalate
 Concen: 58.61 ng
 RT: 14.039 min Scan# 2032
 Delta R.T. -0.000 min
 Lab File: BF123567.D
 Acq: 15 Apr 2021 13:59

Tgt Ion:149 Resp: 1719343


Ion	Ratio	Lower	Upper
149	100		
167	26.6	21.8	32.6
279	3.9	3.3	4.9



Abundance Scan 2137 (14.657 min): BF123565.D\data.ms (-2135) (-)




Di-n-octyl phthalate
Concen: 62.24 ng
RT: 14.657 min Scan# 2137
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060



Tgt Ion:149 Resp: 2656497

Ion	Ratio	Lower	Upper
149	100		
167	1.4	1.2	1.8
43	14.3	11.8	17.6







Abundance Scan 2287 (15.539 min): BF123565.D\data.ms (-2286) (-)


Perylene-d12
Concen: 20.00 ng
RT: 15.539 min Scan# 2287
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:264 Resp: 455606

Ion	Ratio	Lower	Upper
264	100		
260	25.5	18.7	28.1
265	21.4	17.8	26.8






Abundance Scan 2542 (17.039 min): BF123565.D\data.ms (-2537) (-)




Indeno(1,2,3-cd)pyrene
Concen: 49.86 ng
RT: 17.045 min Scan# 2543
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:	276	Resp:	1603224
Ion	Ratio	Lower	Upper
276	100		
138	20.0	16.2	24.2
277	24.9	20.1	30.1






Abundance Scan 2214 (15.110 min): BF123565.D\data.ms (-2188) (-)




Benzo(b)fluoranthene
Concen: 59.96 ng
RT: 15.110 min Scan# 2214
Delta R.T. -0.000 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:	252	Resp:	1843807
Ion	Ratio	Lower	Upper
252	100		
253	22.2	18.1	27.1
125	9.1	6.2	9.4





Abundance Scan 2219 (15.139 min): BF123565.D\data.ms (-2168) (-)




Benzo(k)fluoranthene
Concen: 52.62 ng
RT: 15.145 min Scan# 2220
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060


Tgt Ion:252 Resp: 1599058
Ion Ratio Lower Upper
252 100
253 21.2 17.9 26.9
125 8.5 7.0 10.6



Abundance Scan 2220 (15.145 min): BF123567.D\data.ms (-2168) (-)





Abundance Scan 2277 (15.480 min): BF123565.D\data.ms (-2168) (-)




Benzo(a)pyrene
Concen: 56.09 ng
RT: 15.486 min Scan# 2278
Delta R.T. 0.006 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59


Tgt Ion:252 Resp: 1517549
Ion Ratio Lower Upper
252 100
253 22.0 18.4 27.6
125 9.0 8.0 12.0



Abundance Scan 2278 (15.486 min): BF123567.D\data.ms (-2226) (-)






Abundance Scan 2545 (17.057 min): BF123565.D\data.ms (-2545) (-)




Dibenzo(a,h)anthracene
Concen: 48.45 ng
RT: 17.068 min Scan# 2547
Delta R.T. 0.012 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Instrument :
BNA_F
ClientSampleId :
SSTDIC060

Tgt Ion:278 Resp: 1334299
Ion Ratio Lower Upper
278 100
139 12.6 10.2 15.2
279 23.6 19.4 29.0



Abundance Scan 2618 (17.486 min): BF123565.D\data.ms (-2618) (-)



Benzo(g,h,i)perylene
Concen: 52.60 ng
RT: 17.503 min Scan# 2621
Delta R.T. 0.018 min
Lab File: BF123567.D
Acq: 15 Apr 2021 13:59

Tgt Ion:276 Resp: 1344327
Ion Ratio Lower Upper
276 100
277 23.2 18.2 27.2
138 16.5 13.8 20.6

