



Benzoic acid



$\text{Response} = 2.123\text{e-}001 * \text{Amt} - 4.825\text{e-}002$
 Coef of Det (r^2) = 0.992041 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF041521.M
 Calibration Table Last Updated: Fri Apr 16 10:31:33 2021

Benzo (k) fluoranthene

Response Ratio



$$\text{Response} = 1.220\text{e}+000 * \text{Amt} + 1.005\text{e}-001$$


Coef of Det (r^2) = 0.993130 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF041521.M

Calibration Table Last Updated: Fri Apr 16 10:31:33 2021

Benzo(g,h,i)perylene

Response Ratio



$R = 1.099e-001 A^2 + 6.039e-001 A + 3.854e-002$
 Coef of Det (r^2) = 0.992418 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF041521.M
 Calibration Table Last Updated: Fri Apr 16 10:31:33 2021