

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF041615\
 Data File : BF078563.D
 Acq On : 16 Apr 2015 14:04
 Operator : TP/IZ
 Sample : G1889-01
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-014-A

Quant Time: Apr 17 05:50:44 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 17 00:52:22 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.68	152	27007	20.00	ng	0.00
21) Naphthalene-d8	8.26	136	111589	20.00	ng	-0.01
38) Acenaphthene-d10	10.42	164	58606	20.00	ng	0.00
63) Phenanthrene-d10	12.24	188	115671	20.00	ng	0.00
75) Chrysene-d12	15.51	240	124369	20.00	ng	0.02
86) Perylene-d12	17.28	264	126841	20.00	ng	0.09
System Monitoring Compounds						
5) 2-Fluorophenol	4.91	112	169552	103.51	ng	0.01
7) Phenol-d6	6.30	99	224641	109.43	ng	0.00
23) Nitrobenzene-d5	7.39	82	131609	71.49	ng	0.00
41) 2,4,6-Tribromophenol	11.39	330	56378	115.99	ng	0.00
44) 2-Fluorobiphenyl	9.62	172	223332	54.47	ng	0.00
78) Terphenyl-d14	14.23	244	266223	48.37	ng	0.00
Target Compounds						
49) Dimethylphthalate	10.14	163	21916	4.98	ng	Qvalue # 95
70) Phenanthrene	12.26	178	19249	2.88	ng	99
74) Fluoranthene	13.73	202	40635	5.76	ng	94
77) Pyrene	14.00	202	36180	4.63	ng	# 94
80) Benzo(a)anthracene	15.50	228	19658	2.66	ng	# 92
82) Chrysene	15.53	228	20038m	2.88	ng	
87) Benzo(b)fluoranthene	16.82	252	24112m	3.08	ng	
89) Benzo(a)pyrene	17.21	252	16955	2.48	ng	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF041615\
Data File : BF078563.D
Acq On : 16 Apr 2015 14:04
Operator : TP/IZ
Sample : G1889-01
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FK-GF-01-014-A

Quant Time: Apr 17 05:50:44 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 17 00:52:22 2015
Response via : Initial Calibration

