

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF041615\
 Data File : BF078568.D
 Acq On : 16 Apr 2015 16:27
 Operator : TP/IZ
 Sample : G1891-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-82D

Quant Time: Apr 17 01:49:04 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 17 00:52:22 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.68	152	27254	20.00	ng	0.00
21) Naphthalene-d8	8.27	136	118524	20.00	ng	0.00
38) Acenaphthene-d10	10.42	164	64405	20.00	ng	0.00
63) Phenanthrene-d10	12.24	188	126028	20.00	ng	0.00
75) Chrysene-d12	15.50	240	142204	20.00	ng	0.01
86) Perylene-d12	17.20	264	149497	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	4.91	112	135423	81.93	ng	0.01
7) Phenol-d6	6.30	99	177181	85.53	ng	0.00
23) Nitrobenzene-d5	7.39	82	106065	54.24	ng	0.00
41) 2,4,6-Tribromophenol	11.39	330	45247	84.71	ng	0.00
44) 2-Fluorobiphenyl	9.62	172	195786	43.45	ng	0.00
78) Terphenyl-d14	14.23	244	213209	33.88	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	10.14	163	19587	4.05	ng	97
70) Phenanthrene	12.27	178	146996	20.16	ng	98
71) Anthracene	12.33	178	35709	4.97	ng	99
74) Fluoranthene	13.73	202	272511	35.45	ng	97
77) Pyrene	14.00	202	266535	29.86	ng	# 93
80) Benzo(a)anthracene	15.49	228	144929	17.13	ng	100
82) Chrysene	15.52	228	131127	16.50	ng	97
85) Indeno(1,2,3-cd)pyrene	18.39	276	76688m	8.50	ng	
87) Benzo(b)fluoranthene	16.77	252	147721m	15.99	ng	
88) Benzo(k)fluoranthene	16.79	252	54289m	6.61	ng	
89) Benzo(a)pyrene	17.13	252	120817	14.98	ng	# 94
90) Dibenzo(a,h)anthracene	18.41	278	18988	2.35	ng	# 88
91) Benzo(g,h,i)perylene	18.72	276	68965	8.52	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF041615\
Data File : BF078568.D
Acq On : 16 Apr 2015 16:27
Operator : TP/IZ
Sample : G1891-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-82D

Quant Time: Apr 17 01:49:04 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 17 00:52:22 2015
Response via : Initial Calibration

