

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041715\
 Data File : BF078636.D
 Acq On : 18 Apr 2015 10:56
 Operator : TP/IZ
 Sample : G1854-12
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 LS-SB07

Manual Integrations
 APPROVED

apatel
 4/20/2015 3:57:46 PM

Quant Time: Apr 20 01:43:17 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 18 05:04:14 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.67	152	37490	20.00	ng	0.00
21) Naphthalene-d8	8.25	136	160369	20.00	ng	0.00
38) Acenaphthene-d10	10.41	164	84884	20.00	ng	0.00
63) Phenanthrene-d10	12.23	188	172287	20.00	ng	0.01
75) Chrysene-d12	15.47	240	204324	20.00	ng	0.01
86) Perylene-d12	17.17	264	212211	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	4.90	112	211145	92.86	ng	0.00
7) Phenol-d6	6.28	99	280155	98.31	ng	0.00
23) Nitrobenzene-d5	7.38	82	180744	68.31	ng	0.00
41) 2,4,6-Tribromophenol	11.38	330	76701	108.95	ng	0.00
44) 2-Fluorobiphenyl	9.60	172	276512	46.56	ng	0.00
78) Terphenyl-d14	14.21	244	313448	34.66	ng	0.00
Target Compounds						
10) Phenol	6.31	94	7615	2.23	ng	Qvalue 89
48) Acenaphthylene	10.23	152	20079	2.28	ng	98
49) Dimethylphthalate	10.11	163	43150	6.77	ng	99
70) Phenanthrene	12.25	178	75292	7.55	ng	98
71) Anthracene	12.31	178	30707	3.13	ng	98
74) Fluoranthene	13.70	202	170355	16.21	ng	98
77) Pyrene	13.98	202	146010	11.38	ng	97
80) Benzo(a)anthracene	15.46	228	113470	9.33	ng	96
82) Chrysene	15.50	228	97984	8.58	ng	97
85) Indeno(1,2,3-cd)pyrene	18.34	276	84635	6.53	ng	# 91
87) Benzo(b)fluoranthene	16.73	252	163417m	12.46	ng	
88) Benzo(k)fluoranthene	16.74	252	65836m	5.65	ng	
89) Benzo(a)pyrene	17.10	252	107058	9.35	ng	# 94
91) Benzo(a,h,i)perylene	18.67	276	73080	6.36	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041715\
Data File : BF078636.D
Acq On : 18 Apr 2015 10:56
Operator : TP/IZ
Sample : G1854-12
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
LS-SB07

Manual Integrations
APPROVED

apatel
4/20/2015 3:57:46 PM

Quant Time: Apr 20 01:43:17 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Apr 18 05:04:14 2015
Response via : Initial Calibration

