


Data Path : Z:\HPCHEM1\BNA F\DATA\BF041715\
 Data File : BF078601.D
 Acq On : 17 Apr 2015 12:11
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Mar 31 14:58:34 2015



AutoFind: Scans 1001, 1002, 1003; Background Corrected with Scan 998

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	38.7	5427	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	37.6	5276	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.6	6954	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	14026	PASS
199	198	5	9	6.7	937	PASS
275	198	10	60	24.1	3383	PASS
365	198	1	100	2.7	372	PASS
441	442	0.01	24	15.4	1762	PASS
442	198	50	100	81.6	11442	PASS
443	442	15	24	18.3	2099	PASS