

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041717\
 Data File : BF094474.D
 Acq On : 18 Apr 2017 2:38
 Operator : SJ/MA
 Sample : I2708-16 10X
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-X10-ESW

Manual Integrations
 APPROVED

mohammad
 4/18/2017 4:56:01 PM

Quant Time: Apr 18 05:55:52 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 17 14:59:23 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.76	152	155712	20.00	ng	0.00
21) Naphthalene-d8	7.26	136	615410	20.00	ng	0.00
38) Acenaphthene-d10	9.40	164	244716	20.00	ng	0.00
63) Phenanthrene-d10	11.22	188	425287	20.00	ng	0.00
75) Chrysene-d12	14.46	240	305602	20.00	ng	0.00
86) Perylene-d12	16.09	264	265658	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	4.29	112	165950	17.68	ng	-0.01
7) Phenol-d6	5.37	99	213990	19.34	ng	-0.02
23) Nitrobenzene-d5	6.40	82	118216	11.86	ng	-0.02
41) 2,4,6-Tribromophenol	10.37	330	31755	16.59	ng	0.00
44) 2-Fluorobiphenyl	8.58	172	216880	13.04	ng	0.00
78) Terphenyl-d14	13.19	244	159509	13.95	ng	0.00
Target Compounds						
37) 2-Methylnaphthalene	8.13	142	160703	7.94	ng	99
49) Dimethylphthalate	9.08	163	72624	3.98	ng	# 99
70) Phenanthrene	11.25	178	90535	3.78	ng	99
74) Fluoranthene	12.70	202	67823	2.73	ng	99
77) Pyrene	12.97	202	65140	3.05	ng	97
87) Benzo(b)fluoranthene	15.68	252	45472m	2.80	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041717\
Data File : BF094474.D
Acq On : 18 Apr 2017 2:38
Operator : SJ/MA
Sample : I2708-16 10X
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-X10-ESW

Manual Integrations
APPROVED

mohammad
4/18/2017 4:56:01 PM

Quant Time: Apr 18 05:55:52 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 17 14:59:23 2017
Response via : Initial Calibration

