


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF041717\
 Data File : BF094475.D
 Acq On : 18 Apr 2017 4:28
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF041717.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Apr 17 14:59:23 2017



AutoFind: Scans 1652, 1653, 1654; Background Corrected with Scan 1646

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.6	26133	PASS
68	69	0.00	2	1.7	434	PASS
69	198	0.00	100	35.7	25515	PASS
70	69	0.00	2	0.4	102	PASS
127	198	10	80	48.8	34856	PASS
197	198	0.00	2	0.3	218	PASS
198	198	100	100	100.0	71384	PASS
199	198	5	9	7.0	5012	PASS
275	198	10	60	23.5	16784	PASS
365	198	1	100	2.3	1646	PASS
441	442	0.01	24	15.9	9660	PASS
442	198	50	100	85.1	60778	PASS
443	442	15	24	19.1	11604	PASS