

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041718\
 Data File : BF104568.D
 Acq On : 17 Apr 2018 18:24
 Operator : JU/SJ
 Sample : J2354-04
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-2-4PT-COMP

Manual Integrations
 APPROVED

Sohil
 4/18/2018 3:00:44 PM

Quant Time: Apr 18 01:10:53 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 17 17:08:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	126010	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	474160	20.00	ng	0.00
38) Acenaphthene-d10	9.84	164	187173	20.00	ng	0.00
63) Phenanthrene-d10	11.33	188	317329	20.00	ng	0.00
75) Chrysene-d12	13.97	240	283581	20.00	ng	0.00
86) Perylene-d12	15.44	264	220880	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	809861	98.27	ng	0.00
7) Phenol-d6	6.45	99	966756	95.48	ng	0.00
23) Nitrobenzene-d5	7.37	82	595830	82.05	ng	0.00
41) 2,4,6-Tribromophenol	10.63	330	198218	102.09	ng	0.00
44) 2-Fluorobiphenyl	9.16	172	906995	76.55	ng	0.00
78) Terphenyl-d14	12.92	244	784650	63.08	ng	0.00
Target Compounds						
10) Phenol	6.45	94	33252	3.095	ng	81
49) Dimethylphthalate	9.56	163	129462	8.771	ng	99
83) Bis(2-ethylhexyl)phthalate	13.95	149	55673	4.371	ng	100
87) Benzo(b)fluoranthene	15.01	252	43926m	3.149	ng	
91) Benzo(g,h,i)perylene	17.32	276	22315	2.247	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041718\
Data File : BF104568.D
Acq On : 17 Apr 2018 18:24
Operator : JU/SJ
Sample : J2354-04
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-2-4PT-COMP

Manual Integrations
APPROVED

Sohil
4/18/2018 3:00:44 PM

Quant Time: Apr 18 01:10:53 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 17 17:08:38 2018
Response via : Initial Calibration

