

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041723\
 Data File : BF132884.D
 Acq On : 17 Apr 2023 12:30
 Operator : CG\JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Quant Time: Apr 18 03:58:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF041723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 18 03:56:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.763	152	298025	20.000	ng	0.00
21) Naphthalene-d8	8.051	136	1216069	20.000	ng	0.00
39) Acenaphthene-d10	9.804	164	640514	20.000	ng	0.00
64) Phenanthrene-d10	11.286	188	1160739	20.000	ng	0.00
76) Chrysene-d12	13.922	240	812815	20.000	ng	0.00
86) Perylene-d12	15.351	264	759665	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	107339	5.632	ng	0.00
7) Phenol-d6	6.375	99	140904	5.794	ng	-0.02
23) Nitrobenzene-d5	7.322	82	118475	5.078	ng	-0.01
42) 2,4,6-Tribromophenol	10.586	330	28735	4.473	ng	0.00
45) 2-Fluorobiphenyl	9.122	172	274537	6.585	ng	-0.01
79) Terphenyl-d14	12.869	244	269446	5.656	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.458	88	26591	2.859	ng	96
3) Pyridine	3.187	79	64923	2.567	ng	97
4) n-Nitrosodimethylamine	3.099	42	31526	2.609	ng	# 93
6) Aniline	6.422	93	85630	2.655	ng	98
8) 2-Chlorophenol	6.540	128	54925	2.841	ng	99
9) Benzaldehyde	6.304	77	50276	3.494	ng	93
10) Phenol	6.393	94	71195	2.686	ng	95
11) bis(2-Chloroethyl)ether	6.493	93	59185	2.866	ng	98
12) 1,3-Dichlorobenzene	6.704	146	62252	2.929	ng	99
13) 1,4-Dichlorobenzene	6.781	146	63805	2.996	ng	100
14) 1,2-Dichlorobenzene	6.934	146	61378	3.087	ng	96
15) Benzyl Alcohol	6.898	79	42495	2.550	ng	93
16) 2,2'-oxybis(1-Chloropr...	7.040	45	104551	2.926	ng	96
17) 2-Methylphenol	7.010	107	45914	2.593	ng	96
18) Hexachloroethane	7.281	117	20346	2.753	ng	96
19) n-Nitroso-di-n-propyla...	7.169	70	39519	2.603	ng	96
20) 3+4-Methylphenols	7.163	107	61833	2.954	ng	92
22) Acetophenone	7.169	105	87126	3.054	ng	# 93
24) Nitrobenzene	7.340	77	62596	2.618	ng	99
25) Isophorone	7.581	82	112731	2.547	ng	98
27) 2,4-Dimethylphenol	7.698	122	47745	2.595	ng	97
28) bis(2-Chloroethoxy)met...	7.798	93	73159	2.765	ng	97
29) 2,4-Dichlorophenol	7.898	162	42049	2.455	ng	96
30) 1,2,4-Trichlorobenzene	7.992	180	51565	2.788	ng	98
31) Naphthalene	8.069	128	184333	3.000	ng	96
33) 4-Chloroaniline	8.116	127	65415	2.599	ng	97
34) Hexachlorobutadiene	8.192	225	29207	2.761	ng	99
36) 4-Chloro-3-methylphenol	8.587	107	43088	2.397	ng	98
37) 2-Methylnaphthalene	8.763	142	127348	3.011	ng	98
38) 1-Methylnaphthalene	8.857	142	115930	2.942	ng	96
40) 1,2,4,5-Tetrachloroben...	8.928	216	51364	2.835	ng	99
43) 2,4,6-Trichlorophenol	9.034	196	26700	2.325	ng	99
44) 2,4,5-Trichlorophenol	9.069	196	32922	2.452	ng	94
46) 1,1'-Biphenyl	9.222	154	150671	2.967	ng	98
47) 2-Chloronaphthalene	9.245	162	107295	2.855	ng	99
49) Acenaphthylene	9.663	152	172787	2.883	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041723\
 Data File : BF132884.D
 Acq On : 17 Apr 2023 12:30
 Operator : CG\JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5


Quant Time: Apr 18 03:58:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF041723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 18 03:56:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
50) Dimethylphthalate	9.522	163	121544	2.745	ng	98
51) 2,6-Dinitrotoluene	9.581	165	19538	2.012	ng	89
52) Acenaphthene	9.834	154	115241	2.984	ng	96
53) 3-Nitroaniline	9.745	138	22468	2.034	ng	95
55) Dibenzofuran	10.004	168	163651	2.987	ng	96
58) Fluorene	10.345	166	125909	3.137	ng	97
59) 2,3,4,6-Tetrachlorophenol	10.122	232	24213	2.323	ng	93
60) Diethylphthalate	10.222	149	106368	2.571	ng	98
61) 4-Chlorophenyl-phenyle...	10.345	204	61948	3.034	ng	98
63) Azobenzene	10.498	77	131723	2.850	ng	96
66) n-Nitrosodiphenylamine	10.457	169	104003	2.751	ng	100
67) 4-Bromophenyl-phenylether	10.833	248	35058	2.645	ng	98
68) Hexachlorobenzene	10.892	284	37017	2.692	ng	# 86
69) Atrazine	10.981	200	22587	2.138	ng	95
71) Phenanthrene	11.304	178	188501	3.019	ng	99
72) Anthracene	11.357	178	185445	2.901	ng	98
73) Carbazole	11.510	167	154709	2.798	ng	98
74) Di-n-butylphthalate	11.851	149	120035	2.025	ng	98
75) Fluoranthene	12.492	202	168154	2.708	ng	97
78) Pyrene	12.722	202	175269	2.637	ng	97
81) Benzo(a)anthracene	13.910	228	139416	2.522	ng	98
83) Chrysene	13.945	228	146185	2.661	ng	97
87) Indeno(1,2,3-cd)pyrene	16.739	276	89019	2.051	ng	98
88) Benzo(b)fluoranthene	14.939	252	105379	2.227	ng	97
89) Benzo(k)fluoranthene	14.963	252	114999	2.440	ng	99
90) Benzo(a)pyrene	15.286	252	91207	2.110	ng	99
91) Dibenzo(a,h)anthracene	16.757	278	77714	2.101	ng	98
92) Benzo(g,h,i)perylene	17.157	276	74677	2.109	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 18 03:58:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF041723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 18 03:56:41 2023
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.763 min Scan# 79
 Delta R.T. -0.006 min
 Lab File: BF132884.D
 Acq: 17 Apr 2023 12:30




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5


Tgt Ion:152 Resp: 298025
 Ion Ratio Lower Upper
 152 100
 150 156.9 122.8 184.2
 115 62.9 47.3 70.9



#2
 1,4-Dioxane
 Concen: 2.859 ng
 RT: 2.458 min Scan# 63
 Delta R.T. 0.006 min
 Lab File: BF132884.D
 Acq: 17 Apr 2023 12:30

Tgt Ion: 88 Resp: 26591
 Ion Ratio Lower Upper
 88 100
 58 80.0 67.8 101.6
 43 31.7 25.4 38.0









#3
Pyridine
Concen: 2.567 ng
RT: 3.187 min Scan# 11
Delta R.T. 0.030 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion: 79 Resp: 64923
Ion Ratio Lower Upper
79 100
52 83.9 65.4 98.0
51 38.8 32.6 48.8



#4
n-Nitrosodimethylamine
Concen: 2.609 ng
RT: 3.099 min Scan# 172
Delta R.T. -0.017 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion: 42 Resp: 31526
Ion Ratio Lower Upper
42 100
74 116.0 98.4 147.6
44 15.4 6.1 9.1#









#5
2-Fluorophenol
Concen: 5.632 ng
RT: 5.357 min Scan# 511
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion: 112 Resp: 107339
Ion Ratio Lower Upper
112 100
64 59.9 51.8 77.8
63 29.0 26.3 39.5



#6
Aniline
Concen: 2.655 ng
RT: 6.422 min Scan# 737
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion: 93 Resp: 85630
Ion Ratio Lower Upper
93 100
66 38.6 31.8 47.6
65 19.1 16.3 24.5





#7

Phenol-d6

Concen: 5.794 ng

RT: 6.375 min Scan# 71

Delta R.T. -0.023 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5





Tgt Ion: 99 Resp: 140904

Ion Ratio Lower Upper

99 100

42 20.5 16.1 24.1

71 32.3 26.1 39.1



#8

2-Chlorophenol

Concen: 2.841 ng

RT: 6.540 min Scan# 757

Delta R.T. -0.012 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30




Tgt Ion: 128 Resp: 54925


Ion Ratio Lower Upper

128 100

130 33.5 13.2 53.2

64 51.3 30.9 70.9









#9
Benzaldehyde
Concen: 3.494 ng
RT: 6.304 min Scan# 719
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion: 77 Resp: 50276
Ion Ratio Lower Upper
77 100
105 93.1 80.8 120.8
106 90.8 77.9 117.9



#10
Phenol
Concen: 2.686 ng
RT: 6.393 min Scan# 732
Delta R.T. -0.017 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion: 94 Resp: 71195
Ion Ratio Lower Upper
94 100
65 26.8 9.0 49.0
66 34.0 17.1 57.1









#11
bis(2-Chloroethyl)ether
Concen: 2.866 ng
RT: 6.493 min Scan# 749
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion: 93 Resp: 59185
Ion Ratio Lower Upper
93 100
63 84.0 66.7 106.7
95 32.1 12.8 52.8



#12
1,3-Dichlorobenzene
Concen: 2.929 ng
RT: 6.704 min Scan# 785
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:146 Resp: 62252
Ion Ratio Lower Upper
146 100
148 64.0 51.1 76.7
75 31.3 23.9 35.9





#13

1,4-Dichlorobenzene

Concen: 2.996 ng

RT: 6.781 min Scan# 799

Delta R.T. -0.006 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5





Tgt Ion:146 Resp: 63805

Ion Ratio Lower Upper

146 100

148 63.1 50.7 76.1

111 40.4 32.5 48.7



#14

1,2-Dichlorobenzene

Concen: 3.087 ng

RT: 6.934 min Scan# 824

Delta R.T. -0.006 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30




Tgt Ion:146 Resp: 61378


Ion Ratio Lower Upper

146 100

148 66.6 51.3 76.9

111 42.1 36.0 54.0









#15
Benzyl Alcohol
Concen: 2.550 ng
RT: 6.898 min Scan# 818
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion: 79 Resp: 42495
Ion Ratio Lower Upper
79 100
108 82.6 63.0 94.6
77 71.2 50.4 75.6



#16
2,2'-oxybis(1-Chloropropane)
Concen: 2.926 ng
RT: 7.040 min Scan# 842
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion: 45 Resp: 104551
Ion Ratio Lower Upper
45 100
77 10.6 0.0 31.6
79 7.0 0.0 29.2









#17
2-Methylphenol
Concen: 2.593 ng
RT: 7.010 min Scan# 811
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:	107	Resp:	45914
Ion	Ratio	Lower	Upper
107	100		
108	120.1	91.7	137.5
77	43.4	35.8	53.8
79	41.9	35.0	52.4



#18
Hexachloroethane
Concen: 2.753 ng
RT: 7.281 min Scan# 883
Delta R.T. 0.000 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:	117	Resp:	20346
Ion	Ratio	Lower	Upper
117	100		
119	93.5	76.1	114.1
201	96.6	71.8	107.6





#19

n-Nitroso-di-n-propylamine

Concen: 2.603 ng

RT: 7.169 min Scan# 864

Delta R.T. -0.023 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5

Tgt Ion: 70 Resp: 39519





Ion Ratio Lower Upper

70 100

42 63.4 49.0 73.6

101 8.4 8.1 12.1

130 19.9 18.2 27.2



#20

3+4-Methylphenols

Concen: 2.954 ng

RT: 7.163 min Scan# 863

Delta R.T. -0.012 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30

Tgt Ion: 107 Resp: 61833




Ion Ratio Lower Upper


107 100

108 96.4 68.5 108.5

77 105.9 77.1 117.1

79 27.7 7.4 47.4









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.051 min Scan# 1014
Delta R.T. 0.000 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:136 Resp: 1216069
Ion Ratio Lower Upper
136 100
137 11.0 9.0 13.4
54 9.4 8.9 13.3
68 7.0 6.5 9.7



#22
Acetophenone
Concen: 3.054 ng
RT: 7.169 min Scan# 864
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:105 Resp: 87126
Ion Ratio Lower Upper
105 100
71 7.0 3.1 4.7#
51 32.7 31.4 47.2
120 22.6 18.6 27.8





#23

Nitrobenzene-d5

Concen: 5.078 ng

RT: 7.322 min Scan# 890

Delta R.T. -0.012 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5





Tgt Ion: 82 Resp: 118475

Ion Ratio Lower Upper

82 100

128 43.9 35.7 53.5

54 54.4 44.7 67.1



#24

Nitrobenzene

Concen: 2.618 ng

RT: 7.340 min Scan# 893

Delta R.T. -0.012 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30




Tgt Ion: 77 Resp: 62596


Ion Ratio Lower Upper

77 100

123 45.6 36.3 54.5

65 12.2 11.2 16.8









#25
Isophorone
Concen: 2.547 ng
RT: 7.581 min Scan# 91
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion: 82 Resp: 112731
Ion Ratio Lower Upper
82 100
95 6.7 5.3 7.9
138 18.8 14.2 21.2



#27
2,4-Dimethylphenol
Concen: 2.595 ng
RT: 7.698 min Scan# 954
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion: 122 Resp: 47745
Ion Ratio Lower Upper
122 100
107 110.5 88.5 132.7
121 53.8 47.2 70.8









#28
bis(2-Chloroethoxy)methane
Concen: 2.765 ng
RT: 7.798 min Scan# 91
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion: 93 Resp: 73159
Ion Ratio Lower Upper
93 100
95 31.2 26.3 39.5
123 12.1 9.3 13.9



#29
2,4-Dichlorophenol
Concen: 2.455 ng
RT: 7.898 min Scan# 988
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion: 162 Resp: 42049
Ion Ratio Lower Upper
162 100
164 68.2 45.2 85.2
98 37.5 14.8 54.8









#30
1,2,4-Trichlorobenzene
Concen: 2.788 ng
RT: 7.992 min Scan# 1004
Delta R.T. 0.000 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:180 Resp: 51565
Ion Ratio Lower Upper
180 100
182 95.4 76.6 114.8
145 26.9 25.4 38.0



#31
Naphthalene
Concen: 3.000 ng
RT: 8.069 min Scan# 1017
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:128 Resp: 184333
Ion Ratio Lower Upper
128 100
129 10.5 9.4 14.0
127 12.2 11.0 16.6









#33
4-Chloroaniline
Concen: 2.599 ng
RT: 8.116 min Scan# 1025
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:	127	Resp:	65415
Ion	Ratio	Lower	Upper
127	100		
129	32.0	25.9	38.9
65	29.9	26.7	40.1
92	19.1	16.5	24.7



#34
Hexachlorobutadiene
Concen: 2.761 ng
RT: 8.192 min Scan# 1038
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:	225	Resp:	29207
Ion	Ratio	Lower	Upper
225	100		
223	61.3	49.6	74.4
227	65.0	51.6	77.4









#36
 4-Chloro-3-methylphenol
 Concen: 2.397 ng
 RT: 8.587 min Scan# 1105
 Delta R.T. -0.012 min
 Lab File: BF132884.D
 Acq: 17 Apr 2023 12:30




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5


Tgt Ion:107 Resp: 43088
 Ion Ratio Lower Upper
 107 100
 144 25.3 19.4 29.0
 142 78.9 62.2 93.2



#37
 2-Methylnaphthalene
 Concen: 3.011 ng
 RT: 8.763 min Scan# 1135
 Delta R.T. 0.000 min
 Lab File: BF132884.D
 Acq: 17 Apr 2023 12:30

Tgt Ion:142 Resp: 127348
 Ion Ratio Lower Upper
 142 100
 141 86.3 70.4 105.6





#38

1-Methylnaphthalene

Concen: 2.942 ng

RT: 8.857 min Scan# 1151

Delta R.T. -0.006 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30

Instrument :

BNA_F

ClientSampleId :





SSTDICC2.5

Tgt Ion:142 Resp: 115930

Ion Ratio Lower Upper

142 100

141 95.0 72.6 109.0



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.804 min Scan# 1312

Delta R.T. 0.000 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30




Tgt Ion:164 Resp: 640514


Ion Ratio Lower Upper

164 100

162 100.1 79.1 118.7

160 44.6 35.6 53.4









#40
1,2,4,5-Tetrachlorobenzene
Concen: 2.835 ng
RT: 8.928 min Scan# 1163
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:	216	Resp:	51364
Ion	Ratio	Lower	Upper
216	100		
214	78.6	63.1	94.7
179	20.4	17.4	26.2
108	20.6	16.4	24.6



#42
2,4,6-Tribromophenol
Concen: 4.473 ng
RT: 10.586 min Scan# 1445
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:	330	Resp:	28735
Ion	Ratio	Lower	Upper
330	100		
332	99.2	76.3	114.5
141	35.3	23.9	35.9









#43
2,4,6-Trichlorophenol
Concen: 2.325 ng
RT: 9.034 min Scan# 1182
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:196 Resp: 26700
Ion Ratio Lower Upper
196 100
198 98.5 79.8 119.6
200 32.0 25.4 38.2



#44
2,4,5-Trichlorophenol
Concen: 2.452 ng
RT: 9.069 min Scan# 1187
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:196 Resp: 32922
Ion Ratio Lower Upper
196 100
198 98.2 80.2 120.2
97 47.8 45.0 67.6
132 28.8 25.6 38.4









#45
2-Fluorobiphenyl
Concen: 6.585 ng
RT: 9.122 min Scan# 1198
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:172 Resp: 274537
Ion Ratio Lower Upper
172 100
171 35.2 29.4 44.0
170 23.6 20.1 30.1



#46
1,1'-Biphenyl
Concen: 2.967 ng
RT: 9.222 min Scan# 1213
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:154 Resp: 150671
Ion Ratio Lower Upper
154 100
153 41.2 21.8 61.8
76 15.4 0.0 34.0









#47
2-Chloronaphthalene
Concen: 2.855 ng
RT: 9.245 min Scan# 1217
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:162 Resp: 107295
Ion Ratio Lower Upper
162 100
127 40.4 32.2 48.2
164 32.2 26.5 39.7



#49
Acenaphthylene
Concen: 2.883 ng
RT: 9.663 min Scan# 1288
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:152 Resp: 172787
Ion Ratio Lower Upper
152 100
151 19.4 16.4 24.6
153 12.8 11.4 17.0









#50
Dimethylphthalate
Concen: 2.745 ng
RT: 9.522 min Scan# 1264
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:163 Resp: 121544
Ion Ratio Lower Upper
163 100
194 4.4 3.4 5.0
164 9.5 8.3 12.5



#51
2,6-Dinitrotoluene
Concen: 2.012 ng
RT: 9.581 min Scan# 1274
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:165 Resp: 19538
Ion Ratio Lower Upper
165 100
63 61.5 42.6 63.8
89 59.1 41.0 61.4









#52
Acenaphthene
Concen: 2.984 ng
RT: 9.834 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:154 Resp: 115241
Ion Ratio Lower Upper
154 100
153 106.8 88.8 133.2
152 48.8 41.2 61.8



#53
3-Nitroaniline
Concen: 2.034 ng
RT: 9.745 min Scan# 1302
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:138 Resp: 22468
Ion Ratio Lower Upper
138 100
108 9.8 8.2 12.2
92 111.0 93.8 140.8









#55
Dibenzenofuran
Concen: 2.987 ng
RT: 10.004 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:168 Resp: 163651
Ion Ratio Lower Upper
168 100
139 37.0 31.4 47.0
169 12.6 11.2 16.8



#58
Fluorene
Concen: 3.137 ng
RT: 10.345 min Scan# 1404
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:166 Resp: 125909
Ion Ratio Lower Upper
166 100
165 100.0 78.0 117.0
167 12.4 11.0 16.4









#59
2,3,4,6-Tetrachlorophenol
Concen: 2.323 ng
RT: 10.122 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:232 Resp: 24213
Ion Ratio Lower Upper
232 100
131 50.1 35.9 53.9
130 2.2 1.8 2.8
166 32.3 23.5 35.3



#60
Diethylphthalate
Concen: 2.571 ng
RT: 10.222 min Scan# 1383
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:149 Resp: 106368
Ion Ratio Lower Upper
149 100
177 24.1 18.5 27.7
150 12.6 10.3 15.5









#61
4-Chlorophenyl-phenylether
Concen: 3.034 ng
RT: 10.345 min Scan# 1404
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:204 Resp: 61948
Ion Ratio Lower Upper
204 100
206 32.3 26.6 40.0
141 55.7 43.1 64.7



#63
Azobenzene
Concen: 2.850 ng
RT: 10.498 min Scan# 1430
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion: 77 Resp: 131723
Ion Ratio Lower Upper
77 100
182 20.8 6.0 46.0
105 21.1 1.6 41.6
51 32.2 11.6 51.6





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.286 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5





Tgt Ion:188 Resp: 1160739

Ion Ratio Lower Upper

188 100

94 11.1 9.9 14.9

80 11.4 10.8 16.2



#66

n-Nitrosodiphenylamine

Concen: 2.751 ng

RT: 10.457 min Scan# 1423

Delta R.T. -0.012 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30




Tgt Ion:169 Resp: 104003


Ion Ratio Lower Upper

169 100

168 66.5 53.4 80.0

167 36.5 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 2.645 ng
RT: 10.833 min Scan# 1488
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:248 Resp: 35058
Ion Ratio Lower Upper
248 100
250 98.8 79.1 118.7
141 66.9 51.0 76.6



#68
Hexachlorobenzene
Concen: 2.692 ng
RT: 10.892 min Scan# 1497
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:284 Resp: 37017
Ion Ratio Lower Upper
284 100
142 39.4 22.0 33.0#
249 30.8 22.6 34.0





#69

Atrazine

Concen: 2.138 ng

RT: 10.981 min Scan# 1512

Delta R.T. -0.012 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5





Tgt Ion:200 Resp: 22587

Ion Ratio Lower Upper

200 100

173 28.0 8.7 48.7

215 52.2 27.2 67.2



#71

Phenanthrene

Concen: 3.019 ng

RT: 11.304 min Scan# 1567

Delta R.T. -0.012 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30




Tgt Ion:178 Resp: 188501


Ion Ratio Lower Upper

178 100

176 19.0 15.9 23.9

179 15.3 12.3 18.5





#72

Anthracene

Concen: 2.901 ng

RT: 11.357 min Scan# 11

Delta R.T. -0.006 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5





Tgt Ion:178 Resp: 185445

Ion Ratio Lower Upper

178 100

176 18.4 15.5 23.3

179 15.4 12.8 19.2



#73

Carbazole

Concen: 2.798 ng

RT: 11.510 min Scan# 1602

Delta R.T. -0.006 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30




Tgt Ion:167 Resp: 154709


Ion Ratio Lower Upper

167 100

166 21.0 17.4 26.2

139 12.4 11.0 16.4





#74

Di-n-butylphthalate

Concen: 2.025 ng

RT: 11.851 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5





Tgt Ion:149 Resp: 120035

Ion Ratio Lower Upper

149 100

150 8.9 7.8 11.6

104 6.0 4.2 6.2



#75

Fluoranthene

Concen: 2.708 ng

RT: 12.492 min Scan# 1769

Delta R.T. -0.006 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30




Tgt Ion:202 Resp: 168154


Ion Ratio Lower Upper

202 100

101 14.9 0.0 34.5

203 16.4 0.0 38.2





#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.922 min Scan# 20

Delta R.T. -0.006 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5





Tgt Ion:240 Resp: 812815

Ion Ratio Lower Upper

240 100

120 13.0 10.0 15.0

236 26.0 20.8 31.2



#78

Pyrene

Concen: 2.637 ng

RT: 12.722 min Scan# 1808

Delta R.T. -0.006 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30




Tgt Ion:202 Resp: 175269


Ion Ratio Lower Upper

202 100

200 20.4 17.6 26.4

203 16.8 14.6 22.0





#79

Terphenyl-d14

Concen: 5.656 ng

RT: 12.869 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5





Tgt Ion:244 Resp: 269446

Ion Ratio Lower Upper

244 100

212 6.3 5.4 8.0

122 11.7 9.6 14.4



#81

Benzo(a)anthracene

Concen: 2.522 ng

RT: 13.910 min Scan# 2010

Delta R.T. -0.006 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30




Tgt Ion:228 Resp: 139416


Ion Ratio Lower Upper

228 100

226 26.4 22.4 33.6

229 20.2 16.2 24.2





#83

Chrysene

Concen: 2.661 ng

RT: 13.945 min Scan# 2016

Delta R.T. -0.006 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5





Tgt Ion:228 Resp: 146185

Ion Ratio Lower Upper

228 100

226 28.6 24.3 36.5

229 19.1 16.2 24.4



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.351 min Scan# 2255

Delta R.T. 0.000 min

Lab File: BF132884.D

Acq: 17 Apr 2023 12:30




Tgt Ion:264 Resp: 759665


Ion Ratio Lower Upper

264 100

260 22.5 19.0 28.6

265 20.9 17.7 26.5









#87
Indeno(1,2,3-cd)pyrene
Concen: 2.051 ng
RT: 16.739 min Scan# 2491
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:276 Resp: 89019
Ion Ratio Lower Upper
276 100
138 33.4 25.5 38.3
277 24.4 20.3 30.5



#88
Benzo(b)fluoranthene
Concen: 2.227 ng
RT: 14.939 min Scan# 2185
Delta R.T. -0.006 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:252 Resp: 105379
Ion Ratio Lower Upper
252 100
253 21.1 17.9 26.9
125 11.1 10.0 15.0









#89
Benzo(k)fluoranthene
Concen: 2.440 ng
RT: 14.963 min Scan# 2191
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion:252 Resp: 114999
Ion Ratio Lower Upper
252 100
253 22.6 18.0 27.0
125 13.4 10.0 15.0



#90
Benzo(a)pyrene
Concen: 2.110 ng
RT: 15.286 min Scan# 2244
Delta R.T. -0.012 min
Lab File: BF132884.D
Acq: 17 Apr 2023 12:30

Tgt Ion:252 Resp: 91207
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 13.0 9.9 14.9









#91
 Dibenzo(a,h)anthracene
 Concen: 2.101 ng
 RT: 16.757 min Scan# 2496
 Delta R.T. -0.012 min
 Lab File: BF132884.D
 Acq: 17 Apr 2023 12:30

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tgt Ion: 278 Resp: 77714
 Ion Ratio Lower Upper
 278 100
 139 20.7 16.7 25.1
 279 24.6 18.4 27.6



#92
 Benzo(g,h,i)perylene
 Concen: 2.109 ng
 RT: 17.157 min Scan# 2562
 Delta R.T. -0.017 min
 Lab File: BF132884.D
 Acq: 17 Apr 2023 12:30

Tgt Ion: 276 Resp: 74677
 Ion Ratio Lower Upper
 276 100
 277 20.5 18.9 28.3
 138 26.2 21.7 32.5

