

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041816\
 Data File : BF086278.D
 Acq On : 18 Apr 2016 13:26
 Operator : UM/SJ
 Sample : H2471-12
 Misc :
 ALS Vial : 73 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CDMSH10

Quant Time: Apr 19 01:15:24 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 15 15:56:31 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	356908	20.00	ng	-0.01
21) Naphthalene-d8	8.17	136	1416541	20.00	ng	-0.01
38) Acenaphthene-d10	9.93	164	670383	20.00	ng	-0.01
63) Phenanthrene-d10	11.41	188	1228798	20.00	ng	0.00
75) Chrysene-d12	14.04	240	888090	20.00	ng	-0.01
86) Perylene-d12	15.48	264	737398	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	2055576	94.37	ng	0.01
7) Phenol-d6	6.52	99	2832575	103.96	ng	0.00
23) Nitrobenzene-d5	7.45	82	1618922	65.92	ng	-0.01
41) 2,4,6-Tribromophenol	10.72	330	502999	93.44	ng	-0.01
44) 2-Fluorobiphenyl	9.25	172	2411134	66.37	ng	-0.01
78) Terphenyl-d14	13.00	244	2059785	58.32	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.64	163	410637	8.42	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF041816\
Data File : BF086278.D
Acq On : 18 Apr 2016 13:26
Operator : UM/SJ
Sample : H2471-12
Misc :
ALS Vial : 73 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CDMSH10

Quant Time: Apr 19 01:15:24 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 15 15:56:31 2016
Response via : Initial Calibration

