

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041822\
 Data File : BF127828.D
 Acq On : 18 Apr 2022 16:36
 Operator : CG\JU
 Sample : N2444-05 10X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-4D

Quant Time: Apr 19 05:15:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 13 16:40:15 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.951	152	144484	20.000	ng	-0.01
21) Naphthalene-d8	8.239	136	461007	20.000	ng	# 0.00
39) Acenaphthene-d10	10.004	164	217932	20.000	ng	0.00
64) Phenanthrene-d10	11.498	188	334650	20.000	ng	0.00
76) Chrysene-d12	14.139	240	293089	20.000	ng	0.00
86) Perylene-d12	15.662	264	293134	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.557	112	92851	10.442	ng	-0.02
7) Phenol-d6	6.563	99	115255	9.908	ng	-0.02
23) Nitrobenzene-d5	7.639	82	15486	1.498	ng	0.11
42) 2,4,6-Tribromophenol	10.792	330	19528	8.420	ng	0.00
45) 2-Fluorobiphenyl	9.310	172	118476	7.691	ng	-0.01
79) Terphenyl-d14	13.080	244	104117	5.864	ng	0.00
Target Compounds						
					Qvalue	
9) Benzaldehyde	6.481	77	5305	Below Cal	#	10
18) Hexachloroethane	7.481	117	87922	22.116	ng	# 1
19) n-Nitroso-di-n-propyla...	7.375	70	22682	3.015	ng	# 62
22) Acetophenone	7.269	105	46272	3.735	ng	# 68
25) Isophorone	7.757	82	63726	3.532	ng	# 1
35) Caprolactam	8.686	113	10126	4.509	ng	# 1
37) 2-Methylnaphthalene	9.063	142	1414785	87.566	ng	95
38) 1-Methylnaphthalene	9.063	142	1414785	90.722	ng	100
48) 2-Nitroaniline	9.522	65	15443	3.754	ng	# 50
49) Acenaphthylene	9.863	152	75459	3.746	ng	# 1
51) 2,6-Dinitrotoluene	9.798	165	6319	2.122	ng	# 1
52) Acenaphthene	10.039	154	130322	10.268	ng	# 56
53) 3-Nitroaniline	9.951	138	10994	3.495	ng	# 32
54) 2,4-Dinitrophenol	10.063	184	1279	6.117	ng	# 1
55) Dibenzofuran	10.210	168	132663	7.162	ng	# 1
56) 4-Nitrophenol	10.110	139	108961	44.100	ng	# 28
57) 2,4-Dinitrotoluene	10.210	165	51640	13.988	ng	# 56
58) Fluorene	10.557	166	275044	20.098	ng	# 93
63) Azobenzene	10.698	77	39165	2.319	ng	# 1
66) n-Nitrosodiphenylamine	10.663	169	337681	32.039	ng	# 47
71) Phenanthrene	11.527	178	1152496	64.999	ng	97
72) Anthracene	11.527	178	1152496	65.711	ng	95
75) Fluoranthene	12.716	202	75671	3.911	ng	75
78) Pyrene	12.945	202	348282	14.072	ng	# 92
81) Benzo(a)anthracene	14.127	228	84398	4.311	ng	# 40
83) Chrysene	14.162	228	160859	8.561	ng	# 60


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041822\
Data File : BF127828.D
Acq On : 18 Apr 2022 16:36
Operator : CG\JU
Sample : N2444-05 10X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-4D

Quant Time: Apr 19 05:15:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 13 16:40:15 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.951 min Scan# 795
Delta R.T. -0.012 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36




Instrument :
BNA_F
ClientSampleId :
SB-4D


Tgt Ion:152 Resp: 144484
Ion Ratio Lower Upper
152 100
150 147.0 117.2 175.8
115 61.2 49.8 74.6



#5
2-Fluorophenol
Concen: 10.442 ng
RT: 5.557 min Scan# 556
Delta R.T. -0.024 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36

Tgt Ion:112 Resp: 92851
Ion Ratio Lower Upper
112 100
64 61.6 58.4 87.6
63 33.1 30.1 45.1









#7
Phenol-d6
Concen: 9.908 ng
RT: 6.563 min Scan# 713
Delta R.T. -0.024 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36




Instrument :
BNA_F
ClientSampleId :
SB-4D


Tgt Ion: 99 Resp: 115255
Ion Ratio Lower Upper
99 100
42 19.5 17.8 26.6
71 37.7 29.6 44.4



#9
Benzaldehyde
Concen: Below Cal
RT: 6.481 min Scan# 713
Delta R.T. -0.035 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36

Tgt Ion: 77 Resp: 5305
Ion Ratio Lower Upper
77 100
105 5.7 66.8 106.8#
106 0.0 64.3 104.3#









#18
Hexachloroethane
Concen: 22.116 ng
RT: 7.481 min Scan# 883
Delta R.T. 0.006 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36




Instrument :
BNA_F
Client Sample Id :
SB-4D


Tgt Ion: 117 Resp: 87922
Ion Ratio Lower Upper
117 100
119 414.3 77.6 116.4#
201 0.0 88.2 132.2#



#19
n-Nitroso-di-n-propylamine
Concen: 3.015 ng
RT: 7.375 min Scan# 865
Delta R.T. -0.000 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36

Tgt Ion: 70 Resp: 22682
Ion Ratio Lower Upper
70 100
42 34.9 53.0 79.6#
101 0.0 6.8 10.2#
130 0.0 13.8 20.6#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.239 min Scan# 1014
Delta R.T. -0.006 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36




Instrument :
BNA_F
ClientSampleId :
SB-4D


Tgt Ion:136 Resp: 461007
Ion Ratio Lower Upper
136 100
137 11.4 8.6 13.0
54 11.1 9.2 13.8
68 9.9 5.4 8.0#



#22
Acetophenone
Concen: 3.735 ng
RT: 7.269 min Scan# 847
Delta R.T. -0.106 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36

Tgt Ion:105 Resp: 46272
Ion Ratio Lower Upper
105 100
71 5.6 7.3 10.9#
51 22.9 38.8 58.2#
120 7.7 16.5 24.7#









#23
Nitrobenzene-d5
Concen: 1.498 ng
RT: 7.639 min Scan# 910
Delta R.T. 0.112 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36




Instrument :
BNA_F
ClientSampleId :
SB-4D


Tgt Ion: 82 Resp: 15486
Ion Ratio Lower Upper
82 100
128 40.6 30.6 45.8
54 49.6 48.2 72.2



#25
Isophorone
Concen: 3.532 ng
RT: 7.757 min Scan# 930
Delta R.T. -0.030 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36

Tgt Ion: 82 Resp: 63726
Ion Ratio Lower Upper
82 100
95 173.4 5.8 8.8#
138 8.3 11.6 17.4#









#35
 Caprolactam
 Concen: 4.509 ng
 RT: 8.686 min Scan# 1088
 Delta R.T. -0.012 min
 Lab File: BF127828.D
 Acq: 18 Apr 2022 16:36




Instrument :
 BNA_F
 ClientSampleId :
 SB-4D


Tgt Ion:113 Resp: 10126
 Ion Ratio Lower Upper
 113 100
 55 1066.0 218.7 258.7#
 56 450.7 156.2 196.2#



#37
 2-Methylnaphthalene
 Concen: 87.566 ng
 RT: 9.063 min Scan# 1152
 Delta R.T. 0.106 min
 Lab File: BF127828.D
 Acq: 18 Apr 2022 16:36

Tgt Ion:142 Resp: 1414785
 Ion Ratio Lower Upper
 142 100
 141 93.0 70.6 106.0









#38
1-Methylnaphthalene
Concen: 90.722 ng
RT: 9.063 min Scan# 1152
Delta R.T. 0.006 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36




Instrument :
BNA_F
ClientSampleId :
SB-4D


Tgt Ion:142 Resp: 1414785
Ion Ratio Lower Upper
142 100
141 93.0 74.2 111.2



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.004 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36

Tgt Ion:164 Resp: 217932
Ion Ratio Lower Upper
164 100
162 101.8 81.2 121.8
160 45.5 34.6 52.0





#42

2,4,6-Tribromophenol

Concen: 8.420 ng

RT: 10.792 min Scan# 1446

Delta R.T. -0.006 min

Lab File: BF127828.D

Acq: 18 Apr 2022 16:36

Instrument :

BNA_F

ClientSampleId :

SB-4D





Tgt Ion:330 Resp: 19528

Ion Ratio Lower Upper

330 100

332 96.0 79.1 118.7

141 81.6 26.2 39.4#



#45

2-Fluorobiphenyl

Concen: 7.691 ng

RT: 9.310 min Scan# 1194

Delta R.T. -0.012 min

Lab File: BF127828.D

Acq: 18 Apr 2022 16:36




Tgt Ion:172 Resp: 118476


Ion Ratio Lower Upper

172 100

171 35.1 28.4 42.6

170 23.2 18.6 27.8









#48
2-Nitroaniline
Concen: 3.754 ng
RT: 9.522 min Scan# 11
Delta R.T. -0.024 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36




Instrument :
BNA_F
ClientSampleId :
SB-4D


Tgt Ion: 65 Resp: 15443
Ion Ratio Lower Upper
65 100
92 30.3 50.0 75.0#
138 29.8 64.1 96.1#



#49
Acenaphthylene
Concen: 3.746 ng
RT: 9.863 min Scan# 1288
Delta R.T. -0.006 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36

Tgt Ion: 152 Resp: 75459
Ion Ratio Lower Upper
152 100
151 42.1 15.8 23.8#
153 99.0 10.5 15.7#









#51
2,6-Dinitrotoluene
Concen: 2.122 ng
RT: 9.798 min Scan# 11
Delta R.T. 0.012 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36




Instrument :
BNA_F
ClientSampleId :
SB-4D


Tgt Ion:165 Resp: 6319
Ion Ratio Lower Upper
165 100
63 264.5 58.3 87.5#
89 119.4 51.5 77.3#



#52
Acenaphthene
Concen: 10.268 ng
RT: 10.039 min Scan# 1318
Delta R.T. -0.000 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36

Tgt Ion:154 Resp: 130322
Ion Ratio Lower Upper
154 100
153 145.8 87.6 131.4#
152 99.3 42.9 64.3#









#53
3-Nitroaniline
Concen: 3.495 ng
RT: 9.951 min Scan# 1303
Delta R.T. -0.012 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36




Instrument :
BNA_F
ClientSampleId :
SB-4D


Tgt Ion:138 Resp: 10994
Ion Ratio Lower Upper
138 100
108 61.0 9.4 14.0#
92 58.7 106.1 159.1#



#54
2,4-Dinitrophenol
Concen: 6.117 ng
RT: 10.063 min Scan# 1322
Delta R.T. -0.000 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36

Tgt Ion:184 Resp: 1279
Ion Ratio Lower Upper
184 100
63 1802.2 58.1 87.1#
154 3042.0 53.2 79.8#









#55
Dibenzenofuran
Concen: 7.162 ng
RT: 10.210 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36




Instrument :
BNA_F
ClientSampleId :
SB-4D


Tgt Ion:168 Resp: 132663
Ion Ratio Lower Upper
168 100
139 78.2 30.3 45.5#
169 130.8 10.8 16.2#



#56
4-Nitrophenol
Concen: 44.100 ng
RT: 10.110 min Scan# 1330
Delta R.T. -0.000 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36

Tgt Ion:139 Resp: 108961
Ion Ratio Lower Upper
139 100
109 42.3 64.3 104.3#
65 31.0 117.5 157.5#









#57
2,4-Dinitrotoluene
Concen: 13.988 ng
RT: 10.210 min Scan# 11
Delta R.T. 0.017 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36




Instrument :
BNA_F
ClientSampleId :
SB-4D


Tgt Ion:165 Resp: 51640
Ion Ratio Lower Upper
165 100
63 105.9 53.1 79.7#
89 50.0 69.2 103.8#
182 11.3 2.2 3.2#



#58
Fluorene
Concen: 20.098 ng
RT: 10.557 min Scan# 1406
Delta R.T. -0.000 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36

Tgt Ion:166 Resp: 275044
Ion Ratio Lower Upper
166 100
165 102.8 78.1 117.1
167 20.1 10.5 15.7#









#63
Azobenzene
Concen: 2.319 ng
RT: 10.698 min Scan# 1433
Delta R.T. -0.006 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36




Instrument :
BNA_F
ClientSampleId :
SB-4D


Tgt Ion: 77 Resp: 39165
Ion Ratio Lower Upper
77 100
182 283.0 4.5 44.5#
105 74.2 0.3 40.3#
51 46.5 15.0 55.0



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.498 min Scan# 1566
Delta R.T. 0.006 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36

Tgt Ion: 188 Resp: 334650
Ion Ratio Lower Upper
188 100
94 9.9 7.4 11.0
80 10.5 8.0 12.0









#66
 n-Nitrosodiphenylamine
 Concen: 32.039 ng
 RT: 10.663 min Scan# 1424
 Delta R.T. -0.000 min
 Lab File: BF127828.D
 Acq: 18 Apr 2022 16:36




Instrument :
 BNA_F
 ClientSampleId :
 SB-4D


Tgt Ion:169 Resp: 337681
 Ion Ratio Lower Upper
 169 100
 168 17.4 52.7 79.1#
 167 13.2 28.9 43.3#



#71
 Phenanthrene
 Concen: 64.999 ng
 RT: 11.527 min Scan# 1571
 Delta R.T. 0.006 min
 Lab File: BF127828.D
 Acq: 18 Apr 2022 16:36

Tgt Ion:178 Resp: 1152496
 Ion Ratio Lower Upper
 178 100
 176 19.8 15.4 23.2
 179 17.9 12.4 18.6





#72

Anthracene

Concen: 65.711 ng

RT: 11.527 min Scan# 11

Delta R.T. -0.047 min

Lab File: BF127828.D

Acq: 18 Apr 2022 16:36

Instrument :

BNA_F

ClientSampleId :

SB-4D





Tgt Ion:178 Resp: 1152496

Ion Ratio Lower Upper

178 100

176 19.8 14.6 22.0

179 17.9 12.4 18.6



#75

Fluoranthene

Concen: 3.911 ng

RT: 12.716 min Scan# 1773

Delta R.T. 0.006 min

Lab File: BF127828.D

Acq: 18 Apr 2022 16:36




Tgt Ion:202 Resp: 75671


Ion Ratio Lower Upper

202 100

101 13.9 0.0 29.0

203 31.2 0.0 37.7









#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.139 min Scan# 20
Delta R.T. -0.006 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36




Instrument :
BNA_F
ClientSampleId :
SB-4D


Tgt Ion:240 Resp: 293089
Ion Ratio Lower Upper
240 100
120 9.9 7.1 10.7
236 26.4 20.7 31.1



#78
Pyrene
Concen: 14.072 ng
RT: 12.945 min Scan# 1812
Delta R.T. -0.000 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36

Tgt Ion:202 Resp: 348282
Ion Ratio Lower Upper
202 100
200 20.6 17.1 25.7
203 24.8 14.3 21.5#









#79
 Terphenyl-d14
 Concen: 5.864 ng
 RT: 13.080 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: BF127828.D
 Acq: 18 Apr 2022 16:36




Instrument :
 BNA_F
 ClientSampleId :
 SB-4D


Tgt Ion:244 Resp: 104117
 Ion Ratio Lower Upper
 244 100
 212 9.6 5.7 8.5#
 122 10.7 6.2 9.2#



#81
 Benzo(a)anthracene
 Concen: 4.311 ng
 RT: 14.127 min Scan# 2013
 Delta R.T. -0.006 min
 Lab File: BF127828.D
 Acq: 18 Apr 2022 16:36

Tgt Ion:228 Resp: 84398
 Ion Ratio Lower Upper
 228 100
 226 38.9 21.6 32.4#
 229 70.0 15.4 23.2#









#83
Chrysene
Concen: 8.561 ng
RT: 14.162 min Scan# 20
Delta R.T. -0.006 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36

Instrument :
BNA_F
ClientSampleId :
SB-4D

Tgt Ion:228 Resp: 160859
Ion Ratio Lower Upper
228 100
226 35.9 24.2 36.4
229 58.0 15.4 23.2#



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.662 min Scan# 2274
Delta R.T. -0.000 min
Lab File: BF127828.D
Acq: 18 Apr 2022 16:36

Tgt Ion:264 Resp: 293134
Ion Ratio Lower Upper
264 100
260 25.1 19.1 28.7
265 23.1 17.4 26.2

