

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041823\
 Data File : BF132895.D
 Acq On : 18 Apr 2023 10:29
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 18 11:26:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF041723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 18 04:51:21 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.624	0.643	-3.0	102	0.00
3	Pyridine	1.707	1.792	-5.0	103	0.00
4	n-Nitrosodimethylamine	0.811	0.846	-4.3	103	0.00
5 S	2-Fluorophenol	1.279	1.308	-2.3	102	0.00
6	Aniline	2.164	2.280	-5.4	103	0.00
7 S	Phenol-d6	1.632	1.683	-3.1	104	0.00
8	2-Chlorophenol	1.298	1.342	-3.4	103	0.00
9	Benzaldehyde	0.966	0.942	2.5	104	0.00
10 C	Phenol	1.835	1.899	-3.5	104	0.00
11	bis(2-Chloroethyl)ether	1.386	1.424	-2.7	102	0.00
12	1,3-Dichlorobenzene	1.426	1.463	-2.6	102	0.00
13 C	1,4-Dichlorobenzene	1.429	1.452	-1.6	103	0.00
14	1,2-Dichlorobenzene	1.334	1.343	-0.7	103	0.00
15	Benzyl Alcohol	1.119	1.196	-6.9	102	0.00
16	2,2'-oxybis(1-Chloropropane	2.398	2.488	-3.8	103	0.00
17	2-Methylphenol	1.188	1.235	-4.0	103	0.00
18	Hexachloroethane	0.496	0.513	-3.4	102	0.00
19 P	n-Nitroso-di-n-propylamine	1.019	1.048	-2.8	103	0.00
20	3+4-Methylphenols	1.405	1.434	-2.1	103	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22	Acetophenone	0.469	0.462	1.5	103	0.00
23 S	Nitrobenzene-d5	0.384	0.393	-2.3	103	0.00
24	Nitrobenzene	0.393	0.405	-3.1	103	0.00
25	Isophorone	0.728	0.743	-2.1	104	0.00
26 C	2-Nitrophenol	0.148	0.159	-7.4	102	0.00
27	2,4-Dimethylphenol	0.303	0.307	-1.3	103	0.00
28	bis(2-Chloroethoxy)methane	0.435	0.441	-1.4	103	0.00
29 C	2,4-Dichlorophenol	0.282	0.290	-2.8	104	0.00
30	1,2,4-Trichlorobenzene	0.304	0.309	-1.6	103	0.00
31	Naphthalene	1.011	1.007	0.4	103	0.00
32	Benzoic acid	0.152	0.159	-4.6	102	0.00
33	4-Chloroaniline	0.414	0.423	-2.2	103	0.00
34 C	Hexachlorobutadiene	0.174	0.177	-1.7	103	0.00
35	Caprolactam	0.084	0.091	-8.3	104	0.00
36 C	4-Chloro-3-methylphenol	0.296	0.307	-3.7	103	0.00
37	2-Methylnaphthalene	0.696	0.696	0.0	103	0.00
38	1-Methylnaphthalene	0.648	0.644	0.6	102	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	0.566	0.572	-1.1	102	0.00
41 P	Hexachlorocyclopentadiene	0.298	0.324	-8.7	104	0.00
42 S	2,4,6-Tribromophenol	0.201	0.216	-7.5	104	0.00
43 C	2,4,6-Trichlorophenol	0.359	0.380	-5.8	104	0.00
44	2,4,5-Trichlorophenol	0.419	0.449	-7.2	105	0.00
45 S	2-Fluorobiphenyl	1.302	1.244	4.5	101	0.00
46	1,1'-Biphenyl	1.585	1.588	-0.2	103	0.00
47	2-Chloronaphthalene	1.173	1.191	-1.5	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041823\
 Data File : BF132895.D
 Acq On : 18 Apr 2023 10:29
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 18 11:26:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF041723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 18 04:51:21 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.339	0.381	-12.4	104	0.00
49	Acenaphthylene	1.871	1.882	-0.6	102	0.00
50	Dimethylphthalate	1.383	1.429	-3.3	105	0.00
51	2,6-Dinitrotoluene	0.303	0.324	-6.9	103	0.00
52 C	Acenaphthene	1.206	1.228	-1.8	103	0.00
53	3-Nitroaniline	0.345	0.378	-9.6	105	0.00
54 P	2,4-Dinitrophenol	0.118	0.126	-6.8	102	0.00
55	Dibenzofuran	1.711	1.728	-1.0	103	0.00
56 P	4-Nitrophenol	0.256	0.284	-10.9	105	0.00
57	2,4-Dinitrotoluene	0.374	0.414	-10.7	105	0.00
58	Fluorene	1.253	1.236	1.4	103	0.00
59	2,3,4,6-Tetrachlorophenol	0.325	0.338	-4.0	101	0.00
60	Diethylphthalate	1.292	1.348	-4.3	105	0.00
61	4-Chlorophenyl-phenylether	0.637	0.638	-0.2	103	0.00
62	4-Nitroaniline	0.332	0.363	-9.3	105	0.00
63	Azobenzene	1.443	1.473	-2.1	103	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	0.096	0.101	-5.2	103	0.00
66 c	n-Nitrosodiphenylamine	0.651	0.661	-1.5	103	0.00
67	4-Bromophenyl-phenylether	0.228	0.232	-1.8	102	0.00
68	Hexachlorobenzene	0.237	0.242	-2.1	102	0.00
69	Atrazine	0.182	0.196	-7.7	106	0.00
70 C	Pentachlorophenol	0.153	0.165	-7.8	104	0.00
71	Phenanthrene	1.076	1.065	1.0	103	0.00
72	Anthracene	1.101	1.107	-0.5	103	0.00
73	Carbazole	0.953	0.970	-1.8	103	0.00
74	Di-n-butylphthalate	1.021	1.112	-8.9	106	0.00
75 C	Fluoranthene	1.070	1.100	-2.8	105	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
77	Benzydine	0.312	0.361	-15.7	105	0.00
78	Pyrene	1.635	1.664	-1.8	104	0.00
79 S	Terphenyl-d14	1.172	1.177	-0.4	104	0.00
80	Butylbenzylphthalate	0.325	0.395	-21.5	111	0.00
81	Benzo(a)anthracene	1.360	1.430	-5.1	107	0.00
82	3,3'-Dichlorobenzidine	0.329	0.381	-15.8	108	0.00
83	Chrysene	1.352	1.384	-2.4	104	0.00
84	Bis(2-ethylhexyl)phthalate	0.517	0.611	-18.2	110	0.00
85 c	Di-n-octyl phthalate	1.001	1.131	-13.0	110	0.00
86 I	Perylene-d12	1.000	1.000	0.0	107	0.00
87	Indeno(1,2,3-cd)pyrene	1.143	1.201	-5.1	111	0.00
88	Benzo(b)fluoranthene	1.246	1.262	-1.3	108	0.00
89	Benzo(k)fluoranthene	1.241	1.279	-3.1	103	0.00
90 C	Benzo(a)pyrene	1.138	1.209	-6.2	107	0.00
91	Dibenzo(a,h)anthracene	0.974	1.017	-4.4	110	0.00
92	Benzo(g,h,i)perylene	0.932	0.948	-1.7	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041823\
Data File : BF132895.D
Acq On : 18 Apr 2023 10:29
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 18 11:26:37 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF041723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 18 04:51:21 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0