

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041823\
 Data File : BF132895.D
 Acq On : 18 Apr 2023 10:29
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 18 11:26:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF041723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 18 04:51:21 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	41.225	-3.1	102	0.00
3	Pyridine	40.000	41.985	-5.0	103	0.00
4	n-Nitrosodimethylamine	40.000	41.726	-4.3	103	0.00
5 S	2-Fluorophenol	80.000	81.801	-2.3	102	0.00
6	Aniline	40.000	42.140	-5.4	103	0.00
7 S	Phenol-d6	80.000	82.477	-3.1	104	0.00
8	2-Chlorophenol	40.000	41.371	-3.4	103	0.00
9	Benzaldehyde	40.000	39.019	2.5	104	0.00
10 C	Phenol	40.000	41.397	-3.5	104	0.00
11	bis(2-Chloroethyl)ether	40.000	41.087	-2.7	102	0.00
12	1,3-Dichlorobenzene	40.000	41.033	-2.6	102	0.00
13 C	1,4-Dichlorobenzene	40.000	40.634	-1.6	103	0.00
14	1,2-Dichlorobenzene	40.000	40.260	-0.6	103	0.00
15	Benzyl Alcohol	40.000	42.785	-7.0	102	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.504	-3.8	103	0.00
17	2-Methylphenol	40.000	41.575	-3.9	103	0.00
18	Hexachloroethane	40.000	41.384	-3.5	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.162	-2.9	103	0.00
20	3+4-Methylphenols	40.000	40.843	-2.1	103	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	39.366	1.6	103	0.00
23 S	Nitrobenzene-d5	80.000	81.838	-2.3	103	0.00
24	Nitrobenzene	40.000	41.166	-2.9	103	0.00
25	Isophorone	40.000	40.824	-2.1	104	0.00
26 C	2-Nitrophenol	40.000	42.898	-7.2	102	0.00
27	2,4-Dimethylphenol	40.000	40.633	-1.6	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.579	-1.4	103	0.00
29 C	2,4-Dichlorophenol	40.000	41.206	-3.0	104	0.00
30	1,2,4-Trichlorobenzene	40.000	40.659	-1.6	103	0.00
31	Naphthalene	40.000	39.858	0.4	103	0.00
32	Benzoic acid	40.000	41.978	-4.9	102	0.00
33	4-Chloroaniline	40.000	40.852	-2.1	103	0.00
34 C	Hexachlorobutadiene	40.000	40.633	-1.6	103	0.00
35	Caprolactam	40.000	43.609	-9.0	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.578	-3.9	103	0.00
37	2-Methylnaphthalene	40.000	40.002	-0.0	103	0.00
38	1-Methylnaphthalene	40.000	39.770	0.6	102	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.459	-1.1	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.524	-8.8	104	0.00
42 S	2,4,6-Tribromophenol	80.000	86.277	-7.8	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.345	-5.9	104	0.00
44	2,4,5-Trichlorophenol	40.000	42.840	-7.1	105	0.00
45 S	2-Fluorobiphenyl	80.000	76.445	4.4	101	0.00
46	1,1'-Biphenyl	40.000	40.055	-0.1	103	0.00
47	2-Chloronaphthalene	40.000	40.599	-1.5	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041823\
 Data File : BF132895.D
 Acq On : 18 Apr 2023 10:29
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 18 11:26:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF041723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 18 04:51:21 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.936	-12.3	104	0.00
49	Acenaphthylene	40.000	40.226	-0.6	102	0.00
50	Dimethylphthalate	40.000	41.344	-3.4	105	0.00
51	2,6-Dinitrotoluene	40.000	42.796	-7.0	103	0.00
52 C	Acenaphthene	40.000	40.745	-1.9	103	0.00
53	3-Nitroaniline	40.000	43.801	-9.5	105	0.00
54 P	2,4-Dinitrophenol	40.000	42.791	-7.0	102	0.00
55	Dibenzofuran	40.000	40.400	-1.0	103	0.00
56 P	4-Nitrophenol	40.000	44.336	-10.8	105	0.00
57	2,4-Dinitrotoluene	40.000	44.290	-10.7	105	0.00
58	Fluorene	40.000	39.443	1.4	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.496	-3.7	101	0.00
60	Diethylphthalate	40.000	41.721	-4.3	105	0.00
61	4-Chlorophenyl-phenylether	40.000	40.010	-0.0	103	0.00
62	4-Nitroaniline	40.000	43.736	-9.3	105	0.00
63	Azobenzene	40.000	40.837	-2.1	103	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.323	-5.8	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.577	-1.4	103	0.00
67	4-Bromophenyl-phenylether	40.000	40.661	-1.7	102	0.00
68	Hexachlorobenzene	40.000	40.789	-2.0	102	0.00
69	Atrazine	40.000	43.026	-7.6	106	0.00
70 C	Pentachlorophenol	40.000	43.295	-8.2	104	0.00
71	Phenanthrene	40.000	39.603	1.0	103	0.00
72	Anthracene	40.000	40.218	-0.5	103	0.00
73	Carbazole	40.000	40.737	-1.8	103	0.00
74	Di-n-butylphthalate	40.000	43.549	-8.9	106	0.00
75 C	Fluoranthene	40.000	41.124	-2.8	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzydine	40.000	46.353	-15.9	105	0.00
78	Pyrene	40.000	40.694	-1.7	104	0.00
79 S	Terphenyl-d14	80.000	80.317	-0.4	104	0.00
80	Butylbenzylphthalate	40.000	40.878	-2.2	111	0.00
81	Benzo(a)anthracene	40.000	42.055	-5.1	107	0.00
82	3,3'-Dichlorobenzidine	40.000	46.356	-15.9	108	0.00
83	Chrysene	40.000	40.944	-2.4	104	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	47.275	-18.2	110	0.00
85 c	Di-n-octyl phthalate	40.000	41.273	-3.2	110	0.00
86 I	Perylene-d12	20.000	20.000	0.0	107	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.024	-5.1	111	0.00
88	Benzo(b)fluoranthene	40.000	40.513	-1.3	108	0.00
89	Benzo(k)fluoranthene	40.000	41.236	-3.1	103	0.00
90 C	Benzo(a)pyrene	40.000	42.496	-6.2	107	0.00
91	Dibenzo(a,h)anthracene	40.000	41.786	-4.5	110	0.00
92	Benzo(g,h,i)perylene	40.000	40.690	-1.7	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041823\
Data File : BF132895.D
Acq On : 18 Apr 2023 10:29
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 18 11:26:37 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF041723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 18 04:51:21 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0