

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042015\
 Data File : BF078681.D
 Acq On : 21 Apr 2015 12:07
 Operator : TP/IZ
 Sample : G1902-01
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 LS-SB18A

Manual Integrations
 APPROVED

apatel
 4/21/2015 5:05:08 PM

Quant Time: Apr 21 13:17:50 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 20 16:32:38 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.86	152	39209	20.00	ng	0.00
21) Naphthalene-d8	8.02	136	157287	20.00	ng	0.00
38) Acenaphthene-d10	11.21	164	77100	20.00	ng	0.00
63) Phenanthrene-d10	13.94	188	142635	20.00	ng	0.00
75) Chrysene-d12	17.81	240	150995	20.00	ng	0.00
86) Perylene-d12	19.57	264	152016	20.00	ng	0.03
System Monitoring Compounds						
5) 2-Fluorophenol	3.87	112	233984	98.39	ng	0.00
7) Phenol-d6	5.37	99	297916	99.96	ng	0.01
23) Nitrobenzene-d5	6.80	82	167825	64.67	ng	0.00
41) 2,4,6-Tribromophenol	12.69	330	66417	103.87	ng	0.00
44) 2-Fluorobiphenyl	10.03	172	299356	55.50	ng	0.00
78) Terphenyl-d14	16.47	244	315176	47.17	ng	0.00
Target Compounds						Qvalue
20) 3+4-Methylphenols	6.65	107	6058	2.08	ng	96
49) Dimethylphthalate	10.83	163	20052	3.46	ng	# 97
70) Phenanthrene	13.99	178	39386	4.77	ng	96
71) Anthracene	14.08	178	20392	2.51	ng	# 87
74) Fluoranthene	15.86	202	102874	11.82	ng	98
77) Pyrene	16.18	202	142429	15.03	ng	# 95
80) Benzo(a)anthracene	17.79	228	60152	6.70	ng	96
82) Chrysene	17.84	228	46950m	5.56	ng	
83) Bis(2-ethylhexyl)phthalate	17.93	149	50986	9.00	ng	# 98
85) Indeno(1,2,3-cd)pyrene	20.79	276	29888m	3.12	ng	
87) Benzo(b)fluoranthene	19.12	252	68025m	7.24	ng	
88) Benzo(k)fluoranthene	19.15	252	18633m	2.23	ng	
89) Benzo(a)pyrene	19.50	252	43273	5.28	ng	# 80
91) Benzo(a,h,i)perylene	21.11	276	28876m	3.51	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042015\
Data File : BF078681.D
Acq On : 21 Apr 2015 12:07
Operator : TP/IZ
Sample : G1902-01
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
LS-SB18A

Manual Integrations
APPROVED

apatel
4/21/2015 5:05:08 PM

Quant Time: Apr 21 13:17:50 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 20 16:32:38 2015
Response via : Initial Calibration

