

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042015\
 Data File : BF078677.D
 Acq On : 21 Apr 2015 10:03
 Operator : TP/IZ
 Sample : G1891-12DL 5X
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-69SDL

Manual Integrations
 APPROVED

apatel
 4/21/2015 5:04:51 PM

Quant Time: Apr 21 13:00:48 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 20 16:32:38 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.86	152	6922	20.00	ng	0.00
21) Naphthalene-d8	8.02	136	28248	20.00	ng	0.00
38) Acenaphthene-d10	11.20	164	14397	20.00	ng	-0.01
63) Phenanthrene-d10	13.94	188	27706	20.00	ng	0.00
75) Chrysene-d12	17.81	240	29897	20.00	ng	0.00
86) Perylene-d12	19.55	264	30131	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	3.87	112	39381	93.80	ng	0.00
7) Phenol-d6	5.36	99	51222	97.35	ng	0.00
23) Nitrobenzene-d5	6.79	82	29565	63.44	ng	-0.01
41) 2,4,6-Tribromophenol	12.67	330	11953	100.11	ng	-0.01
44) 2-Fluorobiphenyl	10.02	172	67247	66.77	ng	-0.01
78) Terphenyl-d14	16.46	244	77188	58.34	ng	-0.01
Target Compounds						Qvalue
31) Naphthalene	8.06	128	14798	10.06	ng	99
37) 2-Methylnaphthalene	9.31	142	5767	5.78	ng	94
45) 1,1'-Biphenyl	10.19	154	2393	2.03	ng	91
48) Acenaphthylene	10.94	152	5340	3.57	ng	97
49) Dimethylphthalate	10.83	163	2309	2.13	ng	# 76
51) Acenaphthene	11.26	154	15442	18.33	ng	98
54) Dibenzofuran	11.59	168	15176	11.80	ng	98
57) Fluorene	12.22	166	18601	17.84	ng	99
70) Phenanthrene	13.98	178	221580	138.26	ng	100
71) Anthracene	14.07	178	49848	31.56	ng	99
72) Carbazole	14.40	167	17959	12.13	ng	98
74) Fluoranthene	15.86	202	247014	146.15	ng	96
77) Pvrene	16.17	202	231967	123.60	ng	# 96
80) Benzo(a)anthracene	17.79	228	105302	59.20	ng	100
82) Chrysene	17.83	228	94677	56.65	ng	97
85) Indeno(1,2,3-cd)pyrene	20.77	276	51475	27.14	ng	# 86
87) Benzo(b)fluoranthene	19.11	252	117269m	62.97	ng	
88) Benzo(k)fluoranthene	19.14	252	35153m	21.24	ng	
89) Benzo(a)pvrene	19.49	252	90536	55.71	ng	# 94
90) Dibenzo(a,h)anthracene	20.79	278	11199m	6.88	ng	
91) Benzo(g,h,i)perylene	21.09	276	52641	32.27	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA F\DATA\BF042015\
Data File : BF078677.D
Acq On : 21 Apr 2015 10:03
Operator : TP/IZ
Sample : G1891-12DL 5X
Misc :
ALS Vial : 30 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
SB-69SDL

Manual Integrations
APPROVED

apatel
4/21/2015 5:04:51 PM

Quant Time: Apr 21 13:00:48 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 20 16:32:38 2015
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 5.86 min Scan# 409
Delta R.T. 0.00 min
Lab File: BF078677.D
Acq: 21 Apr 2015 10:03

Instrument :
BNA_F
ClientSampleId :
SB-69SDL

Tot Ion	Ratio	Lower	Upper
152	100		
150	146.1	125.9	188.9
115	54.8	52.7	79.1

Manual Integrations
APPROVED




apatel
4/21/2015 5:04:51 PM



#5

2-Fluorophenol
Concen: 93.80 ng
RT: 3.87 min Scan# 235
Delta R.T. 0.00 min
Lab File: BF078677.D
Acq: 21 Apr 2015 10:03

Tgt Ion	Ratio	Lower	Upper
112	100		
64	59.2	39.9	59.9
63	28.9	20.2	30.4





#7

Phenol-d6

Concen: 97.35 ng

RT: 5.36 min Scan# 365

Delta R.T. 0.00 min

Lab File: BF078677.D

Acq: 21 Apr 2015 10:03

Instrument :

BNA_F





ClientSampleId :

SB-69SDL

Tot Ion:	99	Resp:	51222
Ion Ratio	Lower	Upper	
99	100		
42	16.8	14.8	22.2
71	31.4	26.6	39.8

Manual Integrations
APPROVED

apatel
4/21/2015 5:04:51 PM



#21

Naphthalene-d8

Concen: 20.00 ng




RT: 8.02 min Scan# 598


Delta R.T. 0.00 min

Lab File: BF078677.D

Acq: 21 Apr 2015 10:03

Tgt Ion:	136	Resp:	28248
Ion Ratio	Lower	Upper	
136	100		
137	10.6	9.0	13.4
54	6.9	7.0	10.6#
68	5.0	5.4	8.2#









#23
 Nitrobenzene-d5
 Concen: 63.44 ng
 RT: 6.79 min Scan# 490
 Delta R.T. -0.01 min
 Lab File: BF078677.D
 Acq: 21 Apr 2015 10:03

Instrument :
 BNA_F
 ClientSampleId :
 SB-69SDL

Tot Ion:	82	Resp:	29565
Ion Ratio	Lower	Upper	
82	100		
128	39.2	31.8	47.8
54	54.6	41.5	62.3




Manual Integrations
 APPROVED


apatel
 4/21/2015 5:04:51 PM



#31
 Naphthalene
 Concen: 10.06 ng
 RT: 8.06 min Scan# 601
 Delta R.T. 0.00 min
 Lab File: BF078677.D
 Acq: 21 Apr 2015 10:03

Tgt Ion:	128	Resp:	14798
Ion Ratio	Lower	Upper	
128	100		
129	11.0	8.6	12.8
127	12.6	10.7	16.1









#37
2-Methylnaphthalene
Concen: 5.78 ng
RT: 9.31 min Scan# 711
Delta R.T. -0.01 min
Lab File: BF078677.D
Acq: 21 Apr 2015 10:03

Instrument :
BNA_F
ClientSampled :
SB-69SDL

Tot Ion	Ratio	Lower	Upper
142	100		
141	91.9	70.3	105.5
115	36.1	24.6	37.0




Manual Integrations
APPROVED


apatel
4/21/2015 5:04:51 PM



#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 11.20 min Scan# 876
Delta R.T. -0.01 min
Lab File: BF078677.D
Acq: 21 Apr 2015 10:03

Tgt Ion	Ratio	Lower	Upper
164	100		
162	103.3	82.2	123.2
160	45.1	34.7	52.1









#41
 2,4,6-Tribromophenol
 Concen: 100.11 ng
 RT: 12.67 min Scan# 1005
 Delta R.T. -0.01 min
 Lab File: BF078677.D
 Acq: 21 Apr 2015 10:03

Instrument :
 BNA_F
 ClientSampleId :
 SB-69SDL

Tot Ion	Ratio	Lower	Upper
330	100		
332	97.9	78.6	117.8
141	39.1	31.2	46.8




Manual Integrations
 APPROVED


apatel
 4/21/2015 5:04:51 PM



#44
 2-Fluorobiphenyl
 Concen: 66.77 ng
 RT: 10.02 min Scan# 773
 Delta R.T. -0.01 min
 Lab File: BF078677.D
 Acq: 21 Apr 2015 10:03

Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.5	28.6	42.8
170	23.1	18.5	27.7





#45

1,1'-Biphenyl

Concen: 2.03 ng

RT: 10.19 min Scan# 788

Delta R.T. -0.01 min

Lab File: BF078677.D

Acq: 21 Apr 2015 10:03

Instrument :

BNA_F


ClientSampleId :

SB-69SDL




Tot Ion:	154	Resp:	2393
Ion Ratio	Lower	Upper	
154	100		
153	42.3	20.0	60.0
76	0.0	0.0	30.6

Manual Integrations
APPROVED

apatel
4/21/2015 5:04:51 PM



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 76.00 (75.70 to 76.70): BF0



#48

Acenaphthylene

Concen: 3.57 ng


RT: 10.94 min Scan# 853

Delta R.T. -0.01 min



Lab File: BF078677.D


Acq: 21 Apr 2015 10:03

Tgt Ion:	152	Resp:	5340
Ion Ratio	Lower	Upper	
152	100		
151	21.9	15.7	23.5
153	13.1	10.2	15.2



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E









#49
Dimethylphthalate
Concen: 2.13 ng
RT: 10.83 min Scan# 844
Delta R.T. -0.01 min
Lab File: BF078677.D
Acq: 21 Apr 2015 10:03

Instrument :
BNA_F
ClientSampled :
SB-69SDL

Tot Ion	Ratio	Lower	Upper
163	100		
194	0.0	2.3	3.5#
164	0.0	8.6	13.0#




Manual Integrations
APPROVED


apatel
4/21/2015 5:04:51 PM



#51
Acenaphthene
Concen: 18.33 ng
RT: 11.26 min Scan# 881
Delta R.T. -0.01 min
Lab File: BF078677.D
Acq: 21 Apr 2015 10:03

Tgt Ion	Ratio	Lower	Upper
154	100		
153	113.4	91.4	137.2
152	56.0	43.0	64.6






#54
Dibenzenofuran
Concen: 11.80 ng
RT: 11.59 min Scan# 910
Delta R.T. -0.01 min
Lab File: BF078677.D
Acq: 21 Apr 2015 10:03

Instrument :
BNA_F
ClientSampleId :
SB-69SDL




Tot Ion	Ratio	Lower	Upper
168	100		
139	39.8	31.0	46.4
169	13.2	10.7	16.1

Manual Integrations
APPROVED

apatel
4/21/2015 5:04:51 PM






Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E




#57
Fluorene
Concen: 17.84 ng
RT: 12.22 min Scan# 965
Delta R.T. -0.01 min
Lab File: BF078677.D
Acq: 21 Apr 2015 10:03

Tgt Ion	Ratio	Lower	Upper
166	100		
165	97.6	78.9	118.3
167	14.5	10.7	16.1

Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E









#63
 Phenanthrene-d10
 Concen: 20.00 ng
 RT: 13.94 min Scan# 1116
 Delta R.T. 0.00 min
 Lab File: BF078677.D
 Acq: 21 Apr 2015 10:03

Instrument :
 BNA_F
 ClientSampled :
 SB-69SDL

Tot Ion	Ratio	Resp Lower	Upper
188	100		
94	8.0	9.0	13.4#
80	8.3	9.5	14.3#




Manual Integrations
 APPROVED


apatel
 4/21/2015 5:04:51 PM



#70
 Phenanthrene
 Concen: 138.26 ng
 RT: 13.98 min Scan# 1119
 Delta R.T. -0.01 min
 Lab File: BF078677.D
 Acq: 21 Apr 2015 10:03

Tgt Ion	Ratio	Resp Lower	Upper
178	100		
176	19.3	15.4	23.0
179	15.4	12.2	18.2





#71

Anthracene

Concen: 31.56 ng

RT: 14.07 min Scan# 1127

Delta R.T. -0.01 min

Lab File: BF078677.D

Acq: 21 Apr 2015 10:03

Instrument :

BNA_F

ClientSampleId :

SB-69SDL

Tot Ion:178 Resp: 49848

Ion Ratio Lower Upper

178 100


176 18.7 14.8 22.2

179 16.5 12.4 18.6

Manual Integrations
APPROVED

apatel




4/21/2015 5:04:51 PM



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Carbazole

Concen: 12.13 ng

RT: 14.40 min Scan# 1156

Delta R.T. -0.01 min

Lab File: BF078677.D

Acq: 21 Apr 2015 10:03

Tgt Ion:167 Resp: 17959

Ion Ratio Lower Upper

167 100




166 19.6 16.8 25.2


139 13.1 10.0 15.0

Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#74

Fluoranthene

Concen: 146.15 ng

RT: 15.86 min Scan# 1284

Delta R.T. -0.01 min

Lab File: BF078677.D

Acq: 21 Apr 2015 10:03

Instrument :

BNA_F





ClientSampleId :

SB-69SDL

Tot Ion	Ratio	Resp Lower	Upper
202	100		
101	12.6	0.0	35.2
203	17.3	0.0	38.5

Manual Integrations
APPROVED

apatel
4/21/2015 5:04:51 PM



#75

Chrysene-d12

Concen: 20.00 ng




RT: 17.81 min Scan# 1454


Delta R.T. 0.00 min

Lab File: BF078677.D

Acq: 21 Apr 2015 10:03

Tgt Ion	Ratio	Resp Lower	Upper
240	100		
120	8.5	7.1	10.7
236	25.6	21.4	32.2





#77

Pvrene

Concen: 123.60 ng

RT: 16.17 min Scan# 1311

Delta R.T. -0.01 min

Lab File: BF078677.D

Acq: 21 Apr 2015 10:03

Instrument :

BNA_F


ClientSampleId :

SB-69SDL




Tot Ion	Ratio	Resp Lower	Upper
202	100		
200	20.6	16.3	24.5
203	20.8	13.8	20.6

Manual Integrations
APPROVED

apatel
4/21/2015 5:04:51 PM



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 58.34 ng

RT: 16.46 min Scan# 1336




Delta R.T. -0.01 min


Lab File: BF078677.D

Acq: 21 Apr 2015 10:03

Tgt Ion	Ratio	Resp Lower	Upper
244	100		
212	6.5	5.0	7.4
122	13.6	7.1	10.7

Abundance Ion 244.00 (243.70 to 244.70): E
Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E









#80
Benzo(a)anthracene
Concen: 59.20 ng
RT: 17.79 min Scan# 1453
Delta R.T. 0.00 min
Lab File: BF078677.D
Acq: 21 Apr 2015 10:03

Instrument :
BNA_F
ClientSampleId :
SB-69SDL

Tot Ion	Ratio	Lower	Upper
228	100		
226	26.7	21.3	31.9
229	19.6	15.4	23.2




Manual Integrations
APPROVED


apatel
4/21/2015 5:04:51 PM



#82
Chrysene
Concen: 56.65 ng
RT: 17.83 min Scan# 1456
Delta R.T. -0.01 min
Lab File: BF078677.D
Acq: 21 Apr 2015 10:03

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.2	24.2	36.4
229	22.2	15.8	23.8









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 27.14 ng
 RT: 20.77 min Scan# 1713
 Delta R.T. 0.05 min
 Lab File: BF078677.D
 Acq: 21 Apr 2015 10:03

Instrument :
 BNA_F
 ClientSampleId :
 SB-69SDL

Tot Ion	Ratio	Lower	Upper
276	100		
138	25.6	16.9	25.3#
277	27.9	15.8	23.6#




Manual Integrations
 APPROVED


apatel
 4/21/2015 5:04:51 PM



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 19.55 min Scan# 1607
 Delta R.T. 0.02 min
 Lab File: BF078677.D
 Acq: 21 Apr 2015 10:03

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.9	18.6	27.8
265	23.5	17.0	25.6





#87

Benzo(b)fluoranthene

Concen: 62.97 ng m

RT: 19.11 min Scan# 1568

Delta R.T. 0.01 min

Lab File: BF078677.D

Acq: 21 Apr 2015 10:03

Instrument :

BNA_F


ClientSampleId :

SB-69SDL




Tot Ion:	252	Resp:	117269
Ion Ratio	Lower	Upper	
252	100		
253	22.5	17.2	25.8
125	12.2	7.0	10.4#

Manual Integrations
APPROVED

apatel
4/21/2015 5:04:51 PM



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 21.24 ng m

RT: 19.14 min Scan# 1571




Delta R.T. 0.01 min


Lab File: BF078677.D

Acq: 21 Apr 2015 10:03

Tgt Ion:	252	Resp:	35153
Ion Ratio	Lower	Upper	
252	100		
253	22.6	17.5	26.3
125	10.8	11.6	17.4#

Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#89

Benzo(a)pyrene

Concen: 55.71 ng

RT: 19.49 min Scan# 1601

Delta R.T. 0.02 min

Lab File: BF078677.D

Acq: 21 Apr 2015 10:03

Instrument :

BNA_F





ClientSampleId :

SB-69SDL

Tot Ion:	252	Resp:	90536
Ion Ratio	Lower	Upper	
252	100		
253	23.0	16.8	25.2
125	13.4	7.4	11.0

Manual Integrations
APPROVED

apatel
4/21/2015 5:04:51 PM



#90

Dibenzo(a,h)anthracene

Concen: 6.88 ng m




RT: 20.79 min Scan# 1715




Delta R.T. 0.05 min

Lab File: BF078677.D

Acq: 21 Apr 2015 10:03

Tgt Ion:	278	Resp:	11199
Ion Ratio	Lower	Upper	
278	100		
139	20.8	13.6	20.4
279	30.1	19.0	28.6





#91
 Benzo(a,h,i)perylene
 Concen: 32.27 ng
 RT: 21.09 min Scan# 1741
 Delta R.T. 0.05 min
 Lab File: BF078677.D
 Acq: 21 Apr 2015 10:03

Instrument :
 BNA_F
 ClientSampleId :
 SB-69SDL

Tot Ion	Ratio	Lower	Upper
276	100		
277	25.3	19.0	28.4
138	25.4	19.9	29.9

Manual Integrations
 APPROVED

apatel
 4/21/2015 5:04:51 PM

