



4,6-Dinitro-2-methylphenol



Resp Ratio = $7.20 \times 10^{-2} \times \text{Amt} - 2.18 \times 10^{-2}$
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF042016.M
Calibration Table Last Updated: Fri Apr 29 06:55:27 2016

Hexachlorocyclopentadiene




Resp Ratio = $1.34 \times 10^{-1} \times \text{Amt} - 1.28 \times 10^{-2}$
Coef of Det (r^2) = 0.992 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF042016.M
Calibration Table Last Updated: Wed Apr 20 15:45:26 2016


2-Fluorobiphenyl

Response Ratio



$R = -3.71e-002 A^2 + 1.13e+000 A + 3.30e-001$
Coef of Det (r^2) = 0.994 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF042016.M
Calibration Table Last Updated: Fri Apr 29 06:54:33 2016



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF042016.M
Calibration Table Last Updated: Fri Apr 29 06:54:57 2016