



Data Path : Z:\HPCHEM1\BNA F\DATA\BF042017\
Data File : BF094535.D
Acq On : 20 Apr 2017 10:40
Operator : SJ/MA
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 21 05:36:31 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 17 14:59:23 2017
Response via : Initial Calibration



TIC: BF094535.D

(76) Benzidine



12.883min (-0.003) 0.00ng

response 638992

Ion	Exp%	Act%
184.00	100	100
185.00	19.40	14.00#
183.00	12.20	12.46
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042017\
Data File : BF094535.D
Acq On : 20 Apr 2017 10:40
Operator : SJ/MA
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 21 05:36:31 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 17 14:59:23 2017
Response via : Initial Calibration



TIC: BF094535.D

(69) Pentachlorophenol (C)

10.995min (-0.009) 0.00ng

response 83443

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	64.14
264.00	64.60	62.71
0.00	0.00	0.00