

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042023\
 Data File : BF132947.D
 Acq On : 20 Apr 2023 20:34
 Operator : CG\JU
 Sample : 02386-08
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SPA-2WC-23-04

Quant Time: Apr 21 02:16:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF041723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 20 14:15:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.757	152	290549	20.000	ng	0.00
21) Naphthalene-d8	8.045	136	1171016	20.000	ng	0.00
39) Acenaphthene-d10	9.798	164	614992	20.000	ng	0.00
64) Phenanthrene-d10	11.274	188	1088510	20.000	ng	0.00
76) Chrysene-d12	13.910	240	779292	20.000	ng	# 0.00
86) Perylene-d12	15.339	264	607339	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.381	112	2563281	137.950	ng	0.02
7) Phenol-d6	6.392	99	3066724	129.341	ng	0.00
23) Nitrobenzene-d5	7.322	82	2143860	95.429	ng	0.00
42) 2,4,6-Tribromophenol	10.586	330	986065	159.857	ng	0.00
45) 2-Fluorobiphenyl	9.122	172	3491665	87.220	ng	0.00
79) Terphenyl-d14	12.868	244	4073180	89.173	ng	0.00
Target Compounds						
						Qvalue
13) 1,4-Dichlorobenzene	6.775	146	87845	4.231	ng	99
14) 1,2-Dichlorobenzene	6.928	146	145349	7.498	ng	97
27) 2,4-Dimethylphenol	7.692	122	92129	5.199	ng	99
31) Naphthalene	8.063	128	266926	4.511	ng	99
80) Butylbenzylphthalate	13.339	149	4983	5.423	ng	95
84) Bis(2-ethylhexyl)phtha...	13.910	149	90364	4.483	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042023\
Data File : BF132947.D
Acq On : 20 Apr 2023 20:34
Operator : CG\JU
Sample : 02386-08
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SPA-2WC-23-04

Quant Time: Apr 21 02:16:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF041723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 20 14:15:37 2023
Response via : Initial Calibration

