

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042115\
 Data File : BF078691.D
 Acq On : 21 Apr 2015 19:03
 Operator : TP/IZ
 Sample : G1938-03
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-67S

Manual Integrations
 APPROVED

apatel
 4/22/2015 5:22:17 PM

Quant Time: Apr 22 03:27:09 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 22 00:52:13 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.82	152	35703	20.00	ng	0.00
21) Naphthalene-d8	7.98	136	148563	20.00	ng	-0.01
38) Acenaphthene-d10	11.16	164	76448	20.00	ng	-0.01
63) Phenanthrene-d10	13.90	188	145851	20.00	ng	-0.01
75) Chrysene-d12	17.77	240	159616	20.00	ng	-0.01
86) Perylene-d12	19.46	264	156176	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	3.84	112	244508	112.91	ng	0.01
7) Phenol-d6	5.32	99	308100	113.53	ng	0.00
23) Nitrobenzene-d5	6.75	82	183819	75.00	ng	-0.01
41) 2,4,6-Tribromophenol	12.64	330	78484	123.79	ng	-0.01
44) 2-Fluorobiphenyl	9.99	172	376840	70.46	ng	-0.01
78) Terphenyl-d14	16.43	244	411815	58.30	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	10.80	163	15089	2.63	ng	99
70) Phenanthrene	13.94	178	21374	2.53	ng	99
74) Fluoranthene	15.83	202	36621	4.12	ng	96
77) Pyrene	16.14	202	34573	3.45	ng	# 96
82) Chrysene	17.79	228	18182m	2.04	ng	
87) Benzo(b)fluoranthene	19.04	252	23384m	2.42	ng	
89) Benzo(a)pyrene	19.39	252	17958	2.13	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042115\
Data File : BF078691.D
Acq On : 21 Apr 2015 19:03
Operator : TP/IZ
Sample : G1938-03
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-67S

Manual Integrations
APPROVED

apatel
4/22/2015 5:22:17 PM

Quant Time: Apr 22 03:27:09 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 22 00:52:13 2015
Response via : Initial Calibration

