

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042115\
 Data File : BF078695.D
 Acq On : 21 Apr 2015 21:07
 Operator : TP/IZ
 Sample : G1938-07
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-65AS

Manual Integrations
 APPROVED

apatel
 4/22/2015 5:22:33 PM

Quant Time: Apr 22 03:39:09 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 22 00:52:13 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.82	152	34116	20.00	ng	0.00
21) Naphthalene-d8	7.98	136	142775	20.00	ng	-0.01
38) Acenaphthene-d10	11.17	164	71876	20.00	ng	-0.01
63) Phenanthrene-d10	13.90	188	144008	20.00	ng	-0.01
75) Chrysene-d12	17.77	240	150592	20.00	ng	-0.01
86) Perylene-d12	19.49	264	157142	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	3.84	112	234146	113.16	ng	0.01
7) Phenol-d6	5.32	99	296912	114.50	ng	0.00
23) Nitrobenzene-d5	6.75	82	177210	75.23	ng	-0.01
41) 2,4,6-Tribromophenol	12.64	330	76274	127.95	ng	-0.01
44) 2-Fluorobiphenyl	9.99	172	366439	72.87	ng	-0.01
78) Terphenyl-d14	16.43	244	413518	62.05	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	10.80	163	14496	2.68	ng	# 99
51) Acenaphthene	11.22	154	12586	2.99	ng	97
54) Dibenzofuran	11.55	168	13707	2.13	ng	94
57) Fluorene	12.18	166	15169	2.91	ng	95
70) Phenanthrene	13.94	178	287014	34.45	ng	99
71) Anthracene	14.03	178	56469	6.88	ng	98
72) Carbazole	14.37	167	23315	3.03	ng	99
74) Fluoranthene	15.83	202	335164	38.15	ng	97
77) Pyrene	16.15	202	321410	34.00	ng	# 95
80) Benzo(a)anthracene	17.76	228	171392	19.13	ng	99
82) Chrysene	17.81	228	128420	15.25	ng	97
85) Indeno(1,2,3-cd)pyrene	20.65	276	80333	8.41	ng	# 86
87) Benzo(b)fluoranthene	19.05	252	175117m	18.03	ng	
88) Benzo(k)fluoranthene	19.09	252	46767m	5.42	ng	
89) Benzo(a)pyrene	19.42	252	127975	15.10	ng	# 94
90) Dibenzo(a,h)anthracene	20.67	278	19546m	2.30	ng	
91) Benzo(a,h,i)perylene	20.97	276	79594	9.36	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042115\
Data File : BF078695.D
Acq On : 21 Apr 2015 21:07
Operator : TP/IZ
Sample : G1938-07
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-65AS

Manual Integrations
APPROVED

apatel
4/22/2015 5:22:33 PM

Quant Time: Apr 22 03:39:09 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 22 00:52:13 2015
Response via : Initial Calibration

