

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042115\
 Data File : BF078708.D
 Acq On : 22 Apr 2015 6:07
 Operator : TP/IZ
 Sample : G1891-12DL 5X
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-69SDL

Manual Integrations
 APPROVED

apatel
 4/22/2015 5:23:28 PM

Quant Time: Apr 22 06:54:32 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 22 05:03:12 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.82	152	31710	20.00	ng	0.00
21) Naphthalene-d8	7.98	136	134294	20.00	ng	-0.01
38) Acenaphthene-d10	11.16	164	68750	20.00	ng	-0.01
63) Phenanthrene-d10	13.90	188	133832	20.00	ng	-0.01
75) Chrysene-d12	17.77	240	147002	20.00	ng	0.00
86) Perylene-d12	19.46	264	151024	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	3.84	112	36002	18.72	ng	0.00
7) Phenol-d6	5.32	99	47542	19.72	ng	0.00
23) Nitrobenzene-d5	6.75	82	27123	12.24	ng	-0.01
41) 2,4,6-Tribromophenol	12.64	330	11157	19.57	ng	-0.01
44) 2-Fluorobiphenyl	9.99	172	62620	13.02	ng	-0.01
78) Terphenyl-d14	16.43	244	73910	11.36	ng	0.00
Target Compounds						Qvalue
51) Acenaphthene	11.22	154	14451	3.59	ng	97
54) Dibenzofuran	11.55	168	14229	2.32	ng	96
57) Fluorene	12.18	166	16710	3.36	ng	99
70) Phenanthrene	13.94	178	208071	26.88	ng	99
71) Anthracene	14.03	178	48192	6.32	ng	99
72) Carbazole	14.37	167	16503	2.31	ng	97
74) Fluoranthene	15.83	202	237659	29.11	ng	98
77) Pyrene	16.14	202	220206	23.86	ng	96
80) Benzo(a)anthracene	17.76	228	102972	11.77	ng	98
82) Chrysene	17.79	228	93166	11.34	ng	98
85) Indeno(1,2,3-cd)pyrene	20.63	276	53988	5.79	ng	# 83
87) Benzo(b)fluoranthene	19.04	252	123875m	13.27	ng	
88) Benzo(k)fluoranthene	19.07	252	28027m	3.38	ng	
89) Benzo(a)pyrene	19.41	252	88790	10.90	ng	96
91) Benzo(a,h,i)perylene	20.94	276	51373	6.28	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042115\
Data File : BF078708.D
Acq On : 22 Apr 2015 6:07
Operator : TP/IZ
Sample : G1891-12DL 5X
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-69SDL

Manual Integrations
APPROVED

apatel
4/22/2015 5:23:28 PM

Quant Time: Apr 22 06:54:32 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 22 05:03:12 2015
Response via : Initial Calibration

