



Data Path : Z:\HPCHEM1\BNA_F\DATA\BF042115\
Data File : BF078705.D
Acq On : 22 Apr 2015 3:52
Operator : TP/IZ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 22 07:22:51 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 22 05:03:12 2015
Response via : Initial Calibration



TIC: BF078705.D

(69) Pentachlorophenol (C)



13.600min (+0.000) 0.00ng

response 14285

Ion	Exp%	Act%
265.70	100	100
268.00	62.60	61.88
264.00	62.50	64.75
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF042115\
Data File : BF078705.D
Acq On : 22 Apr 2015 3:52
Operator : TP/IZ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 22 07:22:51 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 22 05:03:12 2015
Response via : Initial Calibration



TIC: BF078705.D

(76) Benzidine

16.091min (-0.011) 0.00ng

response 184229

Ion	Exp%	Act%
184.00	100	100
185.00	15.20	13.79
183.00	11.90	11.59
0.00	0.00	0.00